

NOTE

This manual documents the Models 732A and 732A/AN and their assemblies at the revision levels shown in Appendix 7A.

732A
732A/AN
DC Reference
Standard

Instruction Manual

P/N 788414

MAY 1986

©1986, John Fluke Mfg. Co., Inc. All Rights Reserved. Litho in U.S.A.



WARRANTY

Notwithstanding any provision of any agreement the following warranty is exclusive:

The JOHN FLUKE MFG. CO., INC., warrants each instrument it manufactures to be free from defects in material and workmanship under normal use and service for the period of 1-year from date of purchase. This warranty extends only to the original purchaser. This warranty shall not apply to fuses, disposable batteries (rechargeable type batteries are warranted for 90-days), or any product or parts which have been subject to misuse, neglect, accident, or abnormal conditions of operations.

In the event of failure of a product covered by this warranty, John Fluke Mfg. Co., Inc., will repair and calibrate an instrument returned to an authorized Service Facility within 1 year of the original purchase; provided the warrantor's examination discloses to its satisfaction that the product was defective. The warrantor may, at its option, replace the product in lieu of repair. With regard to any instrument returned within 1 year of the original purchase, said repairs or replacement will be made without charge. If the failure has been caused by misuse, neglect, accident, or abnormal conditions of operations, repairs will be billed at a nominal cost. In such case, an estimate will be submitted before work is started, if requested.

THE FOREGOING WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY, FITNESS, OR ADEQUACY FOR ANY PARTICULAR PURPOSE OR USE. JOHN FLUKE MFG. CO., INC., SHALL NOT BE LIABLE FOR ANY SPECIAL, INCIDENTAL, OR CONSEQUENTIAL DAMAGES, WHETHER IN CONTRACT, TORT, OR OTHERWISE.

If any failure occurs, the following steps should be taken:

1. Notify the JOHN FLUKE MFG. CO., INC., or nearest Service facility, giving full details of the difficulty, and include the model number, type number, and serial number. On receipt of this information, service data, or shipping instructions will be forwarded to you.
2. On receipt of the shipping instructions, forward the instrument, transportation prepaid. Repairs will be made at the Service Facility and the instrument returned, transportation prepaid.

SHIPPING TO MANUFACTURER FOR REPAIR OR ADJUSTMENT

All shipments of JOHN FLUKE MFG. CO., INC., instruments should be made via United Parcel Service or "Best Way" prepaid. The instrument should be shipped in the original packing carton; or if it is not available, use any suitable container that is rigid and of adequate size. If a substitute container is used, the instrument should be wrapped in paper and surrounded with at least four inches of excelsior or similar shock-absorbing material.

CLAIM FOR DAMAGE IN SHIPMENT TO ORIGINAL PURCHASER

The instrument should be thoroughly inspected immediately upon original delivery to purchaser. All material in the container should be checked against the enclosed packing list. The manufacturer will not be responsible for shortages against the packing sheet unless notified immediately. If the instrument is damaged in any way, a claim should be filed with the carrier immediately. (To obtain a quotation to repair shipment damage, contact the nearest Fluke Technical Center.) Final claim and negotiations with the carrier must be completed by the customer.

The JOHN FLUKE MFG. CO., INC., will be happy to answer all applications or use questions, which will enhance your use of this instrument. Please address your requests or correspondence to: JOHN FLUKE MFG. CO., INC., P.O. BOX C9090, EVERETT, WASHINGTON 98206. ATTN: Sales Dept. For European Customers: Fluke (Holland) B.V., P.O. Box 5053, 5004 EB, Tilburg, The Netherlands.

*For European customers, Air Freight prepaid.

John Fluke Mfg. Co., Inc., P.O. Box C9090, Everett, Washington 98206

Table of Contents


SECTION	TITLE	PAGE
1	INTRODUCTION AND SPECIFICATIONS	1-1
	1-1. INTRODUCTION	1-1
	1-8. SPECIFICATIONS	1-1
2	OPERATION	2-1
	2-1. INTRODUCTION	2-1
	2-3. SHIPPING INFORMATION	2-1
	2-7. INSTALLATION	2-1
	2-9. FRONT AND REAR PANEL FEATURES	2-1
	2-11. PRE-OPERATION CONSIDERATIONS	2-1
	2-12. Introduction	2-1
	2-14. Input Power Requirements	2-1
	2-16. AC Line Voltage Selection	2-5
	2-18. Fuse Replacement	2-5
	2-20. Backup Operating Power	2-5
	2-22. Battery Charging	2-6
	2-25. IN CAL Indicator and RESET Terminal	2-6
	2-27. Portability	2-6
	2-29. Guarded Operation	2-6
	2-33. Oven Temperature Thermistor	2-7
	2-36. Minimizing Error Sources	2-8
	2-45. OPERATION	2-9
	2-47. OPERATION ON BATTERY	2-9
3	THEORY OF OPERATION	3-1
	3-1. INTRODUCTION	3-1
	3-3. OVERALL FUNCTIONAL DESCRIPTION	3-1
	3-8. CIRCUIT DESCRIPTION	3-1
	3-10. Power Supplies (A3 and A4)	3-1
	3-14. Voltage Monitor	3-1
	3-17. Reference Circuit, A5	3-3
	3-20. Output Divider	3-3
	3-22. Oven Controller	3-3

TABLE OF CONTENTS, *continued*

SECTION	TITLE	PAGE
3-25.	Battery Charger	3-3
3-29.	Overtemperature Protection	3-3
4	MAINTENANCE	4-1
4-1.	INTRODUCTION	4-1
4-5.	SERVICE INFORMATION	4-1
4-8.	GENERAL MAINTENANCE	4-1
4-9.	Access Procedure	4-1
4-23.	Cleaning	4-6
4-26.	Fuse Replacement	4-6
4-28.	AC Line Voltage Change	4-6
4-30.	ACCEPTANCE TEST	4-8
4-32.	CALIBRATION	4-8
4-35.	Null Verification Procedure	4-9
4-37.	Calibration Procedure	4-9
4-39.	SERVICE/REPAIR PROCEDURES	4-11
4-40.	Introduction	4-11
4-42.	+18.6V Adjustment	4-11
4-44.	Battery Charger Adjustment Procedure	4-11
4-46.	BATTERY CHARGING NOTES	4-13
4-47.	Battery Charge	4-13
4-50.	Battery Discharge	4-13
4-52.	Individual Battery Checkout	4-13
4-54.	TROUBLESHOOTING	4-13
4-55.	Introduction	4-13
4-57.	External Symptom Troubleshooting	4-13
4-59.	Internal Voltage Measurements	4-13
4-61.	Oven Repair	4-14
4-63.	Troubleshooting the A4 Regulator on the Bench	4-14
5	LIST OF REPLACEABLE PARTS	5-1
5-1.	INTRODUCTION	5-1
5-5.	HOW TO OBTAIN PARTS	5-2
6	ACCESSORIES AND OPTIONS	6-1
6-1.	INTRODUCTION	6-1
6-3.	DUAL MOUNTING FASTENER (M00-800-523)	6-1
6-5.	HALF-WIDTH RACK MOUNT KIT (M07-203-601)	6-1
6-7.	FULL-WIDTH RACK MOUNT KIT (M07-200-603)	6-1
6-9.	LOW THERMAL EMF CABLE SET (5440A-7002)	6-1
6-11.	BATTERY PACK (732A-7005)	6-1
6-13.	TRANSIT CASE (732A-7002)	6-1
6-15.	BATTERY CHARGER (732A-7003)	6-1
6-17.	RACK TRAY KIT (732A-7004)	6-1
6-19.	SPECIAL CALIBRATION (732A-000)	6-1
6-21.	SPECIAL CALIBRATION (732A-000R)	6-1


TABLE OF CONTENTS, *continued*

SECTION	TITLE	PAGE
6-23.	SPECIAL CALIBRATION AND DRIFT RATE CHARACTERIZATION (732A-100)	6-2
6-25.	SPECIAL CALIBRATION AND DRIFT RATE CHARACTERIZATION (732A-100R)	6-2
6-27.	DIRECT VOLTAGE MAINTENANCE PROGRAM (732A-200) ..	6-2
6-29.	ADDITIONAL 10V REFERENCE CERTIFICATION (732A-201)	6-2
7	GENERAL INFORMATION	7-1
7A	MANUAL CHANGE INFORMATION	7A-1
7A-1.	INTRODUCTION	7A-1
7A-3.	NEWER INSTRUMENTS	7A-1
8	SCHEMATIC DIAGRAMS	8-1
	TABLE OF CONTENTS	8-1
INDEX		




List of Tables

TABLE	TITLE	PAGE
1-1.	Accessories	1-1
1-2.	732A Specifications	1-2
2-1.	732A Front Panel Controls and Adjustments	2-4
2-2.	732A Rear Panel Features and Controls	2-4
4-1.	Required Test Equipment	4-2
4-2.	External Symptom Troubleshooting	4-15
4-3.	Internal Measurements	4-16



List of Illustrations

FIGURE	TITLE	PAGE
Frontispiece	732A DC Reference Standard	vi
1-1.	732A Outside Dimensions	1-4
2-1.	Front Panel Features	2-2
2-2.	Rear Panel Features	2-3
2-3.	Connector Plug for External Power Input Connector	2-5
2-4.	Guard Connection	2-7
2-5.	732A Precision Divider, Providing a Stable, Adjustable Source	2-8
3-1.	Functional Block Diagram	3-2
4-1.	Cover and Front Panel Screw Locations	4-3
4-2.	Rear Module Mounting Screw Locations	4-3
4-3.	Regulator PCA Removal	4-4
4-4.	Oven Assembly Removal	4-5
4-5.	AC Line Voltage Conversion on A3 Pre-Regulator PCA	4-7
4-6.	732A 10V Calibration	4-10
4-7.	Calibration of Point A to 10V Using 732A	4-10
4-8.	Calibration of 1.018 (and 1V) to 732A	4-10
4-9.	Battery Charger Test Points and Adjustments on A3 Pre-Regulator PCA ..	4-12
4-10.	A4 Regulator (Simplified Diagram)	4-16



732A DC Reference Standard

Section 1

Introduction and Specifications

1-1. INTRODUCTION

NOTE

Throughout this manual, all references to the 732A also pertain to the 732A/AN.

1-2. The Fluke Model 732A is a solid state, dc voltage reference standard. It is highly stable, rugged and transportable. The 732A has 10V, 1.018V and 1V outputs. These outputs are available on front panel binding posts. The calibration adjustments for the 10V, 1.018V, and 1V outputs are accessible through the front panel. A non-conducting adjustment tool is supplied with the unit for this purpose.

1-3. All outputs of the 732A can be shorted indefinitely without damage. Recovery occurs in less than 2 minutes after the short is removed, with no loss of stability.

1-4. The stability and accuracy of the 732A allow direct substitution for saturated standard cells in many applications. The stability specification of 0.5 ppm for 30 days is achieved by enclosing the reference amplifier and output divider of the 732A in a high gain thermal oven. Full accuracy is attained over the specified ambient temperature range of $23 \pm 5^\circ\text{C}$ (64.4 to 82.4°F). Variations in oven temperature may be monitored externally via the OVEN TEMP THERMISTOR terminals on the front panel.

1-5. The 732A may be powered from ac line power, an internal rechargeable battery, or an external low voltage ac or dc source. The 732A is designed to be powered continuously, including during storage or shipment. The backup battery will continue to operate the 732A for up to 24 hours. Either line or battery power may be removed without affecting the output. The battery is kept charged by an internal battery charger when operating from ac

line power, or from an external low voltage ac or dc source.

1-6. Various front panel LEDs (light-emitting diodes) provide a continuous indication of the operating status of the 732A. The AC PWR indicator illuminates in the presence of ac line power. The BTRY CHG indicator is on steadily for normal charging activity, and is off when the battery is charged. The IN CAL indicator monitors the input voltage to the Reference and Oven. Should this voltage fall below that needed to keep the 732A operational, the IN CAL indicator is latched off, indicating a loss of power and standardization. Once power is restored and standardization has been verified, the IN CAL indicator can be reset.

1-7. The 732A may be used on the bench or mounted in a rack. The 732A is a half-rack-width instrument and occupies four standard 1.75-inch rack spaces. The instrument's outside dimensions are given in Figure 1-1. Accessories for the 732A are listed in Table 1-1 and described in more detail in Section 6 of this manual. Calibration options are listed in Section 6.

1-8. SPECIFICATIONS

1-9. Table 1-2 lists the specifications for the 732A.

Table 1-1. Accessories

MODEL NUMBER	DESCRIPTION
M00-800-523	Dual Mounting Fastener
M07-203-601	Half Width Rack Mount Kit
M07-200-603	Full Width Rack Mount Kit
5440A-7002	Low Thermal EMF Test Lead Set
732A-7005	Battery Pack
732A-7002	Transit Case
732A-7003	Battery Charger
732A-7004	Rack Tray

Table 1-2. 732A Specifications

OUTPUT VOLTAGE 10 volts, 1.018 volts, and 1 volt

TRANSFER UNCERTAINTY @18°C to 28°C

Output Voltage	Time Interval			
	30 Days	90 Days	6 Months	1 Year
10V	0.5 ppm	1.0 ppm	1.5 ppm	3.0 ppm
1.018V	1.5 ppm	3.0 ppm	6.0 ppm	12.0 ppm
1V	1.5 ppm	3.0 ppm	6.0 ppm	12.0 ppm

These specifications assume the unit has been continuously powered up with either ac or battery or both. The specifications include effects due to line regulation.

TEMPERATURE COEFFICIENT OF OUTPUT

Range	Temperature Coefficient (ppm/°C)	
	0°C to 18°C	28°C to 40°C
10V	±0.05	±0.05
1.018V	±1.0	±1.0
1V	±1.0	±1.0

OUTPUT ADJUSTMENT AND RESOLUTION

Output	Adj. Range	Adj. Resolution
10V	±50 μ V	<0.05 ppm
1.018V	±50 μ V	<0.25 ppm
1.0V	±5 μ V	<0.10 ppm

OUTPUT IMPEDANCE

10V \leq 1.0 milliohms
 1.018V, 1V \approx 1 kilohm

OUTPUT CURRENT

10V 12 mA maximum
 1.018V, 1V Current limited by 1k Ω source impedance

OUTPUT PROTECTION The output may be shorted indefinitely without damage to the instrument. The instrument is protected against high voltage up to 1000V provided that the net current into the 732A does not exceed 30 mA.

OUTPUT NOISE \leq 1 μ V RMS at 10V output, 0.1-10 Hz.

COMMON MODE REJECTION For common mode voltages up to \pm 20 volts and frequencies up to 400 Hz, applied between ground and ground terminals, the common mode rejection is $>$ 140dB.

LOAD REGULATION AT

12.0mA OUTPUT CURRENT \leq 1.2 ppm

LINE REGULATION \leq 0.05 ppm of output for full \pm 10% power line variation.

Table 1-2. 732A Specifications (cont)

LINE POWER REQUIREMENTS 100, 120, 220, or 240V ac $\pm 10\%$ 50 to 60 Hz $\pm 5\%$, 32W

Nominal Setting	Voltage Limits	Fuse
100V	90-110V	0.375A/250V SLO-BLO
120V	108-132V	0.375A/250V SLO-BLO
220V	198-242V	0.250A/250V SLO-BLO
240V	216-264V	0.250A/250V SLO-BLO

AUXILIARY LOW VOLTAGE POWER

REQUIREMENTS 24-40V dc or 24-30V ac 50-400 Hz

INTERNAL BATTERIES 24V gelled-electrolyte sealed lead-acid

TYPICAL BATTERY LIFE 24 hours at 23°C

PROTECTION CLASS Class 1 as defined in IEC 348.

SIZE (HxWxD) 19.1 cm x 22.1 cm x 60.3 cm
7.5 in. x 8.5 in. x 23.7 in. (see Figure 1-1)

WEIGHT 12.3 kg (27 lbs.)

COMPLIANCE WITH EXTERNAL STANDARDS

- ANSI C39.5 Draft #8
- IEC 348 2nd edition, 1978
- CSA bulletin 556B, 17 Sep 1973
- VDE 0411-1973
- UL 1244

OPERATING TEMPERATURE 0°C to 40°C

ALTITUDE

- Non-operating 0-12,200 meters, (40,000 feet)
- Operating 0-3,050 meters, (10,000 feet)

TEMPERATURE AND HUMIDITY

Condition	Temperature (°C)	% Relative Humidity (Non-condensing)
Non-operating	-40 to +50 0 to 50	Not Controlled 95 $\pm 5\%$
Operating	0 to 30 30 to 40	95 $\pm 5\%$ 75 $\pm 5\%$

VIBRATION

Frequency	G Force Frequency	Double Amplitude
5-55 Hz	2 @ 55 Hz	0.013 inch





Figure 1-1. Outside Dimensions



static awareness



A Message From
John Fluke Mfg. Co., Inc.




Some semiconductors and custom IC's can be damaged by electrostatic discharge during handling. This notice explains how you can minimize the chances of destroying such devices by:

1. Knowing that there is a problem.
2. Learning the guidelines for handling them.
3. Using the procedures, and packaging and bench techniques that are recommended.


The Static Sensitive (S.S.) devices are identified in the Fluke technical manual parts list with the symbol




The following practices should be followed to minimize damage to S.S. devices.




1. MINIMIZE HANDLING




3. DISCHARGE PERSONAL STATIC BEFORE HANDLING DEVICES. USE A HIGH RESISTANCE GROUNDING WRIST STRAP.




2. KEEP PARTS IN ORIGINAL CONTAINERS UNTIL READY FOR USE.




4. HANDLE S.S. DEVICES BY THE BODY




5. USE STATIC SHIELDING CONTAINERS FOR HANDLING AND TRANSPORT




6. DO NOT SLIDE S.S. DEVICES OVER ANY SURFACE



7. AVOID PLASTIC, VINYL AND STYROFOAM® IN WORK AREA



8. WHEN REMOVING PLUG-IN ASSEMBLIES, HANDLE ONLY BY NON-CONDUCTIVE EDGES AND NEVER TOUCH OPEN EDGE CONNECTOR EXCEPT AT STATIC-FREE WORK STATION. PLACING SHORTING STRIPS ON EDGE CONNECTOR HELPS TO PROTECT INSTALLED SS DEVICES.



- 9. HANDLE S.S. DEVICES ONLY AT A STATIC-FREE WORK STATION
- 10. ONLY ANTI-STATIC TYPE SOLDER-SUCKERS SHOULD BE USED.
- 11. ONLY GROUNDED TIP SOLDERING IRONS SHOULD BE USED.

A complete line of static shielding bags and accessories is available from Fluke Parts Department, Telephone 800-526-4731 or write to:

JOHN FLUKE MFG. CO., INC.
 PARTS DEPT. M/S 86
 9028 EVERGREEN WAY
 EVERETT, WA 98204

PORTIONS REPRINTED
 WITH PERMISSION FROM TEKTRONIX, INC.
 AND GENERAL DYNAMICS, POMONA DIV.

Section 2 Operation

2-1. INTRODUCTION

2-2. This section describes the installation and operation of the Model 732A. The contents of this section should be read and understood before any attempt is made to operate the instrument. Should any difficulties arise during operation, contact your nearest John Fluke Sales or Service Representative. To determine the nearest sales or service center, refer to the list of sales and service centers in Section 7 of this manual. The factory may be contacted at:

John Fluke Mfg. Co., Inc.
P.O. Box C9090
Everett, WA 98206
Tel: (206) 347-6100

2-3. SHIPPING INFORMATION

2-4. The 732A is packaged and shipped in a foam-packed container. Upon receipt of the instrument, make a thorough physical and electrical inspection to reveal any possible shipping damage. Special instructions for inspection and claims are included on the shipping carton.

2-5. If reshipment of the instrument is necessary, the original container or equivalent should be used.

2-6. If the instrument is to be shipped with battery power on, use the Transit Case accessory described in Section 6. Alternatively, 24V to 40V dc or 24V to 30V ac may be applied, via the rear panel connector, to supply power during shipment.

NOTE

At an ambient temperature of 23° C, battery current drain is approximately 120 mA. Current drain will increase or decrease, respectively, with higher or lower ambient temperatures.

2-7. INSTALLATION

2-8. The 732A is designed for convenient operation as either a bench or a rack-mount instrument. Rack mounting accessories are described in Section 6.

2-9. FRONT AND REAR PANEL FEATURES

2-10. The Front and Rear panels are shown in Figures 2-1 and 2-2. The various controls and connections are listed and explained in Tables 2-1 and 2-2.

2-11. PRE-OPERATION CONSIDERATIONS

2-12. Introduction

2-13. The following paragraphs describe various conditions that should be considered before operating the 732A. If the 732A is brand new, set the rear panel BATTERY OPR switch to ON, and perform the acceptance test described in Section 4 of this manual.

2-14. Input Power Requirements

2-15. The 732A is designed to be powered continuously (including during storage or shipment) to maintain standardization. Normally, power is continuously applied, either to the ac line input connector or to the low voltage ac or dc input connectors. The ac line power requirements are: 100V, 120V, 220V or 240V ac $\pm 10\%$, at 50 or 60 Hz. Low voltage, 24-40V dc or 24-30V ac, 50-400 Hz may be connected to supply instrument power and charge the internal backup battery through the external input connector on the rear panel (see Table 2-2). The internal, rechargeable gel-cell (sealed, gelled electrolyte lead-acid) battery provides approximately 24 hours of continuous operation when ac power is not available.




Figure 2-1. Front Panel Features




Figure 2-2. Rear Panel Features

Table 2-1. 732A Front Panel Controls and Adjustments

ITEM	FEATURE NAME	DESCRIPTION
1	1.018V Adjustment*	Calibration tool adjustment. $\pm 50 \mu\text{V}$ adjust for 1.018V output.
2	1V Adjustment*	Calibration tool adjustment. $\pm 5 \mu\text{V}$ adjust for 1V output.
3	10V Adjustment*	Calibration tool adjustment. $\pm 50 \mu\text{V}$ adjust for 10V output.
4	AC PWR Indicator	LED that indicates the presence of ac power when illuminated.
5	BTRY CHG Indicator	LED that indicates battery charger operation when illuminated.
6	IN CAL Indicator**	LED that indicates out-of-calibration condition when not illuminated.
7	RESET**	Terminal behind front panel to reset the IN CAL indicator to ON condition.
8	GUARD Terminal	Binding post that connects to internal Guard circuit. Normally connected to OUTPUT LO at some point in the measurement system. 60V is the maximum differential allowed between GUARD and CHASSIS GROUND.
9	CHASSIS GROUND Terminal	Binding post connected to the chassis of the 732A.
10	1V HI Terminal	Binding post on which the 1V output of 732A is available.
11	LO Terminal	Binding post which provides common connection for the 10V output.
12	LO Terminal	Binding post which provides common connection for the 1V and 1.018V outputs.
13	10V HI Terminal	Binding post on which the 10V output of the 732A is available.
14	1.018V HI Terminal	Binding post on which the 1.018V output of the 732A is available.
15	OVEN TEMP THERMISTOR terminals	3/4-inch spaced dual binding posts. Floating thermistor for monitoring oven temperature.

*The 10V adjustment affects both the 1.018V and 1V outputs. This adjustment should be made first when calibrating the 732A. See Section 4.

**The IN CAL indicator detects an out-of-range condition within the power supply of the 732A. If not illuminated, the 732A is not operating at its specified accuracy. Use the RESET terminal to restore the IN CAL indicator after re-calibration. See Section 4.

Table 2-2. 732A Rear Panel Features and Controls

ITEM NO.	FEATURE NAME	DESCRIPTION
1	AC Module	Rear panel module containing the A3, Pre-Regulator PCB Assembly.
2	Battery Module	Rear panel module containing the A6A1, Battery PCB Assembly.
3	SUPPLY SETTING (Ac power requirements)	Specifies the correct ac line voltage required to operate the instrument.
4	 chassis terminal	Binding post that provides a direct chassis connection.
5	BATTERY OPR switch	Slide switch that sets instrument back-up battery supply, on or off.
6	EXTERNAL POWER INPUT connector	External power input connector for connecting an external power source (24-40V dc or 24-30V ac, 50-440 Hz). The internal back-up battery voltage may also be measured at this connector.

Table 2-2. Rear Panel Features (cont)

ITEM	FEATURE NAME	DESCRIPTION
7	Fuse holder	AC line fuse holder.
8	Power connector	IEC 3-wire receptacle, for ac line power connection. See Item 3 for specified ac line voltage.

2-15a. A connector plug (P/N 720847) for the external power input connector (Figure 2-3) is provided with each 732A Battery Module. To wire the plug, use the following procedure:

1. Remove the strain relief nut and the strain relief from the plug housing.
2. Push the contact header out of the plug housing in the direction of the strain relief.
3. Solder a wire to each contact using AWG 19- to 22-gauge wire or 2-conductor cable with an outside diameter smaller than .216 inches (see Figure 2-3).
4. Install the strain relief nut and the strain relief onto the cable in the correct order and orientation.
5. Insert one soldered contact into the contact header hole #1 (+) and the other into hole #2 (-).
6. Reassemble the plug.

2-16. AC Line Voltage Selection

2-17. The Line Voltage Selector switches are located inside the instrument. Their setting is marked on the rear panel (See Figure 2-2). If the marked setting does not agree with the locally available ac power, the settings of the internal Line Voltage Selector switches must be

changed. Refer this and all servicing to qualified personnel. The ac line voltage selection procedure is described in Section 4.

2-18. Fuse Replacement

2-19. The ac line fuse is located on the rear panel of the instrument. If the fuse requires replacement, replace it with one appropriate for the ac line voltage indicated on the rear panel. Use a 3/8A slow-blow fuse (P/N 109264) for ac line voltages from 100V to 120V. Use a 1/4A slow-blow fuse (P/N 166306) for ac line voltages from 220V to 240V.

2-20. Backup Operating Power

2-21. If ac line power fails or drops more than 30% below the nominal value, the internal battery automatically maintains operation. Set the rear panel BATTERY ÖPR switch to ON to enable the internal battery. When line power fails, the AC PWR indicator goes out, but the IN CAL indicator remains on. The unit continues to operate normally until the battery discharges. When the battery is discharged, the IN CAL indicator will go out. When ac power is restored, the BTRY CHG indicator illuminates until the battery is fully charged. If the IN CAL indicator does not illuminate, standardization of the instrument must be verified before the 732A is used. The battery voltage may be measured at the POWER INPUT connector with a high impedance multimeter, such as Multimeter A, specifications for which are given in Table 4-1.




Figure 2-3. Connector Plug for External Input Connector

2-22. Battery Charging**CAUTION**

Permanent battery damage will result if the battery is allowed to discharge below 19V. The degree of damage is a function of the depth of over-discharge and the battery temperature.

CAUTION

The unit must be in an upright position during charging to avoid possible venting of the electrolyte. The battery unit should be kept in an upright position at all times except during transit.

2-23. Under normal operation, battery life should exceed 5 years. For best battery life, minimize the number of charge/discharge cycles and avoid deep (<19V) discharge.

2-24. If the battery is fully discharged, 24 hours is required to recharge it fully when the 732A is operating from ac line power. If the battery is not fully discharged, the charging time will be less, but always in excess of the discharge time. When the battery is charging, the BTRY CHRG indicator illuminates to indicate charging activity; the indicator turns off when the battery is fully charged.

2-25. IN CAL Indicator and RESET Terminal

2-26. If the IN CAL indicator does not illuminate (indicating loss of ac power, a dead battery, or that the instrument is turned off), the output of the 732A may not meet the specifications listed in Section 1. The RESET terminal, located behind the front panel, is used to restore the IN CAL indicator to the on condition. See Section 4 for the reset procedures. Before resetting the indicator, apply power (ac line or low voltage external ac or dc), allow a stabilization period of 24 hours, then check the 732A to ensure that the various outputs are within specification. (See Section 4, Maintenance.)

2-27. Portability

2-28. The 732A is portable and operational at ambient temperatures between 0 and 40°C (32 to 104°F). Normal handling and transportation will not alter accuracy or stability if power is maintained by the internal battery or through the external power connections. The instrument may be used immediately after transportation, provided that the IN CAL indicator is illuminated and that the instrument has not been exposed to ambient temperatures beyond the normal operating range ($23 \pm 5^\circ\text{C}$). If this temperature range has been exceeded, allow one hour for temperature stabilization before using the instrument.

2-29. Guarded Operation**WARNING**

LETHAL VOLTAGES MAY BE PRESENT WHEN OPERATING THE 732A WITH THE GUARD AND CHASSIS GROUND CONNECTIONS SEPARATED. A MAXIMUM POTENTIAL DIFFERENCE OF 60V RMS MAY APPEAR BETWEEN ANY COMBINATION OF THE GUARD TERMINAL, CHASSIS GROUND, REFERENCE STANDARD OUTPUT, OR OVEN TEMPERATURE THERMISTOR OUTPUT. IF THIS LIMITATION IS EXCEEDED, DAMAGE TO THE INSTRUMENT MAY RESULT.

2-30. The 732A is equipped with a guard that isolates the internal circuitry from the chassis and earth ground. A GUARD terminal is provided on the front panel. When the guard is properly used, it can greatly reduce errors caused by common mode voltages. In general, guarded operation is necessary under the following conditions:

1. When a potential exists between equipment power line grounds.
2. When long connection leads are used to connect a high impedance load.
3. When the instrument is operating in the presence of high level radiated noise.

2-31. A potential difference may exist between the power line grounds of the 732A and an instrument to which it is connected. This potential difference can cause circulating ground currents, which cause errors in the output voltage.

2-32. To prevent these errors, connect the 732A GUARD terminal to the load in a manner that provides a separate path for the circulating currents. Connect the GUARD terminal to the grounded side of the load, at the load. Figure 2-4 illustrates the correct GUARD terminal connection and the rerouted ground currents. The circulating current path may also be broken by operating the 732A on battery power as described later in this section.

2-32a. When using the 732A to make transfer measurements involving standard cells, take care to avoid coupling alternating currents into the circuit which may be rectified by the standard cell. Otherwise, erroneous results can be produced. Observe the following precautions:

1. Use low-thermal emf, shielded cables for all interconnections between test instruments.

2. Use only test instruments that are isolated and guarded from power line ground. If necessary, use battery powered instruments to achieve this isolation.
3. Connect the test lead shields to the instrument guard terminals.
4. At one point, and ONE POINT ONLY, connect the test circuit low to the ground circuit, and connect the ground circuit to power line and chassis ground.

2-32b. Careful attention to these precautions usually eliminates the effects of capacitive coupling, which is present in most line-operated instruments, between the ac power line and the instrument circuits. If the effectiveness of these measures is in doubt, the 732A (and possibly other instruments) may be disconnected from the power lines and operated on battery power to determine if the test results change. Line coupling may come from unexpected sources. For example: standard cell enclosures have line-operated heater circuits; potentiometers have line-operated power sources; galvanometers have line-operated illumination sources.

2-33. Oven Temperature Thermistor

2-34. A thermistor, mounted inside the Oven Assembly, senses changes in the internal oven temperature. Use the Oven Temperature Thermistor in conjunction with an external ohmmeter to monitor the temperature stability of the oven vs time. The thermistor terminals are on the front panel of the 732A. Both leads of the thermistor are isolated from all parts of the 732A circuitry. A maximum potential difference of 60V is allowed between either of the thermistor terminals and any other front panel terminal (Guard, Ground, Reference Standard Outputs).

2-35. The nominal value of the Oven Temperature Thermistor is between 4 k Ω and 5 k Ω at the normal oven operating temperature. The thermistor has a temperature coefficient of minus 3.8%/°C. The actual operating value is shipped with the instrument and should be recorded in the instrument data log book as soon as possible.

2-35a. In order to ensure optimum performance of the 732A, the value of the oven temperature thermistor should be recorded to three significant digits whenever the 732A is used (or on a daily basis when the instrument is not in use). The thermistor drift rate is normally ± 50




Figure 2-4. Guard Connection

ohms per year in the first year and ± 10 ohms per year thereafter. A long-term temperature shift of 5% in thermistor resistance would not be excessive (in terms of actual temperature drift of the oven). On a short-term basis, variations of ± 20 ohms from day-to-day indicate probable oven problems. Short-term changes of ± 100 ohms would affect the 1.0 and 1.018V outputs appreciably, but would be unlikely to have a significant effect on the 10V output.

2-36. Minimizing Error Sources

2-37. The inherent accuracy and stability of the 732A may be easily degraded if the effects of thermal emf, lead resistance, and other factors are not considered and minimized.

2-38. THERMAL ERRORS

2-39. When parts of a circuit operate at different temperatures, thermal voltages are normally present at the equipment connections. These thermal voltages can exceed $10 \mu\text{V}$. Use the following techniques to minimize thermal errors:

1. Use the Fluke 5440A-7002 Low Thermal EMF Cable Set. See Section 6, Accessories.
2. Use #24 AWG or larger, bare copper, Teflon[®]-insulated connecting wires. It is preferable to use shielded, twisted pair cable. Avoid splices.

3. Avoid the use of ordinary, nickel-plated, banana plugs for equipment interconnections. Use of low-thermal emf, copper spade, lugs is recommended. Crimp the lug onto the wire before soldering. Loosen the top of the binding post, insert the lug, and tighten the binding post on the lug.

NOTE

Tighten the binding posts with fingers only.

2-40. OTHER ERROR SOURCES

2-41. When using the 732A with a reference divider, the effects of the finite (though very low) output impedance of the 732A, the lead wire resistance, and the loading caused by the reference divider cannot be ignored. Use the following procedure to minimize the effects of test lead resistance and output loading:

1. Connect the equipment as shown in Figure 2-5.
2. Calibrate the 732A/Reference Divider combination at the divider input terminals.
3. When calibration is complete, treat the 732A and the Reference Divider as a system.




Figure 2-5. 732A/Precision Divider, Providing a Stable Adjustable Source

4. Do not disconnect the Reference Divider at any time, even if it is not required in a given procedure. Since the 732A/Reference Divider system was calibrated at the divider input terminals, disconnecting it will change the loading on the 732A and affect the calibration.

2-42. LONG-TERM STABILITY

2-43. The user can determine the long-term stability of the 732A by periodically comparing the output voltage to a known reference standard and plotting the difference as a control chart. Output voltage drift, with respect to the legal voltage, may be positive or negative and is a characteristic of the individual 732A. Accumulated test data have shown that, once established, the measured drift is generally constant, provided the instrument continues to receive uninterrupted operating power. When drift rate is established, extrapolations of output voltage are justified, and allow the use of the reference with less uncertainty than would be obtained from the stability specifications alone. A convenient method of performing the needed periodic calibration is to use the Fluke Direct Voltage Measurement Program, Option 732A-200 (described in the Fluke Catalog), which provides calibration against a traceable reference standard in the user's lab.

2-44. Due to the possibility of undetected damage or malfunction, there is a risk involved in relying completely on any single standard. This risk is reduced significantly by using two or more independent standards (with frequent intercomparisons) to ensure that all are stable within the range of expected normal deviations. An added advantage of having multiple, independent standards is the opportunity to use a statistical average of the individual outputs. Multiple 732A's may be used in this way to gain statistical improvements in the uncertainty specification of the reference voltage.

2-45. OPERATION

2-46. Use the following procedure to prepare the 732A for operation.

1. Check the rear panel for ac power requirements and connect the 732A power cord to an appropriate power source.


2. Set the BATTERY OPR switch to ON.
3. Verify that the AC PWR indicator is illuminated. The BTRY CHG indicator is also illuminated if the batteries are not fully charged.
4. Allow the unit to stabilize for a period of 24 hours if either of the following applies:
 - a. The IN CAL indicator is not illuminated.
 - b. The instrument has been stored in or exposed to ambient temperatures in excess of the normal operating range ($23 \pm 5^\circ\text{C}$).
5. Ensure that the 732A is calibrated according to the procedures described in Section 4.
6. The instrument is now ready for use.
7. If the IN CAL indicator goes out, the output of the 732A is not standardized. Notify the Calibration department or person(s) responsible for maintaining the 732A. Refer to Section 4.

2-47. OPERATION ON BATTERY

2-48. Prior to operating the 732A on battery power, the following procedure should be followed.

1. Connect a voltmeter at the rear terminal POWER INPUT jack to monitor the battery voltage.
2. Note the voltage while the battery is on trickle charge (the AC PWR LED is on, but BTRY CHG off). The voltage should be around 27V.
3. Disconnect the line cord, wait 10 minutes, and monitor the battery voltage for one-half hour. The battery package may have to be replaced if the voltage decreases by more than 0.2V during this period and the load represented by the 732A is correct (see Section 4, Battery Discharge).

2-49. It may also be advisable to perform the above procedure periodically to ensure battery backup capability is maintained.



Section 3

Theory of Operation

3-1. INTRODUCTION

3-2. This section describes the theory of operation of the 732A, including an overall functional description, followed by a circuit description of the 732A. Both descriptions are supported by a block diagram (Figure 3-1). Component-level descriptions contained in the circuit analysis refer to the detailed schematics in Section 8 of this manual.

3-3. OVERALL FUNCTIONAL DESCRIPTION

3-4. The 732A dc Voltage Reference Standard is a highly stable 10V, 12 mA power supply. Refer to Figure 3-1. AC line input power is full wave rectified and fed to a two-stage voltage regulator. The first stage, or Pre-regulator, converts the raw dc to 32V dc. The second stage, or Regulator, converts this voltage to 18.6V dc, which powers the Oven Controller and the Reference.

3-5. The Voltage Monitor disables the Oven Controller and latches the IN CAL indicator off when the output of the Regulator is insufficient for proper operation. The RESET terminal is used to restore the IN CAL indicator to the on condition after standardization of the instrument has been performed.

3-6. If ac line power fails or is not available, an internal, sealed, lead-acid battery maintains operating power to the 732A. When ac power is available, a battery charger charges the battery. While the battery is charging, the BTRY CHG indicator is on; when the battery is fully charged, the BTRY CHG goes out.

3-7. When ac line power is not available, the battery may be charged by a low voltage external ac or dc source

connected at the rear panel POWER INPUT connector. The external source can also supply operating power for the instrument. The battery voltage can also be measured at the rear panel connectors.

3-8. CIRCUIT DESCRIPTION

3-9. The information in this section describes the circuitry of the 732A to the functional block diagram level. Refer to the detailed schematics in Section 8.

3-10. Power Supplies (A3 and A4)

3-11. The 732A has two cascaded regulators. The Pre-regulator (A3Q1) is a simple emitter follower regulator that clamps the full wave rectified power from the bridge rectifier to approximately 32V dc.

3-12. The Regulator (located on A4) supplies operating voltages to all of the circuitry in the 732A except the battery charger. During battery operation, the battery drives the Regulator input.

3-13. The Regulator (Q1, Q2, Q3, Q4) is a conventional series-pass transistor, error-amplifier design that regulates the 32V to 18.6V dc.

3-14. Voltage Monitor

3-15. The Voltage Monitor circuit (Q5, Q6, Q7, Q8) checks the regulator output and disables the instrument when the supply voltage falls below a critical value. When this happens, the Oven Controller is disabled and the IN CAL indicator is latched off. The reset circuit is used to turn the IN CAL indicator back on after standardization has been verified by qualified personnel. The Voltage Monitor is located on the A4 Regulator PCA (printed circuit assembly).




Figure 3-1. Functional Block Diagram

3-16. Transistor Q8 is turned on by the voltage drop across the Regulator circuit series-pass transistor. This causes switching transistor Q5 to saturate, supplying power to the Oven Controller circuit and the IN CAL indicator circuit. When the output falls below that needed for normal operation, Q8 and Q5 turn off, shutting down the Oven Controller and removing drive from Q7, a Programmable Unijunction Transistor (PUT). This removes the drive from Q6, shutting off the IN CAL indicator on the front panel. When power is restored, Q7 remains latched off until its emitter is connected momentarily to the COMMON output terminal via the RESET connection, which is accessible through the front panel.

3-17. Reference Circuit, A5

3-18. The Reference Circuit (A4Q12, Q1, Q2, Q5, U1, U2) reduces the 18.6V output of the Regulator to precisely 10V. The Reference Circuit is a highly stable series-pass voltage regulator. The entire reference supply (except the pass transistor) is enclosed in an oven to provide the consistent thermal environment necessary for the stability of the output.

3-19. U2, the Ref-Amp, is a transistor and zener diode mounted on a common substrate. This construction compensates for ambient temperature changes; thus U2 has an extremely low temperature coefficient. The Ref-Amp compares the 10V output to its internal zener reference to derive an error voltage that is amplified by op amp U1. U1 drives the series-pass element (Q1, A4Q12). Q2 provides current limiting to protect the series-pass element under short-circuit conditions. Variable resistor R20 allows a small adjustment ($\pm 50 \mu\text{V}$) in the output voltage of the Reference. Larger adjustments can be made by jumper changes on the Calibration PCA, A7.

3-20. Output Divider

3-21. Two precision resistive voltage dividers divide the precise 10V output down to 1V and 1.018V. Each of these dividers is adjustable over a limited range to allow calibration. Both dividers are enclosed in the oven with the reference.

3-22. Oven Controller

3-23. The Oven Controller (A4Q13, A4Q14, Q3, Q4, U3, U4, U5) maintains the internal temperature of the oven at a nominal temperature of $45 \pm 2^\circ\text{C}$. The Oven Controller is a high gain thermal proportional control

circuit. The Oven Controller circuit is partially located on the A5 Reference PCA, inside the oven. The oven driver and output transistors are located on the A4 Regulator PCA.

3-24. Thermistors RT1, series connected RT2, and RT3 are connected in a bridge configuration with R28 and R29, and are located inside the oven. U3 buffers the bridge output and drives differential amplifier/integrator U5, which drives the driver (A4Q13) and output transistors (A8Q1, A8Q2) and subsequently the oven heaters. U4 shapes the overall loop frequency response.

3-25. Battery Charger


3-26. The battery charger determines the state of charge of the internal battery and sets the charging current accordingly: constant current charging for deep discharge or constant voltage trickle charging for charge maintenance. The Battery Charger circuit is located on A3.

3-27. Transistor Q2 is a current source that supplies all the charging current. Transistors Q3 and Q4 form a Schmitt trigger. Transistor Q6 supplies a constant voltage output for trickle charging and thus maintains the battery at full charge. Two thermistors monitor the battery temperature (A6RT1, A6RT2) and adjust the charging rate accordingly.

3-28. During initial charging, Q3 enables Q2 and the high charge rate. When the battery voltage rises to approximately 32V, Q4 turns off, shutting off the constant current charge. The battery is then constant voltage charged by Q6 (approximately 27V at 23°C). Potentiometer R10 sets the threshold point of this transition and hence the end of charge current. At this point, Q6 supplies a constant voltage trickle charge to the battery, and R20 sets this voltage level. Thermistors RT1 and RT2, located on the battery pca, and Q5 prevent high current charging at temperatures below 5°C , and/or high temperatures.

3-29. Overtemperature Protection

3-30. Protection against overtemperature is provided by a 58°C thermal fuse (F1), which is placed directly on top of the ovenized reference module. When F1 opens, it interrupts the +18.6V supply connection to the oven heater, and also disconnects the +18.6V from the latch circuit Q6 on the regulator PCA, causing the IN CAL LED to go out (and stay out) until the fault is corrected.



Section 4

Maintenance

WARNING

THESE SERVICING INSTRUCTIONS ARE FOR USE BY QUALIFIED PERSONNEL ONLY. TO AVOID ELECTRICAL SHOCK, DO NOT PERFORM ANY SERVICING OTHER THAN THAT CONTAINED IN THE OPERATING INSTRUCTIONS UNLESS YOU ARE QUALIFIED TO DO SO.

4-1. INTRODUCTION

4-2. This section provides maintenance information for the 732A. It includes general maintenance procedures, an acceptance test, a calibration test, calibration procedures, and troubleshooting information.

4-3. The acceptance test is used to verify that the instrument is operating within specifications. Perform the acceptance test upon receipt of the instrument.

4-4. The instrument should be calibrated at intervals commensurate with the accuracy and stability requirements of the user. Necessary test equipment is listed in Table 4-1. Equivalent instruments may be used, provided that they meet the minimum specification(s).

NOTE

To limit thermally induced errors, use the Fluke Low Thermal EMF Cable Set (accessory 5440 A-7002). Alternatively, copper wire, (preferably shielded twisted pair), with crimped and soldered low-thermal lugs may be used. Avoid the use of ordinary nickel-plated banana plugs.

CAUTION

To avoid cracking the plastic binding post insulator, tighten only with finger pressure. Do not use tools.

4-5. SERVICE INFORMATION

4-6. The 732A is warranted for a period of one (1) year upon delivery to the original purchaser. The warranty is given on the back of the title page located in the front of this manual.

4-7. Factory-authorized calibration and service for each Fluke product is available at various worldwide locations. A complete list of Fluke service centers is included in Section 7 of this manual. Shipping information is given in Section 2 of this manual. If requested, an estimate will be provided to the customer before any repair work is begun on instruments that are not currently under warranty.

4-8. GENERAL MAINTENANCE

4-9. Access Procedure

4-10. Use the following procedures to disassemble the 732A for adjustment or repair. Disconnect ac power connections before disassembling the 732A.

Table 4-1. Required Test Equipment

TYPE	REQUIRED SPECIFICATIONS	RECOMMENDED MODEL	PROCEDURE*
Certified 732A	As required by the user	Fluke 732A**	A, B
Voltage Divider	7 decade, 0.1 ppm resolution 0.1 ppm absolute linearity	Fluke 720A	B
Null Detector	1 μ V full-scale sensitivity. 10 M Ω input resistance. ZERO/OPR switch must open circuit input terminals in ZERO position.	Fluke 845AB, AR	B
Adjustable Source	10V dc output 1 μ V resolution 0.3 ppm + 2 μ V stability	Fluke 5440A, 5440B or 5442A	B
Multimeter A	4½-digit display 20 k Ω resistance range 200 mV to 200V ac or dc	Fluke 8050A, 8060A	A,C,D
Multimeter B	6½-digit display 10V dc range, 100 μ V resolution 1V dc range, 10 μ V resolution	Fluke 8500A, 8502A	B,D
Rheostat	50 k Ω , ½W	Fluke P/N 501601	C
Variac	120V, 1A, metered	GenRad W5MT3A	C
Load Resistor	1 k Ω , ½W Carbon Composition	Fluke P/N 108597	B,C
Adjustment Tool	Supplied with 732A	Fluke P/N 686113	A,B

* A = Acceptance Test
B = Calibration procedure
C = Battery charger adjustment
D = Troubleshooting

**The 732A selected for use as the Certified 732A in the Calibration Procedure should be calibrated at a calibration facility whose transfer uncertainties are consistent with the user's needs.

4-11. COVER REMOVAL

4-12. Use the following procedure to access the interior of the 732A. (Refer to Figure 4-1.)

1. Remove all screws securing the top and/or bottom cover(s).
2. Lift the cover(s) off the instrument.

4-13. REAR MODULE REMOVAL

4-14. The AC Module and the Battery Module are located in the rear of the 732A. Use the following procedure to remove either of the rear modules. (Refer to Figure 4-2.)

NOTE

Either module, but NOT both, may be removed without loss of standardization. If the AC Module is removed, ensure that the rear panel, BATTERY OPR switch is set to ON and that the battery is charged before removing the AC Module. This will ensure continued standardization.

1. Remove the screws securing the module to the rear of the instrument.
2. Pull the module out from the rear of the instrument.




Figure 4-1. Cover and Front Panel Screw Locations




Figure 4-2. Rear Module Mounting Screw Locations

4-15. REGULATOR PCA (PRINTED CIRCUIT ASSEMBLY) REMOVAL

NOTE

Since the Regulator PCA removal requires the removal of BOTH rear modules, standardization will not be maintained after this procedure.

4-16. Use the following procedure to remove the Regulator PCA from the 732A. (Refer to Figure 4-3.)


1. Remove the top and bottom covers.

2. With the 732A resting on its bottom, remove the screws securing the inner shield top cover and remove the shield.
3. Remove both of the rear modules.
4. Remove the screws that fasten the T0-220 power transistor to the bottom of the chassis. Save the insulator and the shoulder washer. Note the positions of the insulating hardware so they can be reassembled properly.
5. Unplug the Regulator PCA from the motherboard by pulling it out towards the rear of the 732A.

4-17. OVEN REMOVAL

4-18. Use the following procedure to remove the Oven Assembly from the 732A. (Refer to Figure 4-4.)

1. Remove the top and bottom covers.
2. With the 732A resting on its bottom, remove the screws securing the inner shield cover and remove the cover.
3. Carefully pry out the top foam insulating block from the front of the instrument using a blade type screwdriver.



*CAUTION: DO NOT LOSE INSULATING WASHER
AND SHOULDER WASHER

REMOVE SCREW*

REGULATOR PCB ASSEMBLY

Figure 4-3. Regulator PCA Removal




Figure 4-4. Oven Assembly Removal

4. Do the same for the foam block that is now exposed.
5. Locate the two mylar tabs on each side of the Oven Assembly.
6. Grasp both mylar tabs and pull them steadily and evenly upwards.
7. Disconnect the Oven Assembly cable harness at the motherboard and at the front panel.

4-19. OVEN DISASSEMBLY

4-20. Use the following procedure to disassemble the Oven Assembly only if access is necessary to make repairs on the Oven Controller circuit. Do not attempt to repair the Reference circuit.

1. Remove the Oven Assembly from the 732A.
2. Lay the instrument on its side, with its top facing you, and lay the Oven Assembly on the work surface.
3. Remove the four screws holding the inside clamshell. (The inside clamshell contains the adjustment holes for the calibration potentiometers.)

NOTE

Do not turn the screws on the outside clamshell, as this will cause difficult disassembly and reassembly.

4. Move the wire bundle to the side and lift the heater assembly free of the Oven Assembly.
5. Lay the heater assembly to the side. The Reference PCA circuitry is now accessible.

NOTE

In most cases, repairs to the pca can be more easily performed from the component side of the pca. If access to the bottom of the pca is necessary, unscrew the outside four Teflon standoffs.

4-21 FRONT PANEL REMOVAL

4-22. Use the following procedure to detach the front panel from the 732A:

1. Remove the front screws of the top and bottom cover.

2. Peel the decal from both of the front corner side moldings and remove the exposed screws. Refer to Figure 4-1 for screw locations.
3. Remove the front corner side moldings from the instrument.
4. The front panel is now free. Be extremely careful of the wire harness connected to the front panel binding posts. The service loop provided is quite limited.

4-23. Cleaning

CAUTION

To prevent possible damage to the front panel, do not use aromatic hydrocarbons or chlorinated solvents on the front panel of the 732A.

4-24. When the 732A is properly cared for and kept in a controlled atmosphere, cleaning is seldom required. However, any contamination in the instrument, particularly oil, can contribute to an increase in leakage, which may impair accuracy.

4-25. Clean the exterior and the front panel of the 732A with a soft cloth dampened in a mild solution of detergent and water. Do not attempt to clean the interior of the instrument.

4-26. Fuse Replacement

4-27. The power fuse, F1, is located on the rear panel of the 732A. If replacement is necessary, use the following rated fuses:

100V or 120V ac operation — MDL $\frac{3}{8}$ ($\frac{3}{8}$ A slow-blow, Fluke P/N 109264)

230V or 240V ac operation — MDL $\frac{1}{4}$ ($\frac{1}{4}$ A slow blow, Fluke P/N 166306)

4-28. AC Line Voltage Change

4-29. The 732A may be operated from 100V, 120V, 220V, or 240V ac $\pm 10\%$. The assigned line voltage may be changed to match the available source using the following procedure. Refer to Figure 4-5.

1. Ensure that the battery is charged or that an appropriate external ac or dc source is connected to the POWER INPUT jack on the rear panel. This will maintain the unit's standardization when ac line power is removed. The BTRY CHG indicator on the front panel will extinguish when the battery is fully charged and the 732A is still connected to the ac power source.




Figure 4-5. AC Line Voltage Conversion on A3 Pre-Regulator PCA

2. Set the BATTERY OPR switch to ON and remove ac line power from the instrument.
3. Remove the AC Module.
4. Locate the voltage selector switches (the slide switches on the top of pca near the rear panel). Set the switches so that the dots on the switch actuators select the correct line voltage, as shown in Figure 4-5.
5. Reinsert the AC Module; replace the screws.
6. On the rear panel, change the mark to the appropriate box (under the SUPPLY/SETTING heading) to indicate the present power configuration.
7. Replace the line fuse with one of the appropriate value.
8. After verifying that the local ac line voltage matches the voltage selected on the 732A, apply ac line power to the instrument.

4-30. ACCEPTANCE TEST

4-31. Use the following procedure to verify that the instrument is operational. The required test equipment is listed in Table 4-1. Equivalent instruments may be used, provided the minimum specifications are met.

1. Check the IN CAL indicator on the front panel. If the indicator is illuminated, proceed to step 2. If the indicator is not illuminated, complete steps a through g.
 - a. If the IN CAL indicator was not on, set the rear panel BATTERY PWR switch to OFF. Apply ac power to the instrument using the Variac to the Supply Setting limit listed on the rear panel.
 - b. Adjust the Variac for 120V ac output. The ac line current should be less than 0.3A.
 - c. Set the BATTERY PWR switch to ON. The ac line current should be less than 0.35A. If the battery is very low, the BTRY CHG indicator will blink.
 - d. Allow the 732A to stabilize (under power) for 24 hours.
 - e. Measure the value of the Oven Temperature Thermistor at the front panel binding posts

with Multimeter A. The value should be within ± 20 ohms of the value shipped with the instrument.

- f. Perform or obtain calibration as described in the paragraphs following the heading, CALIBRATION, below.
 - g. Once the 732A has been calibrated, proceed to step 2.
2. Apply ac power of the correct voltage and frequency to the instrument. The AC PWR and BTRY CHG indicators should both be on.
 3. Repeat step 1(e).
 4. Check the output voltage at the 10V output using Multimeter B. It should be accurate within the performance limitations of the Multimeter.
 5. Measure load regulation. Connect the 732A under test as shown in figure 4-6. Use copper hook-up wire under the binding posts, leaving the connector socket free. Measure the null voltage, then insert the leads from a 1000 Ohm load resistor into the connector sockets and determine the shift in the null voltage. The voltage change should be less than $10 \mu\text{V}$ or 1.0 ppm.
 6. Repeat step 4 for the 1V and 1.018V outputs.
 7. If a standards laboratory is available, verify stability by comparison to another precertified 732A. This step is optional.
 8. The instrument is operational.

4-32. CALIBRATION

4-33. Complete the following calibration procedure to certify the 732A. The procedure uses direct comparison between the Unit Under Test (UUT) and a Certified 732A to calibrate the 10V output. The 10V output of the UUT is then transferred to a stable adjustable voltage source. The voltage source is then divided down, as required, for comparison with the UUT 1.018V and 1V outputs.

4-34. The 732A is normally shipped from the factory with the BATTERY OPR switch OFF. Upon receipt, the instrument must be powered up and stabilized for a minimum of 24 hours before calibration against traceable standards. To maintain calibration as a traceable standard, the 732A must continue to receive uninterrupted operating voltage from line power or from the

self-contained batteries, which provide approximately 24 hours of off-line operation when fully charged. If operating power is interrupted, the IN CAL light on the instrument front panel will be extinguished and recalibration will be necessary.

4-34a. The 732A may be calibrated in several ways. For calibration by the Fluke Standards Laboratory and shipment under power, refer to Options 732A-000 and 732A-100 described under Direct Voltage Maintenance Program in the Fluke catalog. For calibration in the user's lab, the Direct Voltage Maintenance Program (Option 732A-200) also provides for shipment of a Fluke-owned certified 732A, together with all necessary connecting cables and operating instructions, for on-site comparison. Calibration may be obtained directly from the National Bureau of Standards in Gaithersburg, Md., providing the necessary arrangements are made to have the 732A transported back to the user's lab under power. The Bureau of Standards should be contacted directly for details on this special calibration service (listed as 3.4H, Solid State Voltage Ref. Devices Calibration). Calibration by comparison with another certified 732A is described in detail in the following paragraphs. Other methods, using standard cells for reference, may be considered. However, these methods are not recommended due to the difficulty in obtaining the transfer from a standard cell voltage of 1.018 up to 10V without introducing significant errors from voltage dividers and thermal emfs.

4-35. Null Verification Procedure

4-36. Use the following procedure to verify the accuracy of null in the calibration procedure. The Null Verification procedure identifies the thermal voltages present and allows the null adjustment to be made independently of them. Use the Null Verification procedure when instructed to "verify the null" in the Calibration Procedure.

1. Adjust the UUT for zero on the Null Detector.
2. Reverse the HI and LO (positive and negative) leads on the UUT and the certified 732A.
3. Observe the Null Detector reading. If the reading does not equal zero, adjust the UUT for one-half of the Null Detector reading.
4. Reverse the HI and LO (positive and negative) leads on the UUT and the Certified 732A. The Null Detector should have the same reading as it did at the end of step 3. If it does not, adjust the UUT for one-half the difference.

5. Repeat steps 2 through 4 until the Null reading does not change when the UUT and Certified 732A leads are reversed.
6. The residual reading on the Null Detector equals the sum of the thermal voltages in the circuit.

4-37. Calibration Procedure

4-38. Complete the following procedure to standardize the outputs of the 732A to a Certified 732A. Battery operation of the 732A and 845AB/AR is preferred. Set the Null Detector to ZERO when changing leads. Use the supplied adjustment tool for all adjustments (Fluke P/N 686113).

1. Obtain a certified 732A.
2. Perform the self-calibration procedure on the Precision Divider immediately prior to this procedure.
3. Connect the UUT and the Certified 732A as shown in Figure 4-6.
4. Set the OPR switch on the Null Detector to the ZERO position. Adjust the Null Detector for zero on the $3 \mu\text{V}$ range.
5. Set the Null Detector to the $30 \mu\text{V}$ range and the OPR switch to OPR.
6. Decrease the range setting on the Null Detector slowly, while simultaneously adjusting the 10V calibration potentiometer (through the front panel opening on the UUT) for a null indication on the Null Detector on the $3 \mu\text{V}$ range. Let the system stabilize for about 1 minute before final adjustment. Use the non-conducting adjustment tool supplied with instrument.
7. Verify the null, using the Null Verification Procedure.
8. Connect the equipment as shown in Figure 4-7. Set the Precision Divider ratio switches to 0.999999X.
9. Adjust the Adjustable Source for a null indication.
10. Verify the null.
11. Connect the equipment as shown in Figure 4-8. Set the Precision Divider ratio switches to 0.1018000.




Figure 4-6. 732A 10V Calibration




Figure 4-7. Calibration of Point A to 10V Using 732A




Figure 4-8. Calibration of 1.018V (and 1V) to 732A

12. Set the Null Detector RANGE switch to the 30 μ V volt. Switch the Null Detector to OPR.
13. Adjust the 1.018V calibration potentiometer on the UUT while decreasing the RANGE setting on the Null Detector to obtain a null on the 3 μ V range. Use the non-conducting adjustment tool.
14. Verify the null.
15. Transfer the Null Detector input lead from the 1.018V terminal to the 1V terminal on the UUT.
16. Set the Precision Divider ratio switches to 0.1000000.
17. Set the RANGE control on the Null Detector to the 30 μ V position. Adjust the 1V calibration potentiometer on the UUT while decreasing the RANGE setting on the Null Detector to obtain a null on the 3 μ V range. Use the non-conducting adjustment tool.
18. Verify the null.
19. If the IN CAL indicator is illuminated, proceed to step 20. If the indicator is not illuminated, connect a short wire to one of the front panel COMMON terminals. touch the other end of this wire to the circuit board behind the RESET hole. The IN CAL indicator should illuminate.
20. Calibration is complete. Record all test results. Disconnect all test equipment. Cover the output adjustment access holes and the RESET hole with tamper-proof calibration seals.

4-39. SERVICE/REPAIR PROCEDURES

4-40. Introduction

4-41. The Battery Charger Adjustment procedure is the only field service procedure for the 732A. There is no field serviceable circuitry within the oven/reference supply assembly. All adjustments within the oven must be made at the Factory or at a Fluke Technical Service Center. The following paragraphs describe the Battery Charger adjustments for the 732A.

4-42. +18.6V Adjustment

4-43. To perform the +18.6V adjustment, perform the following:

1. Connect a dc voltmeter to TP2 (LOW) and TP1 (HI) on the A4 Regulator PCA.

2. Adjust R22 for 18.6V as indicated on the dc meter.

4-44. Battery Charger Adjustment Procedure

CAUTION

This procedure will cause loss of standardization. Calibration must be performed before reusing the instrument.

4-45. Refer to Figure 4-9. Perform this procedure to calibrate the battery charger after repair of the battery charger circuit. The equipment required is listed in Table 4-1.

1. Remove ac power from the instrument.
2. Set the BATTERY OPR switch to OFF.
3. Remove the top cover from the instrument.
4. Remove the AC Module from the instrument.
5. Locate test points TP1, TP2, and TP5 on the A3, Pre-Regulator PCA (part of the AC Module). Locate trimpots R20 and R10 and jumper wire W1, also on the AC Module.
6. Connect a 50 k Ω rheostat between TP1 and TP2. Adjust the Rheostat for maximum resistance.
7. Connect Multimeter A between TP5 and TP1. TP5 is positive with respect to TP1.
8. Reinstall the AC Module in the instrument.
9. Apply ac power to the UUT using the Variac. Adjust the Variac for the line voltage indicated on the rear of the instrument.
10. Adjust R20 for a 28.2V dc reading on Multimeter A.
11. Turn the ac power off by reducing the Variac to zero volts or by unplugging the UUT.
12. Remove jumper W1 on A3.
13. Restore ac power.
14. Connect Multimeter A between TP2 and TP1. TP2 is positive with respect to TP1. Set Multimeter to read dc volts, 200V range.
15. Set the BATTERY OPR switch to ON.




Figure 4-9. Battery Charger Test Points and Adjustments on A3 Pre-Regulator PCA

16. Set R10 fully clockwise (CW). Multimeter A should read approximately 45 to 50V dc.
 17. While observing Multimeter A, adjust the rheostat toward minimum resistance. At approximately 26V dc, the BTRY CHG indicator and DS1 should come on. (DS1 is the voltage reference, located on A3, for the constant current source in the battery charger circuit.) The ac line current should increase approximately 110 mA at 115V ac (55 mA at 220V ac).
 18. Adjust the rheostat for a Multimeter A reading of 32V dc.
 19. Turn R10 counter-clockwise (ccw) until the BTRY CHG indicator and DS1 go out. Note that the ac line current has dropped.
 20. Adjust the rheostat toward minimum resistance while observing the BTRY CHG indicator. When the BTRY CHG indicator lights, DS1 lights, and the ac line current increases suddenly. Multimeter A should read between 24.5 and 26.5V dc.
 21. Adjust the rheostat until the BTRY CHG and DS1 indicators turn off. Multimeter A should indicate a dc voltage greater than +31V.
 22. Disconnect all test equipment and the rheostat.
 23. Remove the AC Module from the 732A.
 24. Reinstall jumper W1.
 25. Reinstall the AC Module.
 26. Apply ac power to the UUT.
 27. After a 30 minute warm up, perform the following procedure to make the final adjustment of trickle charge voltage.
 - a. With the battery pack installed and the unit on trickle charge (BTRY CHG LED not on), measure the battery pack voltage. It must be between 27.0 and 27.6V.
 - b. If the trickle charge voltage at the battery pack is not within the required range, record the value and connect the multimeter between TP5 and TP1 (TP5 positive with respect to TP1). Turn the battery switch off.
 28. Adjust R20 the same amount as the difference between the recorded value and 27.0V.
 29. Turn the battery switch on.
- NOTE**
- After making the adjustment, allow 30 minutes for the battery voltage to achieve its final value.*
30. Recheck the battery voltage.
 31. Remove the multimeter.
 32. Battery Charger adjustment is now complete. Perform the Calibration Adjustment Procedure described earlier in this section.
- 4-46. BATTERY CHARGING NOTES**
- 4-47. Battery Charge**
- 4-48. To check the charging and discharging current on the Battery Pack, connect a milliammeter to the two black wires on the BATTERY OPR switch. Then set the switch to OFF.
- 4-49. The charger has two modes: constant current mode and constant voltage mode. In the constant current mode, charging current should be about 200 to 400 mA if the battery is not fully charged. This mode lasts until battery voltage reaches about 31V, at which time it switches to the constant voltage mode (27V) for a few milliamps trickle charge.
- 4-50. Battery Discharge**
- 4-51. At 23°C (the normal oven temperature), with the line cord unplugged, the nominal current drain from the battery is approximately 110 mA.
- 4-52. Individual Battery Checkout**
- 4-53. Individual batteries will accept a charge of 300 to 400 mA at 7.75V max (31V divided by 4) if the terminal voltage is below 6V.
- 4-54. TROUBLESHOOTING**
- 4-55. Introduction**
- 4-56. The following paragraphs describe troubleshooting procedures for the 732A.
- 4-57. External Symptom Troubleshooting**
- 4-58. Using malfunction symptom(s), refer to Table 4-2 to isolate problems within the 732A. Table 4-1 lists the required test equipment for troubleshooting.

4-59. Internal Voltage Measurements

WARNING

TO AVOID ELECTRICAL SHOCK HAZARD, OBSERVE THE FOLLOWING PRECAUTIONS WHILE WORKING ON THE INSIDE OF THE 732A. REMOVE ANY JEWELRY BEFORE BEGINNING TESTING. LINE VOLTAGE AC MAY BE PRESENT DURING THE FOLLOWING TESTS, WHICH SHOULD NOT BE PERFORMED WHILE WORKING ALONE. EXERCISE APPROPRIATE CAUTION TO AVOID ELECTRICAL SHOCK WHEN WORKING IN OR AROUND THE VICINITY OF THE AC POWER CONNECTOR, FUSEHOLDER, AND POWER TRANSFORMER. THE BATTERY ASSEMBLY IS CAPABLE OF GENERATING EXTREMELY HIGH PEAK CURRENTS. AVOID ACCIDENTAL SHORTING OF BATTERY TERMINALS.

CAUTION

The following tests are conducted with power applied to the instrument. To avoid instrument damage, exercise appropriate caution to avoid inadvertently shorting adjacent test points or circuit board traces with test probes or other instrument(s).

CAUTION

To ensure continued instrument performance, do not attempt to replace individual wires in the reference output wiring harness. Replace the entire harness.

4-60. Use the tests shown in Table 4-3 to isolate problems to the major functional circuit groups of the 732A. It is assumed that the external symptoms given in Table 4-2 have been examined and that the primary circuit of the power transformer is operable. This procedure is conducted with the instrument energized. Observe the previously stated WARNINGS and CAUTIONS.

4-61. Oven Repair

4-62. Shifts in the output level that cannot be compensated for by adding or removing jumpers from the A7 Calibration PCA will require the entire Oven Assembly to be returned to Fluke and exchanged for a working unit. Do not attempt to repair the circuitry involving U1,

U2, Q1, Q2, Q5, the resistors associated with TP1 through TP4, or any other component(s) associated with the aforementioned components. Special procedures and auxiliary test equipment are necessary for component replacement within the Reference Circuit. Module exchange is the most economical and expedient method of repair for the user.

4-63. Troubleshooting the A4 Regulator on the Bench

NOTE

No extender is available for the A4 Regulator PCA, but both sides of it are easily accessed while it is installed. To gain better access to the A3 Pre-Regulator PCA, remove the A4 board. Then, when certain that the A3 Pre-Regulator is functioning per Table 4-3 (Internal Measurements), reinstall and test the A4.

4-64. The following procedure may be used to troubleshoot the A4 PCA on a bench top totally isolated from the 732A. An 18 to 32V dc source with current limiting (such as the Fluke 33X series, Fluke 343A, or any common lab supply) is required.

1. Refer to Figure 4-10, a simplified drawing of the A4 Regulator.
2. First, apply +18.6V between TP2 (-) and TP1 (+) to the test sense circuitry. Troubleshoot as needed to obtain the values of circuit voltage shown.

NOTE

Normal current drain is less than 5 mA for either sense of input circuit, therefore set the dc source limit to 10 mA.

3. Then apply 32V between TP3 (-) and TP1 (+) to test the remaining circuitry.

CAUTION

Do not apply a load to the A4 regulator except when it is installed in the 732A with the required heat sinking. Otherwise, damage to power transistor (Q4) could occur.

Table 4-2. External Symptom Troubleshooting

SYMPTOM	PROBABLE CAUSE	ACTION
732A Inoperative.	Fuse blown. Battery dead. Battery opr switch set to OFF. 732A not plugged in.	Check fuse. Measure battery voltage at rear panel jacks. Recharge battery. Visual check. Restore power.
IN CAL indicator off.	Lost ac power, battery dead.	Charge battery, verify instrument calibration.
Repeated fuse blowing.	AC line primary circuit. Power transformer. Bridge rectifier. Battery charger rectifier.	Visual inspection. (2) Use ohmmeter. Use ohmmeter.
Will not run on external ac or dc source.	Ballast lamp open.	Replace lamp.
Output voltage drifts.	Oven or reference.	(1)
Temperature sensitive.	Oven.	Check oven controller circuit.
Output voltage not correct.	Reference.	Perform calibration procedure.
Output voltages not adjustable to specifications.	Reference.	(1)
Battery won't charge.	Defective battery. Battery charger defective.	Replace. Troubleshoot and repair.
Battery won't charge from external source.	Ballast lamp open.	Replace lamp.
<p>(1) The Reference portion of the Oven/Reference Supply assembly is not field repairable. Refer repair to a Fluke Technical Service Center.</p> <p>(2) Return instrument to Fluke Technical Service Center for service.</p>		

Table 4-3. Internal Measurements*

PCB	TEST POINTS	CORRECT VOLTAGE READING	CORRECTIVE ACTION
A3	TP3, TP4	$\leq 60V$ dc	AC line voltage, Rectifier, Power Transformer
A3	TP6, TP4	32V dc	Pre-Regulator
	TP2, TP1	$\leq 31V$ dc	Battery Charger**
A3	TP5, TP1	28.2V dc	Battery Charger**
A4	TP1, TP3	32V dc	Pre-Regulator, Motherboard
A4	TP1, TP2	$\approx 18.6V$ dc	Regulator
Front Panel	10V, COM	10.00000V dc	Oven, Reference Supply
Front Panel	1V, COM	1.000000V dc	Output Divider***
Front Panel	1.018V, COM	1.018000V dc	Output Divider***
Rear Panel	EXT. PWR.	$\geq 24V$ dc	Battery

*Voltage measurements taken with Multimeter A, except for those marked with *** in corrective action column.

**Conditions: battery installed, BATTERY OPR switch ON.

***Calibration of 10V output affects calibration of this output.




Figure 4-10. A4 Regulator (Simplified Diagram)

Section 5

List of Replaceable Parts

TABLE OF CONTENTS

ASSEMBLY NAME	DRAWING NO.	TABLE NO. PAGE	FIGURE NO. PAGE
732A Final Assembly	732A-7201, 732T&B	5-1 5-3	5-1 5-5
A1 LED PCA	732A-4006	5-2 5-9	5-2 5-9
A2 Motherboard PCA	732A-4005T	5-3 5-10	5-3 5-10
A3 Pre-Regulator PCA	732A-4003	5-4 5-11	5-10 5-12
A4 Regulator PCA	732A-4002T	5-5 5-14	5-15 5-15
A5 Reference PCA	732A-4001	5-6 5-16	5-16 5-17
A5A8 Piggyback PCA	732A-4008	5-7 5-18	5-7 5-18
A6 Battery Module PCA	732A-4054	5-8 5-19	5-8 5-20
A7 Calibration PCA	732A-4007	5-9 5-21	5-9 5-21
A8 Preheater PCA	732A-4053	5-10 5-22	5-10 5-22

5-1. INTRODUCTION

5-2. This section contains the parts lists for the 732A DC Reference Standard. Components are listed alphanumerically.

5-3. The parts lists include the following information:

1. Reference Designation.
2. Description of each Part.
3. FLUKE Stock Number.
4. Federal Supply Code for Manufacturers.
5. Manufacturer's Part Number.
6. Total Quantity of Components Per Assembly.

5-4. Although Fluke recommends module exchange in place of component-level repair, this manual also includes schematics (Section 8) and a discussion of the theory of operation (Section 3). Service by non-factory personnel voids the warranty. Use of parts not approved by Fluke may compromise board specifications and operation.

CAUTION

*

**INDICATED DEVICES ARE SUBJECT TO
DAMAGE BY STATIC DISCHARGE.**

5-5. HOW TO OBTAIN PARTS

5-6. Components may be ordered directly from the John Fluke Mfg. Co., Inc. (or its authorized representative) by using appropriate Fluke Stock Number(s), or from the manufacturer by using the manufacturer's part number.

5-7. In the event the part you order has been replaced by a new or improved part, the replacement will be accompanied by an explanatory note and installation instructions, if necessary. To ensure prompt handling of your order, include the following information:

1. Quantity.
2. Fluke Stock Number.
3. Description.
4. Reference Designation.
5. Printed Circuit Board Part Number and Revision Letter.

5-8. Parts price information is available from the John Fluke Mfg. Co., Inc. or from its representatives.

TABLE 5-1. 732A FINAL ASSEMBLY
(SEE FIGURE 5-1.)

REFERENCE DESIGNATOR	FLUKE STOCK	MFRS SPLY	MANUFACTURERS PART NUMBER	TOT QTY	R S	O T	N E
A->NUMERICS----->	S-----DESCRIPTION-----	--NO--	---OR GENERIC TYPE---		-Q	-	-E
A 1	LED PCA	642280	89536 642280	1			
A 2	MOTHERBOARD PCA	650994	89536 650994	1			1
A 3	PRE-REGULATOR PCA	642264	89536 642264	1			
A 4	REGULATOR PCA	642256	89536 642256	1			
A 5	REFERENCE PCA	642272	89536 642272	1			
A 6	BATTERY PCA	732628	89536 732628	1			
A 7	CALIBRATION PCA	645028	89536 645028	1			
A 8	PREHEATER PCA	788323	89536 788323	1			
C 1	CAP. TA, 82UF, +-20%, 20V	357392	12954 D82GS2D20M	1			
CR 1	* DIODE, SI, 100 PIV. 1.5 AMP	116111	05277 1N4817	1			
DS 1	LAMP, NEON, 105-125V, 700MA, WIRE LEADS	100347	74276 T2-24-2	1		1	
E 1-3	BINDING-POST-ASSY-CP-RED	637892	89536 637892	3			
E 4, 5	BINDING-POST-ASSY-CP-BLK	637900	89536 637900	2			
E 6	BINDING-POST-ASSY-BR-BLU	637876	89536 637876	1			
E 7	BINDING-POST-ASSY-BR-GRN	637868	89536 637868	1			
E 8, 9	BINDING-POST-ASSY-BR-WHT	637884	89536 637884	2			
E 10	BINDING POST, BRASS, 1/4-28	102707	89536 102707	1			
H 1	WASHER	644740	89536 644740	2			
H 2	WASHER, FLAT, BRASS, #8, 0.032 THK	631606	89536 631606	16			
H 3	R-P-WASHER	606293	89536 606293	7			
H 4	NUT, MACH, HEX, BR, 8-32	631614	89536 631614	9			
H 5	WASHER, FLAT, S STEEL, #6, 0.016 THK	260471	86928 5710-23-16	2			
H 6	NUT, HEX, MINI, S. STL, 6-32	110569	89536 110569	2			
H 7	WASHER, LOCK, INTRNL, STEEL, #8	110320	89536 110320	9			
H 8	NUT, NYLON, PUSH-IN	222414	83058 PC-97726	16			
H 9	NUT, MACH, HEX, BR, 1/4-28	110619	89536 110619	1			
H 10	WASHER, LOCK, SPLIT, STEEL, 0.263 ID	111518	89536 111518	1			
H 11	WASHER, SHLDR, NYLON, #3	485417	89536 485417	1			
H 12	SCREW, MACH, PHP, STL, 4-40X5/16	152116	89536 152116	1			
H 13	SCREW, MACH, PHP, STL, 6-32X1/4	152140	89536 152140	12			
H 14	SCREW, MACH, FHP, STL, 8-32X1/2	114355	89536 114355	2			
H 15	SCREW, MACH, PHP SEMS, STL, 6-32X1/2	177030	89536 177030	4			
H 16	SCREW, MACH, FHUP, S. STL, 6-32X1/4	320093	89536 320093	4			
H 17	SCREW, MACH, PHP, STL, 8-32X5/8	114983	89536 114983	4			
H 18	SCREW, MACH, FHP, STL, 8-32X5/16	281725	89536 281725	8			
H 19	SCREW, MACH, FHP, STL, 8-32 X 7/16	306159	89536 306159	12			
H 20	SCREW, THD FORM, PHP, S. STL, 8-18X1/2	306233	89536 306233	16			
H 21	NUT, MACH, HEX, STL, 4-40	184044	73734 8002A-NF	2			
H 22	SCREW, MACH, PHP, STL, 6-32X1-1/4	159756	89536 159756	4			
H 23	SCREW, MACH, FHP, STL, 6-32X3/8	114363	73734 182444	4			
H 24	WASHER, SHLDR, NYLON, #3	485417	89536 485417	2		1	
H 25	NUT, MACH, HEX, STL, 4-40	184044	73734 8002A-NF	2			
H 26	SCREW, MACH, FHUP, S. STL, 6-32X1/4	320093	89536 320093	4			
H 27	SCREW, MACH, PHP, STL, 6-32X1/4	152140	89536 152140	20			
H 28	SCREW, MACH, PHP, STL, 6-32X3/4	114223	73734 19048	4			
HR 1, 2	STRIP TYPE	643387	89536 643387	2			
HR 3, 4	PATCH TYPE, 4" X6"	643411	89536 643411	2			
#J 1	CONN, CIRC, CABLE, PLUG, 3 CONTACT	720847	89536 720847	1			
MP 1	FRONT PANEL	641902	89536 641902	1			1
MP 2	CABLE TIE, 4"L, 0.100"W, 0.75 DIA	172080	89536 172080	1			
MP 3	TERM STRIP, LUG, RT ANG, 0.375CTR, 2 POS	654988	89536 654988	1			
MP 4	SLEEVE, POLYOL, SHRINK, .500-.250ID, BLACK	196683	89536 196683	1			1
MP 5	SLEEVE, TEFLON, 0.018ID, NATURAL	325506	89536 325506	1			
MP 6	REAR PANEL	641910	89536 641910	1			
MP 7	CHASSIS SIDE	641928	89536 641928	2			
MP 8	BOTTOM COVER	641944	89536 641944	1			
MP 9	CHASSIS GUARD	641951	89536 641951	1			
MP 10	BULKHEAD, REAR GUARD	641977	89536 641977	1			
MP 11	BULKHEAD, FRONT GUARD	641985	89536 641985	1			
MP 12	SHIM-OVEN	788349	89536 788349	2			
MP 13	INSULATOR, CHASSIS	644906	89536 644906	4			
MP 14	SIDE TRIM	642298	89536 642298	2			
MP 15	INSERT, SIDE TRIM	642306	89536 642306	1			
MP 16	HANDLE	642314	89536 642314	1			
MP 17	STRAP, HANDLE	644880	89536 644880	1			
MP 18	HANDLE RETAINER	579052	89536 579052	2			
MP 19	BRACKET, HANDLE SUPPORT	632414	89536 632414	2			
MP 20	ADHESIVE, SIDE TRIM	680850	89536 680850	2			
MP 21	OVEN-INSUL-OUTER-SIDE-TOP-BOTTOM	654251	89536 654251	3			
MP 22	OVEN-INSUL-INNER-SIDE-TOP-BOTTOM	654269	89536 654269	3			
MP 23	OVEN-INSUL-OUTER-FRONT-BACK	654277	89536 654277	2			
MP 24	OVEN-INSUL-INNER-FRONT-BACK	654285	89536 654285	2			
MP 25	CORNER ANGLE BRACKET-7	298166	89536 298166	2			
MP 26	NAMEPLATE, SERIAL -REAR PANEL-	472795	89536 472795	1			
MP 27	CORNER, PLASTIC 7 INCH	656231	89536 656231	4			
MP 28	FOOT, REAR PANEL 1.25 INCH	657064	89536 657064	4			
MP 29	FOOT, SINGLE BAIL TYPE (DARK UMBER)	653923	89536 653923	4			
MP 30	CABLE TIE, 4"L, 0.100"W, 0.75 DIA	172080	89536 172080	2			

TABLE 5-1. 732A FINAL ASSEMBLY
(SEE FIGURE 5-1.)

REFERENCE DESIGNATOR	A--NUMERICS-->	S	DESCRIPTION	FLUKE STOCK --NO--	MFRS SPLY CODE-	MANUFACTURERS PART NUMBER --OR GENERIC TYPE--	TOT QTY	R S -Q	N D T -E
MP	31		CORD, LINE, R/A 5-15/IEC, 3-18AWG, SVT	363481	89536	363481	1		
MP	32		HOLE PLUG, POLYETHYLENE, F/5/16 HOLE	187799	89536	187799	4		
MP	33		GUIDE, CARD, PCB, 6-1/2 X .312 X .125	326009	89536	326009	4		
MP	34		BANANA PLUG, PANEL	101543	89536	101543	2		
MP	35		BINDING POST PART, HEAD, BRASS, 1/4-28	102889	89536	102889	1		
MP	36		INSUL PART, TRANS, SILICONE, POWER	508630	89536	508630	1		
MP	37		TAPE, FOAM, VINYL, 1/2 W, 1/8 THK	330449	89536	330449	1		
MP	38		CABLE TIE, ANCHOR, ADHSV, 0.122" TIE	407908	89536	407908	2		
MP	39		TAPE, FOAM, ADHSV, 1"W, 1/2 THK	229005	89536	229005	16		
MP	40		COMPONENT HOLDER	104794	98159	2829-115-3	1		
MP	41		PLATE, HEATER TOP	644609	89536	644609	1		
MP	42		PLATE, HEATER BOTTOM	644617	89536	644617	1		
MP	43		COVER, HEATER BOTTOM	644625	89536	644625	1		
MP	44		COVER, HEATER TOP	644633	89536	644633	1		
MP	45		PLATE, SENSOR	644641	89536	644641	2		
MP	46		HOLDER, HEATER TOP A	644658	89536	644658	1		
MP	47		HOLDER, HEATER TOP B	644666	89536	644666	1		
MP	48		HOLDER, HEATER BOTTOM	644773	89536	644773	1		
MP	49		SHIM, HEATER	644781	89536	644781	4		
MP	50		INSULATOR SHEET	650788	89536	650788	2		
MP	51		INSULATOR, THERMISTOR	788356	89536	788356	2		
MP	52		SLEEVE, TEFLON, 0.066ID, NATURAL	113845	89536	113845	1		
MP	53		SPACER, RND, NYLON, 6-32X0.610	643361	89536	643361	4	1	
MP	54		INSUL PART, TRANS, SILICONE, POWER	508630	89536	508630	2		
MP	55		TOP COVER	641936	89536	641936	1		
MP	56		COVER, GUARD	641969	89536	641969	1		
MP	57		GUIDE, CARD, PCB, 6-1/2 X .312 X .125	326009	89536	326009	2		
MP	58		DECAL, FRONT CORNER	659235	89536	659235	2		
MP	59		DECAL, REAR CORNER	685206	89536	685206	2		
MP	60		STRAP, OVEN	644799	89536	644799	1		
MP	61		OVEN-INSUL-OUTER-SIDE-TOP-BOTTOM	654251	89536	654251	1		
MP	62		OVEN-INSUL-INNER-SIDE-TOP-BOTTOM	654269	89536	654269	1		
MP	63		SPACER, RND, NYLON, 6-32X1.200	394262	89536	394262	4		
MP	64		SLEEVE, TEFLON, 0.042ID, NATURAL	175976	89536	175976	1		
MP	65		SLEEVE, PVC, 0.505ID, CLEAR	113761	89536	113761	1		
MP	66		ALIGNMENT TOOL	686113	89536	686113	2		
Q	1,	2	* TRANSISTOR, SI, BV= 60V, 40W, TO-220	642694	89536	642694	2		
R	1		RES, CC, 2.7, +-5%, 1W	159376	89536	159376	1		
R	21		RES, MF, 5.36K, +- 1%, 0.125W, 25 PPM	782714	89536	782714	1		
RT	3,	4	THERMISTOR, DISC, NEG., 10K, +-10%, 25C	644054	89536	644054	2		
RV	1		VARISTOR, 22V, +-20%, 1.0MA	500777	03508	V22ZA1	1	1	
TM	1		732A & 732A/AN INSTRUCTION MANUAL	788414	89536	788414	1		
TM	2		OVEN THERMISTOR CARD, 732A DC REF STD	715698	89536	715698	1		
W	1		WIRE, TEF, EE, UL1180, 22AWG, STRN, BLU	115675	89536	115675	1		
W	2		WIRE, TEF, EE, UL1180, 22AWG, STRN, GRN/YEL	386136	89536	386136	1		
W	3		CABLE ASSY, DIVIDER OUTPUT & REF OUT	644997	89536	644997	1		
W	4		CABLE ASSY, THERMISTOR	651067	89536	651067	1		
W	5		WIRE, BUS, 22 AWG, TINNED COPPER	115469	89536	115469	1		
			RECOMMENDED SPARE PARTS KIT	648845	89536	648845	1		

NOTE 1 = FOR S/N'S 3345000 AND BELOW.
CONTACT YOUR NEAREST SERVICE CENTER
FOR REPLACEMENT INFORMATION.
(REFERENCE PCN 1025)




Figure 5-1. 732A Final Assembly

732A T&B
(1 of 4)






Figure 5-1. 732A Final Assembly (cont)



OVEN TOP VIEW



REAR PANEL VIEW

Figure 5-1. 732A Final Assembly (cont)





Figure 5-1. 732A Final Assembly (cont)

TABLE 5-2. A1 LED PCA
(SEE FIGURE 5-2.)

REFERENCE DESIGNATOR	A->NUMERICS->	S	DESCRIPTION	FLUKE STOCK NO	MFRS SPLY CODE	MANUFACTURERS PART NUMBER OR GENERIC TYPE	TOT QTY	R S T	N O T
DS	1-3	*	LED, RED, LIGHT BAR, PCB MOUNT	534834	28480	HLMP 2300	3		
MP	2		SPACER, SWAGED, RND, BRASS, 0.150IDX0.187	357269	89536	357269	2		




732A-1606

Figure 5-2. A1 LED PCA

TABLE 5-3. A2 MOTHERBOARD PCA
(SEE FIGURE 5-3.)

REFERENCE DESIGNATOR A->NUMERICS->>	S	DESCRIPTION	FLUKE STOCK ---NO---	MFRS SPLY CODE-	MANUFACTURERS PART NUMBER --OR GENERIC TYPE--	TOT. QTY	R S T -Q -E	N O -E
J 1		CONN ACC,PWB EDGE POLARIZING INSERT	291716	89536	291716	8		
J 2, 3, 4		CONN,PWB EDGE,REC,0.150 CTR,6 PDS	291708	91662	6308-006-313-001	17		
TP 1- 2		TERM,FASTON,TAB,SOLDR,0.110 WIDE	512889	02660	62395	2		
W 1, 2		CABLE SET ASSY, REF AMP & LED	651059	89536	651059	1		




732A-1605

Figure 5-3. A2 Motherboard PCA


TABLE 5-4. A3 PRE-REGULATOR PCA
(SEE FIGURE 5-4.)

REFERENCE DESIGNATOR	A->NUMERICS->>	S	DESCRIPTION	FLUKE STOCK NO	MFRS SPLY CODE	MANUFACTURERS PART NUMBER OR GENERIC TYPE	TOT QTY	R S -Q	N O T -E
C	1		CAP, AL, 220UF, +75-20%, 100V	381947	74840	ITA	1		
C	2		CAP, CER, 0.1UF, +-20%, 100V, Z5V	149146	56289	33C4186	1	1	
C	3		CAP, AL, 330UF, +75-20%, 80V	292862	89536	292862	1		
C	4		CAP, CER, 4700PF, +-20%, 100V, X7R	362871	72982	8121-A100-WSR-472M	1		
C	5		CAP, TA, 1UF, +-10%, 35V	161919	56289	196D010X0035G	1		
C	6		CAP, CER, 0.01UF, +-20%, 100V, X7R	407361	72982	8121-A100-WSR-103M	1		
C	9, 10		CAP, CER, 0.05UF, +-20%, 100V, Z5V	149161	56289	55C23A1	2		
CR	1		* DIODE, SI, RECT, BRIDGE, BV=200V, IO=1.0A	295509	09423	FB200	1	1	
CR	2, 3, 13,		* DIODE, SI, 100 PIV, 1.5 AMP	116111	05277	1N4817	4	1	
CR	19		*	116111					
CR	4		* ZENER, UNCOMP, 40.0V, 5%, 6.5MA, 1.0W	407825	89536	1047825	1	1	
CR	5- 9, 11,		* DIODE, SI, BV= 75.0V, IO=150MA, 500 MW	203323	07910	1N4448	8	1	
CR	16, 23		*	203323					
CR	10		* ZENER, UNCOMP, 30.0V, 10%, 4.2MA, 0.4W	272633	04713	1N972A	1	2	
CR	12, 14		* ZENER, UNCOMP, 5.2V, 1.0MA, 0.4W	233627	89536	233627	2	1	
CR	15		* ZENER, UNCOMP, 5.6V, 5%, 3.0MA, 0.5W	535559	89536	535559	1	1	
CR	18		* DIODE, GER, BV=100.0V, IO= 90MA, 80 MW	149187	93332	1N270	1	1	
CR	20- 22, 28		* DIODE, SI, N-JFET, CURRENT REG, IF=1.0 MA	334839	11532	TCR5297	4	1	
CR	24, 26		* DIODE, SI, N-JFET, CURRENT REG, IF=3.0 MA	429373	89536	429373	2	1	
CR	25		* DIODE, SI, N-JFET, CURRENT REG, IF=5.3 MA	334714	07910	TCR5315	1	1	
CR	29		* ZENER, UNCOMP, 20.0V, 5%, 6.2MA, 0.4W	180463	89536	180463	1		
CR	30		* ZENER, COMP, 6.4V, 3%, 1PPM TC, 2.0MA	357848	89536	357848	1		
DS	1		* LED, VISIBLE RED, PCB MNT, LUM INT=0.5MCD	369777	28480	5082-4480	1		
E	1		TERM, RING .072 & .196, SOLDR	101030	79963	174	1		
E	1- 6		TERM, FASTON, TAB, SOLDR, 0.110 WIDE	512889	02660	62395	6		
F	1		FUSE, 1/4 X 1-1/4, SLOW, 0.375A, 250V	109264	71400	MDX3-8	1		
FL	1		FILTER, LINE, 250V/1A MAX	649988	89536	649988	1	5	
H	1		WASHER, SHLDR, NYLON, #3	485417	89536	485417	2		
H	2		SCREW, MACH, PHP, STL, 4-40X5/16	152116	89536	152116	2		
H	3		SCREW, MACH, PHP, STL, 6-32X5/16	152157	89536	152157	4		
H	5		SCREW, MACH, PHP, STL, 8-32X1/4	228890	89536	228890	2		
H	6		SCREW, THD FORM, PHP, S STL, 8-18X5/8	574673	89536	574673	4		
H	7		SCREW, MACH, PHP, STL, 6-32X3/8	152165	89536	152165	2		
H	8		WASHER, FLAT, MYLAR, #8, 0.010 THK	197426	89536	197426	2		
H	9		WASHER, FLAT S STEEL, #4, 0.032 THK	146225	89536	146225	2		
H	10		NUT, MACH, HEX, STL, 4-40	110635	89536	110635	2		
H	11		NUT, NYLON, PUSH-IN	222406	83058	PC97476	4		
MP	1		INSUL PART, TRANS, SILICONE, POWER	508630	89536	508630	2		
MP	2		SPACER, RND, SOLUBLE, 0.062IDX0.1560D	296319	32559	T0806	2		
MP	3		BRACKET, PRE-REGULATOR	641993	89536	641993	1		
MP	4		PANEL, PRE-REGULATOR	644583	89536	644583	1		
MP	5		HLDR, FUSE, 1/4 X 1-1/4, LOPROFILE, PNLMT	424416	89536	424416	1		
MP	6		ROUND, L=1.750, THREAD 6-32, BLACK	650242	89536	650242	1		
MP	7		HEAT SINK	644674	89536	644674	1		
MP	8		CLIP-ON TYPE FOR T0-220	644062	89536	644062	1		
MP	9		SPACER L.E.D	541284	89536	541284	1	1	
Q	1		* TRANSISTOR, SI, BV= 60V, 65W, T0-220	386128	01295	T1P120	1		
Q	2		* TRANSISTOR, SI, BV= 60V, 40W, T0-220	642694	89536	642694	1	1	
Q	3		* TRANSISTOR, SI, BV=100V, 40W, T0-220	454033	04713	MJE15028	1		
Q	4		* TRANSISTOR, SI, NPN, SMALL SIGNAL	242065	04713	2N5089	1	1	
Q	5		* TRANSISTOR, SI, PNP, SMALL SIGNAL	195974	64713	2N3906	1	1	
Q	6		* TRANSISTOR, SI, NPN, SMALL SIGNAL	168716	07263	S19254	1	1	
R	1		WM RESISTOR	717892	89536	717892	1		
R	3		RES, CC, 3.3, +-5%, 0.5W	188482	89536	188482	1		
R	4		RES, MF, 1.54K, +-1%, 0.125W, 25PPM	335331	89536	335331	1		
R	6		RES, CC, 510, +-5%, 0.25W	218032	01121	CR5115	1		
R	7		RES, MF, 22.6, +-1%, 0.125W, 100PPM	296640	89536	296640	1	1	
R	8		RES, MF, 402, +-1%, 0.125W, 100PPM	289611	91637	CMF554020F	1		
R	9		RES, CC, 10K, +-5%, 0.5W	109165	01121	EB1035	1		
R	10		RES, VAR, CERM, 500, +-20%, 0.5W	267849	11236	190FC501B	1	1	
R	11		RES, MF, 6.65K, +-1%, 0.125W, 100PPM	294918	91637	CMF551272F	1		
R	12		RES, MF, 16.2K, +-1%, 0.125W, 100PPM	226233	89536	226233	1		
R	13		RES, MF, 33.2K, +-0.5%, 0.125W, 100PPM	334102	89536	334102	1		
R	14		RES, MF, 17.4K, +-1%, 0.125W, 50PPM	349175	89536	349175	1		
R	15		RES, CC, 10K, +-5%, 0.25W	148106	01121	CB1035	1		
R	16		* RES, CC, 100K, +-5%, 0.25W	148189	89536	148189	1		
R	19		RES, CF, 15K, +-5%, 0.25W	348854	80031	CR251-4-5P15K	1		
R	20		* RES, VAR, CERM, 5K, +-20%, 0.5W	267872	89536	267872	1		
RT	1		THERMISTOR, DISC, NEG., 10K, +-10%, 25C	104596	73168	JA41J	1		
S	1, 2		SWITCH, SLIDE, DPDT, POWER	234278	89536	234278	2	1	
T	1		POWER TRANSFORMER	645036	89536	645036	1		
W	1		JUMPER, WIRE, NONINSUL, 0.275CTR	529271	89536	529271	1		



732A-1603

Figure 5-4. A3 Pre-Regulator PCA



732A-4003

Figure 5-4. A3 Pre-Regulator PCA (cont)

TABLE 5-5. A4 REGULATOR PCA
(SEE FIGURE 5-5.)

REFERENCE DESIGNATOR	S	DESCRIPTION	FLUKE STOCK --NO--	MFRS SPLY CODE	MANUFACTURERS PART NUMBER --OR GENERIC TYPE--	TOT QTY	R S -R	N O T -E
C 1		CAP, AL, 330UF, +75-20%, 80V	292862	89536	292862	1		
C 2		CAP, TA, 82UF, +-20%, 20V	357392	12934	D82GS2D20M	1		
C 3, 4		CAP, TA, 10UF, +-20%, 35V	417683	56289	196D106X0035KA1	2		
C 5, 8, 10		CAP, CER, 0.22UF, +-20%, 50V, Z5U	309849	71590	CW3C0C224K	3		
C 6		CAP, CER, 0.047UF, +-20%, 50V, Z5U	460733	71590	CW20C473M	1		
C 7		CAP, CER, 0.01UF, +80-20%, 50V, Y5R	369579	89536	369579	1		
C 9		CAP, TA, 22UF, +-20%, 35V	394775	56289	196D226X0035TE4	1		
CR 3, 5, 6	*	DIODE, SI, BV= 75.0V, IO=150MA, 500 MW	203323	07910	1N4448	3	1	
CR 7, 8	*	DIODE, GER, BV=100.0V, IO= 80MA, 80 MW	149187	93332	1N270	2	1	
CR 10	*	DIODE, SI, N-JFET, CURRENT REG, IF=0.43MA	393454	89536	393454	1	1	
CR 12, 14	*	DIODE, SI, N-JFET, CURRENT REG, IF=1.0 MA	334839	11532	1CR5297	2	1	
CR 13	*	DIODE, SI, N-JFET, CURRENT REG, IF=1.0 MA	348482	89536	348482	1	1	
CR 15	*	DIODE, SI, 100 PIV, 1.5 AMP	116111	05277	1N4817	1	1	
DS 1		LED, VISBLE RED, PCB MNT, LUM INT=0.5MCD	369777	28480	5082-4480	1		
E 1- 3		TERM, FASTON, TAB, SOLDR, 0.110 WIDE	512889	02660	62395	3		
MP 1		SPACER L.E.D.	541284	89536	541284	1	1	
Q 1, 5	*	TRANSISTOR, SI, PNP, SMALL SIGNAL	229898	04713	MP56522	2		
Q 2	*	TRANSISTOR, SI, PNP, SMALL SIGNAL	218388	07263	2N3645	1	1	
Q 3, 8	*	SMALL SIGNAL NPN DOUBLE DIFF, SILICON	352138	89536	352138	2	1	
Q 4	*	TRANSISTOR, SI, BV=100V, 40W, T0-220	454033	04713	MJE15028	1	1	
Q 6	*	TRANSISTOR, SI, NPN, SMALL SIGNAL	218396	04713	2N3904	1	1	
Q 7	*	TRANSISTOR, SI, UJT, PNP	268110	03508	2N6027	1	1	
Q 12, 13	*	TRANSISTOR, SI, PNP, SMALL SIGNAL	195974	64713	2N3906	2	1	
R 1		RES, MF, 348, +-1%, 0.125W, 100PPM	236778	89536	236778	1		
R 2		RES, MF, 1.21K, +-1%, 0.125W, 100PPM	229146	91637	CMF551211F	1		
R 3		RES, MF, 24.3, +-1%, 0.125W, 100PPM	281816	91637	CMF5524R2F	1		
R 4		RES, MF, 8.66K, +-1%, 0.125W, 25PPM	330738	91637	CMF558661F	1		
R 5		RES, CC, 3K, +-5%, 0.25W	193508	01121	CB3025	1		
R 6		RES, CC, 4.3K, +-5%, 0.25W	193375	01121	CB4325	1		
R 7		RES, CC, 18K, +-5%, 0.25W	148122	01121	CB1835	1		
R 8, 9		RES, CC, 91K, +-5%, 0.25W	193300	89536	193300	2		
R 10		RES, MF, 5K, +-0.1%, 0.125W, 25PPM	340240	91637	CMF555001B	1		
R 11		RES, CC, 10K, +-5%, 0.25W	148106	01121	CB1035	1		
R 12		RES, MF, 0.39, +-5%, 2W	219386	89536	219386	1		
R 13		RES, CC, 2.7, +-5%, 1W	159376	89536	159376	1		
R 14		RES, CC, 150K, +-5%, 0.25W	182212	01121	CB1545	1		
R 15, 25		RES, CC, 51K, +-5%, 0.25W	193334	01121	CB5135	2		
R 16		RES, CC, 1M, +-5%, 0.25W	182204	01121	CB1055	1		
R 17		RES, CC, 2.7K, +-5%, 0.25W	170720	01121	CB2725	1		
R 18		RES, MF, 80.6K, +-1%, 0.125W, 25PPM	312710	91637	CMF558062F	1		
R 19		RES, CC, 18K, +-5%, 0.5W	187898	01121	EB1835	1		
R 20		RES, CC, 8.2K, +-5%, 0.25W	160796	01121	CB8225	1		
R 21		RES, CC, 1K, +-5%, 0.25W	148023	01121	CB1025	1		
R 22		RES, VAR, CERH, 5K, +-20%, 0.75W	159905	89536	159905	1	1	
R 23		RES, CC, 270K, +-5%, 0.25W	220061	89536	220061	1		
R 24		RES, CC, 10, +-5%, 0.25W	147868	01121	CB1005	1		
R 27		RES, MF, 316K, +-1%, 0.125W, 100PPM	289496	91637	CMF553163F	1		
VR 1	*	ZENER, UNCOMP, 10.0V, 5%, 20.0MA, 0.5W	473744	07910	1N5240	1	1	
VR 2	*	ZENER, COMP, 6.4V, 5%, 5 PPM TC, 1.0MA	330829	07910	1N4571	1	1	
VR 9	*	ZENER, UNCOMP, 10.0V, 5%, 20.0MA, 0.5W	473744	07910	1N5240	1	1	
W 1		WIRE, BUS, 22 AWG, TINNED COPPER	115469	89536	115469	1		





Figure 5-5. A4 Regulator PCA

732A-1602

TABLE 5-6. AS REFERENCE PCA
 (SEE FIGURE 5-6.)

REFERENCE DESIGNATOR A->NUMERICS	S	DESCRIPTION	FLUKE STOCK --NO--	MFRS SFLY CODE	MANUFACTURERS PART NUMBER --OR GENERIC TYPE--	TOT QTY	R S	N O T -E
A	8	ASSY, PIGGYBACK	751578	89536	751578	1		
C	1, 2, 5,	CAP, CER, 0.22UF, +-20%, 50V, Z5U	309849	71590	CW3CC0224K	5	1	
C	6, 15		309849					
C	3	CAP, POLYCA, 1UF, +-10%, 50V	271219	84411	X4630W1029, 50W	1		
C	4	CAP, CER, 330PF, +-5%, 100V, C0G	528620	51406	RFE121	1		
C	7	CAP, CER, 0.0047UF, +-20%, 50V, Z5U	255471	51442	200-050-601-502M	1		
C	8	CAP, POLYCA, 5UF, +-10%, 50V	313254	84411	X4630W5.0UF-10P-50	1		
C	9	CAP, POLYCA, 0.47UF, +-10%, 100V	288860	84411	X4630W	1		
C	10	CAP, POLYCA, 4UF, +-20%, 50V	340281	89536	340281	1		
C	11	CAP, MICA, 270PF, +-5%, 500V	140452	72136	DM15F271J	1		
C	12, 13	CAP, MICA, 100PF, +-5%, 500V	148494	72136	DM15F101J	2		
C	14	CAP, CER, 180PF, +-10%, 1000V, S3N	105890	56289	C023B102E181M	1		
C	16	CAP, CER, 0.047UF, +-20%, 50V, Z5U	440733	71590	CW20C473H	1		
C	17	CAP, CER, 1200PF, +-20%, 100V, X7R	358283	72982	8121-A100-W5R-122M	1		
CK	3, 4, 7	* DIODE, SI, BV= 75.0V, IO=150MA, 500 MW	203323	07910	1N4448	3	1	
E	1-14	TERM, FASTON, TAB, SOL, DR, 0.110 WIDE	512889	02660	62395	4		
H	1	SCRW, SET, HH, STL, 6-32X3/4	643395	89536	643395	1		
MP	1	T0-5 CASE THERMALLOY #11158	380220	89536	380220	1		
MP	2	SPACER, RND, NYLON, 6-32X0.610	643361	89536	643361	1		
MP	3	SPACER, RND, NYLON, 6-32X1.200	394262	89536	394262	1		
MP	4	SPACER, RND, SOLUBLE, 0.062IDX0.1560D	296319	32559	T0806	16		
MP	5	INSUL PART, TRANS, SILICONE, T0-5	658807	89536	658807	1		
MP	6	CABLE TIE, 4"L, 0.100"W, 0.75 DIA	172080	89536	172080	1		
MS	1	* REF AMP SET	645010	89536	645010	1	1	1
Q	1, 2	* TRANSISTOR, SI, NPN, SMALL SIGNAL	218394	04713	2N3904	2		
Q	3, 4	* TRANSISTOR, SI, PNP, SMALL SIGNAL	195974	64713	2N3906	2	1	
Q	5	* DIODE, SI, N-JFET, CURRENT REG, IF=1.0 MA	334839	11532	TCR5297	1	1	
Q	5	* SMALL SIGNAL NPN DOUBLE DIFF, SILICON	352138	89536	352138	1	1	
R	1	RES, CC, 200, +-5%, 0.25W	193482	89031	CR251-4-5P200E	1		
R	2, 3	MATCHED RESISTOR SET, 10.0V	634824	89536	634824	1	1	
R	4	1.27K	634915	89536	634915	1	1	
R	7, 8	REF AMP DIVIDER SET	715706	89536	715706	1	1	
R	12	20K	634840	89536	634840	1	1	
R	13	W W RESISTOR	711184	89536	711184	1	1	
R	14	RES, WW 250K +/-0.6%, .5W	238485	89536	238485	1	1	
R	15, 18	RES, WW CARD TYPE 500K	195388	89536	195388	2	1	
R	16, 56	RES, WW CARD, 1/2W, 1K,	131706	89536	131706	2	1	
R	17	RES, WW CARD, 1/2W, 2K,	131714	89536	131714	1	1	
R	19, 48, 49,	RES, WW, 35 +/-1.5%, 1/4W	634907	89536	634907	5	1	
R	54, 55		634907					
R	20	RES, VAR, CERM, 100, +-20%, 0.75W	159889	89536	159889	1		
R	22	RES, MF, 17.4K, +-1%, 0.125W, 25PPM	335372	89536	335372	1		
R	23, 26, 64	RES, CC, 51, +-5%, 0.25W	221879	01121	CB5105	3		
R	24	RES, CC, 10, +-5%, 0.25W	147868	01121	CB1005	1		
R	25	RES, CC, 30K, +-5%, 0.25W	193417	01121	CB3035	1		
R	27	RES, MF, 10K, +-0.1%, 0.125W, 25PPM	435065	89536	435065	1		
R	28	RES, MF, 7.5K, +-1%, 0.125W, 25PPM	484881	89536	484881	1		
R	29	RES, MF, 19.1K, +-1%, 0.125W, 25PPM	291518	89536	291518	1		
R	30	RES, MF, 2.15K, +-1%, 0.125W, 50PPM	347039	89536	347039	1		
R	31	RES, MF, 1K, +-0.1%, 0.125W, 25PPM	340380	91637	CMF55	1	1	
R	32	RES, CC, 6.2M, +-5%, 0.25W	221960	01121	CB6255	1		
R	33	RES, CC, 5.1M, +-5%, 0.25W	296467	01121	CB5155	1		
R	34	RES, CC, 2.4M, +-5%, 0.25W	221945	01121	CB2455	1		
R	35, 36	RES, CC, 27M, +-5%, 0.25W	221994	89536	221994	2		
R	37	RES, CC, 1K, +-5%, 0.25W	148023	01121	CB1025	1		
R	38, 39, 42	RES, CC, 51K, +-5%, 0.25W	193334	01121	CB5135	3		
R	40	RES, CC, 10K, +-5%, 0.25W	148106	01121	CB1035	1		
R	41	RES, CC, 6.8M, +-5%, 0.25W	394064	01121	CB	1		
R	43	RES, CC, 100K, +-5%, 0.25W	148189	01121	CB1045	1		
R	57	350K	642801	89536	642801	1	1	
R	58	RES, VAR, CERM, 200, +-20%, 0.75W	186213	89536	186213	1		
R	59	RES, VAR, WW, 200, +-3%, 2W, BSH, MTG	542928	89536	542928	1		
R	61	RES, CC, 1M, +-5%, 0.25W	182204	01121	CB1055	1		
R	62, 63	RES, CC, 2.7, +-5%, 0.5W	218743	89536	218743	2		
RT	1, 2	THERMISTOR, DISC, NEG., 10K, +-10%, 25C	104596	73168	JA41J1	2		
U	1	* IC, OP AMP, GEN PURPOSE, TO-78 METAL CAN	288928	12040	LM308AH	1		
U	3	* IC, OP AMP, SELECTED DIFF OUT VOLT=2MV	473777	89536	473777	1	1	
U	4, 5	* IC, OP AMP, GENERAL PURPOSE, TO-5 CASE	284760	12040	LM308H	2	1	
VR	1	* ZENER, UNCOMP, 5.2V, 1.0MA, 0.4W	233527	89536	233627	1	1	
VR	2	* ZENER, COMP, 6.4V, 5%, 5 PPM TC, 1.0MA	330829	07910	1N4571	1	1	
VR	6, 8	* ZENER, UNCOMP, 20.0V, 5%, 23.0MA, 1.0W	276980	12969	UZ8712	2	1	




732A-1601

Figure 5-6. A5 Reference PCA

TABLE 5-7. A5A8 PIGGYBACK PCA
(SEE FIGURE 5-7.)

REFERENCE DESIGNATOR	S	DESCRIPTION	FLUKE STOCK NO	MFRS SPLY CODE	MANUFACTURERS PART NUMBER OR GENERIC TYPE	TOT QTY	R S Q	N O T E
J 1- 5		PIN, SINGLE, FWR, 0.025 SQ	267633	00779	86144-1	5		
MF 1		SPACER, SWAGED, RND, BRASS, 2-56X0.375	342956	89536	342956	1	1	
R 44, 46		MATCHED RESISTOR SET, 1.0V	751917	89536	751917	1		
R 45, 47		MATCHED RESISTOR SET, 1.018V	751925	89536	751925	1	1	



732A-1608

Figure 5-7. A5A8 Piggyback PCA

TABLE 5-8. A6 BATTERY MODULE PCA
(SEE FIGURE 5-8.)

REFERENCE DESIGNATOR		FLUKE STOCK	MFRS SPLY	MANUFACTURERS PART NUMBER	TOT	R	N
A--NUMERICS---	S	---NO---	CODE--	---OR GENERIC TYPE---	QTY	--Q	--E
BT	1- 4	BATTERY, LEAD-ACID, GEL, 6.0V, 4.0AH	739961	89536 739961	4		
CR	1	* DIODE, SI, 100 PIV, 1.5 AMP	116111	05277 1N4817	1	1	
DS	1	LAMP, END CAPS, 13.5V, 520MA	643346	89536 643346	1		
E	9, 11, 13,	WIRE ASSY, BLUE	738385	89536 738385	4		
E	15		738385				
E	10, 12, 14,	WIRE ASSY, RED	738377	89536 738377	4		
E	16		738377				
H	1	RIVET PLATED BRASS, 1/8 X 5/32 L	161315	89536 161315	2		
H	2	SCREW, THD FORM, PHP, S STL, 8-18X5/8	574673	89536 574673	4		
H	3	SCREW, THD CUT, PHP, S STL, 4-24X3/8	183574	89536 183574	3		
H	4	SCREW, MACH, PHP, STL, 6-32X1/4	152140	89536 152140	12		
H	5	SCREW, MACH, PHP, S STL, 6-32X1/4	320093	89536 320093	13		
H	6	SCREW, MACH, PHP, STL, 6-32X3/8	152165	89536 152165	2		
H	7	WASHER, FLAT, MYLAR, #8, 0.010 THK	197426	89536 197426	2		
J	10	CONNECTOR	720844	89536 720844	1		
MP	1	BANANA JACK, PANEL, RED	162065	89536 162065	1		
MP	2	BANANA JACK, PANEL, BLACK	162073	89536 162073	3		
MP	3	FRONT END, BATTERY ENCLOSURE	644682	89536 644682	1		
MP	4	REAR END, BATTERY ENCLOSURE	644732	89536 644732	1		
MP	5	TOP, BATTERY ENCLOSURE	644690	89536 644690	1		
MP	6	BOTTOM, BATTERY ENCLOSURE	644708	89536 644708	1		
MP	7	INBD SIDE, BATTERY ENCLOSURE	644716	89536 644716	1		
MP	8	PANEL, BATTERY	644591	89536 644591	1		
MP	9	OUTBD SIDE, BATTERY ENCLOSURE	644724	89536 644724	1		
MP	10	NUT, INSULATOR	279398	89536 279398	4		
MP	11	TAPE, FOAM, VINYL, 1/2 W, 1/8 THK	330449	89536 330449	1		
MP	12	ROUND, L=1.750, THREAD 6-32, BLACK	650242	89536 650242	1		
MP	13	BATTERY PANEL	732602	89536 732602	1		
R	1	RES. CC, 51K, +-5%, 0.25W	193334	01121 CB5135	1	1	
RT	1, 2	THERMISTOR, DISC, NEG., 10K, +-10%, 25C	104596	73168 JA41J1	2		
XDS	1	HLDK, COMPONENT, 13/32, PWB MT	103028	89536 103028	2		





Figure 5-8. A6 Battery Module PCA

TABLE 5-9. A7 CALIBRATION PCA
(SEE FIGURE 5-9.)

REFERENCE DESIGNATOR A--NUMERICS-->	S-----DESCRIPTION-----	FLUKE STOCK --NO--	MFRS SPLY CODE--	MANUFACTURERS PART NUMBER --OR GENERIC TYPE--	TOT QTY	R S	N D T E
P 1	CABLE ASSY, CALIBRATION	644989	89536	644989	1		



732A-4007

Figure 5-9. A7 Calibration PCA

TABLE 5-10. A8 PREHEATER PCA
(SEE FIGURE 5-10.)

REFERENCE DESIGNATOR	A-NUMERICS	S	DESCRIPTION	FLUKE STOCK NO	MFRS SPLY CODE	MANUFACTURERS PART NUMBER OR GENERIC TYPE	TOT QTY	R S T	N O E
F	1		FUSE THERMAL 136F +-2.5F	715110	89536	715110	1		
R	1, 2		RES, CC, 2K, +-5%, 0.25W	202879	01121	CR2025	2		




Figure 5-10. A8 Preheater PCA

Section 6

Accessories and Options

6-1. INTRODUCTION

6-2. This section describes accessories and options available with the model 732A.

6-3. DUAL-MOUNTING FASTENER (M00-800-523)

6-4. The Dual-Mounting Fastener is a 8-32 threaded fastener designed for bolting two half-rack-width instruments together. The Dual-Mounting Fasteners may be used for either dual rack mounting applications (as used in the M07-200-603 Full-Width Rack Mount Kit) or dual table top applications. Four M00-800-523 fasteners are required for each pair of half-rack width instruments.

6-5. HALF-WIDTH RACK MOUNT KIT (M07-203-601)

6-6. The Half-Width Rack Mounting kit permits the 732A to be rack mounted. A blank filler panel is supplied, allowing left or right hand offset mounting. Assembly instructions are supplied with the kit.

6-7. FULL-WIDTH RACK MOUNT KIT (M07-200-603)

6-8. The Full-Width Rack Mounting kit permits the 732A to be rack mounted side-by-side with another half-rack-width instrument. This rack mounting method requires the 732A to be bolted to the adjacent instrument. To facilitate bolting the instruments together, four M00-800-523, Dual Mounting Fasteners are included with the kit. Assembly instructions are supplied with the kit.

6-9. LOW THERMAL EMF CABLE SET (5440A-7002)

6-10. This set of three dual-conductor, shielded cables is terminated in low thermal emf banana-plug-type connectors. The set consists of one 4-foot long cable and two 2-foot long cables. Individual cables are identified as P/N 750133 for the 2-foot cable and P/N 750141 for the 4-foot cable.

6-11. BATTERY PACK (732A-7005)

6-12. The Battery Pack is a replacement module for the rear panel Battery Module on the 732A. It may be used as an additional auxiliary source or as a spare.

6-13. TRANSIT CASE (732A-7002)

6-14. The Transit Case provides a means of transporting the 732A while it is continuously powered by a battery source contained within the Transit Case. This allows continuity of standardization during transportation over long distances.

6-15. BATTERY CHARGER (732A-7003)

6-16. The Battery Charger provides the capability to charge up to four battery packs at once. This unit is designed to be used with the transit case for extended battery operation during transit.

6-17. RACK TRAY KIT (732A-7004)

6-18. The 732A-7004 Rack Tray Kit provides for mounting the 732A DC Reference Standards and/or the 752A Reference Divider in a standard 19-inch equipment rack. The Rack Tray Kit holds up to two instruments.

6-19. SPECIAL CALIBRATION (732A-000)

6-20. The 732A-000 Special Calibration option provides output voltage calibration of a new 732A in the Fluke Standards Lab, a Traceability Test Report and shipment of the unit under power to maintain traceability.

6-21. SPECIAL CALIBRATION (732A-000R)

6-22. The 732A-000R Special calibration option provides the same calibration (or recalibration) as 732A-000, but for a previously purchased instrument.

6-23. SPECIAL CALIBRATION AND DRIFT RATE CHARACTERIZATION (732A-100)

6-24. The 732A-100 option provides output voltage calibration and drift rate characterization of a new 732A for a period of sixty days in the Fluke Standards Lab, a Traceability Test Report, and shipment of the instrument under power.

6-25. SPECIAL CALIBRATION AND DRIFT RATE CHARACTERIZATION (732A-100R)

6-26. The 732A-100R option provides the same calibration (or recalibration) as 732A-100, but for a previously purchased instrument.

6-27. DIRECT VOLTAGE MAINTENANCE PROGRAM (732A-200)

6-28. The 732A-200 option provides 10V reference certification of a 732A at the customer's facility. It includes 1-week use of a Fluke owned and certified 732A (for on-site calibration of a 10V reference) and a Traceability Report.

6-29. ADDITIONAL 10V REFERENCE CERTIFICATION (732A-201)

6-30. The 732A-201 option provides for certification of additional 10V references at the same site of a 732A-200 option, using the Fluke owned 732A. This option may only be ordered with option 732A-200.

Section 7

General Information

7-1. This section of the manual contains generalized user information as well as supplemental information to the List of Replaceable Parts contained in Section 5.

List of Abbreviations and Symbols

A or amp	ampere	hf	high frequency	(+) or pos	positive
ac	alternating current	Hz	hertz	pot	potentiometer
af	audio frequency	IC	integrated circuit	p-p	peak-to-peak
a/d	analog-to-digital	If	intermediate frequency	ppm	parts per million
assy	assembly	in	inch(es)	PROM	programmable read-only memory
AWG	american wire gauge	intl	internal	psi	pound-force per square inch
B	bel	I/O	input/output	RAM	random-access memory
bcd	binary coded decimal	k	kilo (10 ³)	rf	radio frequency
°C	Celsius	kHz	kilohertz	rms	root mean square
cap	capacitor	kΩ	kilohm(s)	ROM	read-only memory
ccw	counterclockwise	kV	kilovolt(s)	s or sec	second (time)
cer	ceramic	lf	low frequency	scope	oscilloscope
cermet	ceramic to metal(seal)	LED	light-emitting diode	SH	shield
ckt	circuit	LSB	least significant bit	Si	silicon
cm	centimeter	LSD	least significant digit	serno	serial number
cmrr	common mode rejection ratio	M	mega (10 ⁶)	sr	shift register
comp	composition	m	milli (10 ⁻³)	Ta	tantalum
cont	continue	mA	milliampere(s)	tb	terminal board
crt	cathode-ray tube	max	maximum	tc	temperature coefficient or temperature compensating
cw	clockwise	mf	metal film	tcxo	temperature compensated crystal oscillator
d/a	digital-to-analog	MHz	megahertz	tp	test point
dac	digital-to-analog converter	min	minimum	u or μ	micro (10 ⁻⁶)
dB	decibel	mm	millimeter	uhf	ultra high frequency
dc	direct current	ms	millisecond	us or μs	microsecond(s) (10 ⁻⁶)
dmm	digital multimeter	MSB	most significant bit	uut	unit under test
dvm	digital voltmeter,	MSD	most significant digit	V	volt
elect	electrolytic	MTBF	mean time between failures	v	voltage
ext	external	MTTR	mean time to repair	var	variable
F	farad	mV	millivolt(s)	vco	voltage controlled oscillator
°F	Fahrenheit	mv	multivibrator	vhf	very high frequency
FET	Field-effect transistor	MΩ	megohm(s)	vlf	very low frequency
ff	flip-flop	n	nano (10 ⁻⁹)	w	watt(s)
freq	frequency	na	not applicable	ww	wire wound
FSN	federal stock number	NC	normally closed	xfmr	transformer
g	gram	(-) or neg	negative	xstr	transistor
G	giga (10 ⁹)	NO	normally open	xtal	crystal
gd	guard	ns	nanosecond	xtlo	crystal oscillator
Ge	germanium	opnl ampl	operational amplifier	Ω	ohm(s)
GHz	gigahertz	p	pico (10 ⁻¹²)	μ	micro (10 ⁻⁶)
gmV	guaranteed minimum value	para	paragraph		
gnd	ground	pcb	printed circuit board		
H	henry	pF	picofarad		
hd	heavy duty	pn	part number		

Federal Supply Codes for Manufacturers

D9816 Westermann Wilhelm Augusta-Anlage Mannheim-Nackarau Germany	02533 Leigh Instruments Ltd. Frequency Control Div. Don Mills, Ontario, Canada	04713 Motorola Inc. Semiconductor Group Phoenix, Arizona	06665 Precision Monolithics Sub of Bourns Inc. Santa Clara, California
00199 Marcon Electronics Corp Keamy, New Jersey	02606 Fenwal Labs Division of Travenal Labs Morton Grove, Illinois	05236 Jonathan Mfg. Co. Fullerton, California	06666 General Devices Co. Inc. Indianapolis, Indiana
00213 Nytronics Comp. Group Inc. Dandlinton, South Carolina	0266 Bunker Ramo-Eltra Corp. Amphenol NA Div. Broadview, Illinois	05245 Corcom Inc. Libertyville, Illinois	06739 Electron Corp. Littleton, Colorado
00327 Welwyn International Inc. Westlake, Ohio	02735 RCA-Solid State Div. Somerville, New Jersey	05276 ITT Pomona Electronics Div. Pomona, California	06743 Gould Inc. Foil Div. Eastlake, Ohio
00656 Acrovox Corp. New Bedford, Massachusetts	02799 Arco Electronics Inc. Chatsworth, California	05277 Westinghouse Elec. Corp. Semiconductor Div. Youngwood, Pennsylvania	06751 Components Inc. Semcor Div. Phoenix, Arizona
00686 Film Capacitors Inc. Passaic, New Jersey	03508 General Electric Co. Semiconductor Products & Batteries Auburn, New York	05397 Union Carbide Corp. Materials Systems Div. Cleveland, Ohio	06776 Robinson Nugent Inc. New Albany, Indiana
00779 AMP, Inc. Harrisburg, Pennsylvania	03797 Genisco Technology Corp. Eltronics Div. Rancho Dominguez, Calif.	05571 Sprague Electric Co. (Now 56289)	06915 Richco Plastic Co. Chicago, Illinois
01121 Allen Bradley Co. Milwaukee, Wisconsin	03877 Gilbert Engineering Co. Inc Incon Sub of Transitron Electronic Corp. Glendale, Arizona	05574 Viking Connectors Inc Sub of Criton Corp. Chatsworth, Calif.	06961 Vernitron Corp. Piezo Electric Div. Bedford, Ohio
01281 TRW Electronics & Defense Sector Lawndale, California	03888 KDI Electronics Inc. Pyrofilm Div. Whippany, New Jersey	05820 EG & G Wakefield Engineering Wakefield, Massachusetts	06980 Varian Associates Inc. Eimac Div. San Carlos, California
01295 Texas Instruments Inc. Semiconductor Group Dallas, Texas	03911 Clairex Corp. Clairex Electronics Div. Mount Vernon, New York	05972 Loctite Corp. Newington, Connecticut	07047 Ross Milton Co., The Southampton, Penna.
01537 Motorola Communications & Electronics Inc. Franklin Park, Illinois	03980 Muirhead Inc. Mountainside, New Jersey	06001 General Electric Co. Electric Capacitor Product Section Columbia, S. Carolina	07138 Westinghouse Electric Corp. Industrial & Government Tube Div. Horsesheds, New York
01686 RCL Electronics/Shallcross Inc. Electro Components Div. Manchester, New Hampshire	04009 Cooper Industries, Inc. Arrow Hart Div. Hartford, Connecticut	06141 Fairchild Weston Systems Inc. Data Systems Div. Sarasota, Florida	07233 Benchmark Technology Inc. City of Industry, Calif.
01884 Sprague Electric Co. (Now 56289)	04217 Essex International Inc. Wire & Cable Div. Anaheim, California	06192 La Deau Mfg. Co. Glendale, California	07239 Biddle Instruments Bluc Bell, Penna.
01961 Varian Associates Inc. Pulse Engineering Div. Convoy, Connecticut	04221 Midland-Ross Corp. Midtex Div. N. Mankato, Minnesota	06229 Electrovert Inc. Elmsford, New York	07256 Silicon Transistor Corp. Sub of BBF Inc. Chelmsford, Massachusetts
02111 Spectrol Electronics Corp. City of Industry, California	04222 AVX Corp. AVX Ceramics Div. Myrtle Beach, S. Carolina	06383 Panduit Corp. Tinley Park, Illinois	07261 Avnet Corp. Culver City, California
02114 Ampere Electronic Corp. Ferrox Cube Div. Saugerties, New York	04423 Telonic Berkley Inc. Laguna Beach, California	06473 Bunker Ramo Corp. Amphenol NA Div. SAMS Operation Chatsworth, California	07263 Fairchild Camera & Instrument Semiconductor Div. Mountain View, California
02131 General Instrument Corp. Government Systems Div. Westwood, Massachusetts		06555 Beede Electrical Instrument Penacook, New Hampshire	07344 Bircher Co. Inc., The Rochester, New York
02395 Sonar Radio Corp. Hollywood, Florida			

Federal Supply Codes for Manufacturers (cont)

07557 Campion Co. Inc. Philadelphia, Penna.	09423 Scientific Components Inc. Santa Barbara, California	11711 General Instrument Corp. Rectifier Div. Hicksville, New York	12954 Microsemi Corp. Components Group Scottsdale, Arizona
07597 Burndy Corp. Tape/Cable Div. Rochester, New York	09579 CTS of Canada, Ltd Streetsville, Ontario	11726 Qualidyne Corp. Santa Clara, California	12969 Unitrode Corp. Lexington, Massachusetts
07716 TRW Inc. (Can use 11502) IRC Fixed Resistors/ Burlington Burlington, Iowa	09922 Burndy Corp. Norwalk, Connecticut	12014 Chicago Rivet & Machine Co. Naperville, Illinois	13050 Potter Co. Wesson, Mississippi
07792 Lenma Engineering Corp. Northampton, Massachusetts	09969 Dale Electronics Inc. Yankton, South Dakota	12040 National Semiconductor Corp. Danbury, Connecticut	13103 Thermalloy Co., Inc. Dallas, Texas
07810 Bock Corp. Madison, Wisconsin	09975 Burroughs Corp. Electronics Components Detroit, Michigan	12060 Diodes Inc. Northridge, California	13327 Solitron Devices Inc. Tappan, New York
07933 Raytheon Co. Semiconductor Div. Mountain View, Calif.	10059 Barker Engineering Corp. Kenilworth, New Jersey	12136 PHC Industries Inc. Formerly Philadelphia Handle Co. Camden, New Jersey	13511 Bunker-Ramo Corp. Amphenol Cadre Div. Los Gatos, California
08235 Industro Transistor Corp. Long Island City, New York	10389 Illinois Tool Works Inc. Licon Div. Chicago, Illinois	12300 AMF Canada Ltd. Potter-Brumfield Guelph, Ontario, Canada	13606 Sprague Electric Co. (Use 56289)
08261 Spectra-Strip An Eltra Co. Garden Grove, Calif.	10582 CTS of Asheville Skyland, N. Carolina	12323 Practical Automation Inc. Shelton, Connecticut	13689 SPS Technologies Inc. Hatfield, Pennsylvania
08530 Reliance Mica Corp. Brooklyn, New York	11236 CTS Corp. Beme Div. Beme, Indiana	12327 Freeway Corp. Cleveland, Ohio	13919 Burr-Brown Research Corp. Tucson, Arizona
08718 ITT Cannon Electric Phoenix Div. Phoenix, Arizona	11237 CTS Corp of California Paso Robles Div. Paso Robles, California	12443 Budd Co., The Plastics Products Div. Phoenixville, Pennsylvania	14099 Semtech Corp. Newbury Park, California
08806 General Electric Co. Miniature Lamp Products Cleveland, Ohio	11295 ECM Motor Co. Schaumburg, Illinois	12581 Hitachi Metals International Ltd. Hitachi Magna-Lock Div. Big Rapids, Missouri	14140 McGray-Edison Co. Commercial Development Div. Manchester, New Hampshire
08863 Nylomatic Fallsington, Penna.	11358 Columbia Broadcasting System CBS Electronic Div. Newburyport, Massachusetts	12615 US Terminals Inc. Cincinnati, Ohio	14193 Cal-R-Inc. Santa Monica, California
08988 Skottie Electronics Inc. Archbald, Pennsylvania	11403 Vacuum Can Co. Best Coffee Maker Div. Chicago, Illinois	12617 Hamlin Inc. Lake Mills, Wisconsin	14298 American Components Inc. an Insilco Co. RPC Div. Conshohocken, Pennsylvania
09021 Aircro Inc. Aircro Electronics Bradford, Penna.	11502 TRW Inc. TRW Resistive Products Div. Boone, North Carolina	12697 Clarostat Mfg. Co. Inc. Dover, New Hampshire	14298 ACIC Inc. Sub of Insilco Corp. Research Triangle Park, NC
09023 Cornell-Dublier Electronics Fuquay-Varina, N. Carolina	11503 Keystone Columbia Inc. Freemont, Indiana	12749 James Electronic Inc. Chicago, Illinois	14329 Wells Electronics Inc. South Bend, Indiana
09214 General Electric Co. Semiconductor Products Dept. Auburn, New York	11532 Teledyne Relays Teledyne Industries Inc. Hawthorne, California	12856 MicroMetals Inc. Anaheim, California	14482 Watkins-Johnson Co. Palo Alto, California
09353 C and K Components Inc. Newton, Massachusetts		12881 Metex Corp. Edison, New Jersey	14552 Microsemi Corp. Santa Ana, California
		12895 Cleveland Electric Motor Co. Cleveland, Ohio	14655 Cornell-Dublier Electronics Div. of Federal Pacific Electric Co. Govt Cont Dept. Newark, New Jersey

Federal Supply Codes for Manufacturers (cont)

14704 Crydom Controls (Division of Int Rectifier) El Segundo, California	16733 Cablewave Systems Inc. North Haven, Connecticut	18927 GTE Products Corp. Precision Material Products Business Parts Div. Tinsville, Pennsylvania	23936 William J. Pardy Co. Pamotor Div. Burlingame, California
14752 Electro Cube Inc. San Gabriel, California	16742 Paramount Plastics Fabricators Inc. Downey, California	19315 Bendix Corp., The Navigation & Control Group Teterboro, New Jersey	24347 Penn Engineering Co. S. El Monte, California
14936 General Instrument Corp. Discrete Semi Conductor Div. Hicksville, New York	16758 General Motors Corp. Delco Electronics Div. Kokomo, Indiana	19451 Perine Machinery & Supply Co. Kent, Washington	24355 Analog Devices Inc. Norwood, Massachusetts
14949 Trompeter Electronics Chatsworth, California	17069 Circuit Structures Lab Burbank, California	19613 Minnesota Mining & Mfg. Co. Textool Products Dept. Electronic Product Div. Irving, Texas	24444 General Semiconductor Industries, Inc. Tempe, Arizona
15412 Amtron Midlothian, Illinois	17117 Electronic Molding Corp. Woonsocket, Rhode Island	19647 Caddock Electronics Inc. Riverside, California	24655 Genrad Inc. Concord, Massachusetts
15542 Scientific Components Corp. Mini-Circuits Laboratory Div. Brooklyn, New York	17338 High Pressure Eng. Co. Inc. Oklahoma City, Oklahoma	19701 Mepco/Centralab Inc. A N. American Philips Co. Mineral Wells, Texas	24759 Lenox-Fugle Electronics Inc. South Plainfield, New Jersey
15636 Elec-Trol Inc. Saugus, California	17545 Atlantic Semiconductors Inc. Asbury Park, New Jersey	20584 Enochs Mfg. Inc. Indianapolis, Indiana	24796 AMF Inc. Potter & Brumfield Div. San Juan Capistrano, Calif.
15782 Bausch & Lomb Inc. Graphics & Control Div. Austin, Texas	17745 Angstrom Precision, Inc. Hagerstown, Maryland	20891 Cosar Corp. Dallas, Texas	24931 Specialty Connector Co. Greenwood, Indiana
15801 Fenwal Electronics Inc. Div. of Kidde Inc. Framingham, Massachusetts	17856 Siliconix Inc. Santa Clara, California	21317 Electronics Applications Co. El Monte, California	25088 Siemen Corp. Isilon, New Jersey
15818 Teledyne Inc. Co. Teledyne Semiconductor Div. Mountain View, California	18178 E G & Gvactee Inc. St. Louis, Missouri	21604 Buckeye Stamping Co. Columbus, Ohio	25099 Cascade Gasket Kent, Washington
15849 Useco Inc. (Now 88245)	18324 Signetics Corp. Sacramento, California	21845 Solitron Devices Inc. Semiconductor Group Riviera Beach, Florida	25403 Amperex Electronic Corp. Semiconductor & Micro-Circuit Div. Slatersville, Rhode Island
15898 International Business Machines Corp. Essex Junction, Vermont	18520 Sharp Electronics Corp. Paramus, New Jersey	22526 DuPont, El DeNemours & Co. Inc. DuPont Connector Systems Advanced Products Div. New Cumberland, Pennsylvania	25706 Daburn Electronic & Cable Corp. Norwood, New Jersey
16245 Conap Inc. Olean, New York	18542 Wabash Inc. Wabash Relay & Electronics Div. Wabash, Indiana	22767 ITT Semiconductors Palo Alto, California	26629 Frequency Sources Inc. Sources Div. Chelmsford, Massachusetts
16258 Space-Lok Inc. Burbank, California	18565 Chomerics Inc. Woburn, Massachusetts	22784 Palmer Inc. Cleveland, Ohio	26806 American Zettler Inc. Irvine, California
16352 Codi Corp. Linden, New Jersey	18612 Vishay Intertechnology Inc. Vishay Resistor Products Group Malvern, Pennsylvania	23050 Product Comp. Corp. Mount Vernon, New York	27014 National Semiconductor Corp. Santa Clara, California
16469 MCL Inc. LaGrange, Illinois	18632 Norton-Chemplast Santa Monica, California	23732 Tracor Applied Sciences Inc. Rockville, Maryland	27167 Coming Glass Works Coming Electronics Wilmington, North Carolina
16473 Cambridge Scientific Industries Div. of Chemed Corp. Cambridge, Maryland	18677 Scanbe Mfg. Co. Div. of Zero Corp. El Monte, California	23880 Stanford Applied Engineering Santa Clara, California	27264 Molex Inc. Lisle, Illinois
	18736 Voltronics Corp. East Hanover, New Jersey		27440 Industrial Screw Products Los Angeles, California

Federal Supply Codes for Manufacturers (cont)

27745 Associated Spring Barnes Group Inc. Syracuse, New York	30800 General Instrument Corp. Capacitor Div. Hicksville, New York	33297 NEC Electronics USA Inc. Electronic Arrays Inc. Div. Mountain View, California	49956 Raytheon Company Executive Offices Lexington, Massachusetts
27956 Relcom (Now 14482)		33919 Nortek Inc. Cranston, Rhode Island	50088 Thomson Components-Mostek Corp. Carrollton, Texas
28198 Positronic Industries Springfield, Missouri	31019 Solid State Scientific Inc. Willow Grove, Pennsylvania	34333 Silicon General Inc. Garden Grove, California	50120 Eagle-Picher Industries Inc. Electronics Div. Colorado Springs, Colorado
28213 Minnesota Mining & Mfg. Co. Consumer Products Div. 3M Center Saint Paul, Minnesota	31091 Alpha Industries Inc. Microelectronics Div. Hatfield, Pennsylvania	34225 Advanced Micro Devices Sunnyvale, California	50157 Midwest Components Inc. Muskegon, Mississippi
28425 Serv-O-Link Euless, Texas	31323 Metro Supply Company Sacramento, California	34359 Minnesota Mining & Mfg. Co. Commercial Office Supply Div. Saint Paul, Minnesota	50541 Hypertronics Corp. Hudson, Massachusetts
28478 Deltrol Corporation Deltrol Controls Div. Milwaukee, Wisconsin	31448 Army Safeguard Logistics Command Huntsville, Alabama	34371 Harris Corp. Harris Semiconductor Products Group Melbourne, Florida	50579 Litronix Inc. Cupertino, California
28480 Hewlett Packard Co. Corporate HQ Palo Alto, California	31746 Cannon Electric Woodbury, Tennessee	34649 Intel Corp. Santa Clara, California	51167 Aries Electronics Inc. Frenchtown, New Jersey
28484 Emerson Electric Co. Gearmaster Div. McHenry, Illinois	31827 Budwig Ramona, California	34802 Electromotive Inc. Kenilworth, New Jersey	51372 Verbatim Corp. Sunnyvale, California
28520 Heyco Molded Products Kenilworth, New Jersey	31918 ITT-Schadow Eden Prairie, Minnesota	34848 Hartwell Special Products Placentia, California	51406 Murata Eric, No. America Inc. (Also see 72982) Marietta, Georgia
29083 Monsanto Co. Santa Clara, California	32293 Intersil Cupertino, California	35009 Renfrew Electric Co. Ltd. IRC Div. Toronto, Ontario, Canada	51499 Amtron Corp. Boston, Massachusetts
29604 Stackpole Components Co. Raleigh, North Carolina	32539 Mura Corp. Westbury, Long Island, N.Y.	36665 Mitel Corp. Kanata, Ontario, Canada	51605 CODI Semiconductor Inc. Kenilworth, New Jersey
29907 Omega Engineering Inc. Stamford, Connecticut	32767 Griffith Plastics Corp. Burlingame, California	37942 Mallory Capacitor Corp. Sub of Emhart Industries Indianapolis, Indiana	51642 Centre Engineering Inc. State College, Pennsylvania
30035 Jolo Industries Inc. Garden Grove, California	32879 Advanced Mechanical Components Northridge, California	39003 Maxim Industries Middleboro, Massachusetts	51791 Statek Corp. Orange, California
30146 Symbex Corp. Painesville, Ohio	32897 Murata Erie North America Inc. Carlisle Operations Carlisle, Pennsylvania	40402 Rodenstein Electronics Inc. Statesville, North Carolina	51984 NEC America Inc. Falls Church, Virginia
30148 AB Enterprise Inc. Ahoskie, North Carolina	32997 Boums Inc. Trimpot Div. Riverside, California	42498 National Radio Melrose, Massachusetts	52063 Exar Integrated Systems Sunnyvale, California
30161 Aavid Engineering Inc. Laconia, New Hampshire	33096 Colorado Crystal Corp. Loveland, Colorado	43543 Nytronics Inc.(Now 53342)	52072 Circuit Assembly Corp. Irvine, California
30315 Itron Corp. San Diego, California	33173 General Electric Co. Owensboro, Kentucky	44655 Ohmitc Mfg. Co. Skokie, Illinois	52152 Minnesota Mining & Mfg. Saint Paul, Minnesota
30323 Illinois Tool Works Inc. Chicago, Illinois	33246 Epoxy Technology Inc. Billerica, Massachusetts	49671 RCA Corp. New York, New York	52333 API Electronics Haugpage, Long Island, New York

Federal Supply Codes for Manufacturers (cont)

52361
Communication Systems
Piscataway, New Jersey

52525
Space-Lok Inc.
Lercio Div.
Burbank, California

52531
Hitachi Magnetics
Edmore, Missouri

52745
Timco
Los Angeles, California

52763
Stettner-Electronics Inc.
Chattanooga, Tennessee

52769
Sprague-Goodman Electronics Inc.
Garden City Park, New York

52771
Monitem Corp.
Amatrom Div.
Santa Clara, California

52840
Western Digital Corp.
Costa Mesa, California

53021
Sangamo Weston Inc.
(See 06141)

53217
Technical Wire Products Inc.
Santa Barbara, California

53342
Opt Industries Inc.
Phillipsburg, New Jersey

53944
Glow-Lite
Pauls Valley, Oklahoma

54294
Shallcross Inc.
Smithfield, North Carolina

54453
Sullins Electronic Corp.
San Marcos, California

54473
Matsushita Electric Corp.
(Panasonic)
Secaucus, New Jersey

54583
TDK
Garden City, New York

54869
Piber International Corp.
Arlington Heights, Illinois

54937
DeYoung Mfg.
Bellevue, Washington

54590
RCA Corp.
Electronic Components Div.
Cherry Hill, New Jersey

55026
American Gage & Machine Co.
Simpson Electric Co. Div.
Elgin, Illinois

55112
Plessey Capacitors Inc.
(Now 60935)

55261
LSI Computer Systems Inc.
Melville, New York

55285
Beroquist Co.
Minneapolis, Minnesota

55576
Synertek
Santa Clara, California

55680
Michicon/America/Corp.
Schaumburg, Illinois

56282
Utek Systems Inc.
Olathe, Kansas

56289
Sprague Electric Co.
North Adams, Massachusetts

56365
Square D Co.
Corporate Offices
Palatine, Illinois

56375
DAL Industries Inc.
Wescorp Div.
Mountain View, California

56481
Shugart Associates
Sub of Xerox Corp.
Sunnyvale, California

56708
Zilog Inc.
Campbell, California

56856
Vamistor Corp. of Tennessee
Sevierville, Tennessee

56880
Magnetics Inc.
Baltimore, Maryland

57026
Endicott Coil Co. Inc.
Binghamton, New York

57053
Gates Energy Products
Denver, Ohio

58014
Hitachi Magnalock Corp.
(Now 12581)

58104
Simco
Atlanta, Georgia

58474
Superior Electric Co.
Bristol, Connecticut

59124
KOA-Speer Electronics Inc.
Bradford, Pennsylvania

59640
Supertex Inc.
Sunnyvale, California

59660
Tusonix Inc.
Tucson, Arizona

59730
Thomas and Betts Corp.
Iowa City, Iowa

59831
Semtronics Corp.
Watchung, New Jersey

60395
Xicor Inc.
Milpitas, California

60399
Torin Engineered Blowers
Div. of Clevepak Corp.
Torrington, Connecticut

60705
Cera-Mite Corp.
(formerly Sprague)
Grafton, Wisconsin

60935
Westlake Capacitor Inc.
Tantalum Div.
Greencastle, Indiana

61804
M/A Com Inc.
Burlington, Massachusetts

61857
SAN-O Industrial Corp.
Bohemia, Long Island, NY

61935
Schurter Inc.
Petaluma, California

62351
Apple Rubber
Lancaster, New York

62793
Lear Siegler Inc.
Energy Products Div.
Santa Ana, California

63743
Ward Leonard Electric Co. Inc.
Mount Vernon, New York

64154
Lamb Industries
Portland, Oregon

64155
Linear Technology
Milpitas, California

64834
West M G Co.
San Francisco, Calif.

65092
Sangamo Weston Inc.
Weston Instruments Div.
Newark, New Jersey

65940
Rohm Corp & Whatney
Irvine, California

65964
Evox Inc.
Bannockburn, Illinois

66150
Entron Inc.
Winslow Teltronics Div.
Glendale, New York

66608
Bezing Industries
Fremont, California

70290
Almetal Universal Joint Co.
Cleveland, Ohio

70485
Atlantic India Rubber Works Inc.
Chicago, Illinois

70563
Amperite Company
Union City, New Jersey

70903
Belden Corp.
Geneva, Illinois

71002
Bimbach Co. Inc.
Farmingdale, New York

71034
Bliley Electric Co.
Erie, Pennsylvania

71183
Westinghouse Electric Corp.
Bryant Div.
Bridgeport, Connecticut

71400
Bussman Manufacturing
Div. McGraw-Edison Co.
St. Louis, Missouri

71450
CTS Corp.
Elkhart, Indiana

71468
ITT Cannon Div. of IIT
Fountain Valley, California

71482
General Instrument Corp.
Clare Div.
Chicago, Illinois

Federal Supply Codes for Manufacturers (cont)

71590 Mepco/Centralab A North American Philips Co. Fort Dodge, Iowa	73445 Amperex Electronic Corp. Hicksville, New York	75378 CTS Knights Inc. Sandwich, Illinois	79727 C - W Industries Southampton, Pennsylvania
71707 Coto Corp. Providence, Rhode Island	73559 Cadingswitch Inc. Hartford, Connecticut	75382 Kulka Electric Corp. (Now 83330) Mount Vernon, New York	79963 Zierick Mfg. Corp. Mount Kisco, New York
71744 General Instrument Corp. Lamp Div./Worldwide Chicago, Illinois	73586 Circle F Industries Trenton, New Jersey	75915 Tracor Littlefuse Des Plaines, Illinois	80009 Tektronix Beaverton, Oregon
71785 TRW Inc. Cinch Connector Div. Elk Grove Village, Illinois	73734 Federal Screw Products Inc. Chicago, Illinois	76854 Oak Switch Systems Inc. Crystal Lake, Illinois	80031 Mepco/Electra Inc. Morristown, New Jersey
71984 Dow Corning Corp. Midland, Michigan	73743 Fischer Special Mfg. Co. Cold Spring, Kentucky	77122 TRW Assemblies & Fasteners Group Fastener Div. Moutainside, New Jersey	80032 Ford Aerospace & Communications Corp. Western Development Laboratories Div. Palo Alto, California
72005 AMAX Specialty Metals Corp. Newark, New Jersey	73893 Microdot Mt. Clemens, Mississippi	77342 AMF Inc. Potter & Brumfield Div. Princeton, Indiana	80145 LFE Corp. AMC Process Control Div. Clinton, Ohio
72136 Electro Motive Mfg. Corp. Florence, South Carolina	73899 JFD Electronic Components Div. of Murata Erie Oceanside, New York	77542 Ray-O-Vac Corp Madison, Wisconsin	80183 Sprague Products (Now 56289)
72228 AMCA International Corp. Continental Screw Div. New Bedford, Massachusetts	73905 FL Industries Inc. San Jose, California	77638 General Instrument Corp. Rectifier Div. Brooklyn, New York	80294 Boums Instruments Inc. Riverside, California
72259 Nytronics Inc. New York, New York	73949 Guardian Electric Mfg. Co. Chicago, Illinois	77900 Shakeproof Lock Washer Co. (Now 78189)	80583 Hammerlund Mfg. Co. Inc. Paramus, New Jersey
72619 Amperex Electronic Corp. Dialight Div. Brooklyn, New York	74199 Quam Nichols Co. Chicago, Illinois	77969 Rubbercraft Corp. of CA Ltd. Torrance, California	80640 Computer Products Inc. Stevens-Arnold Div. South Boston, Mass.
72653 G C Electronics Co. Div. of Hydrometals Inc. Rockford, Illinois	74217 Radio Switch Co. Marlboro, New Jersey	78189 Illinois Tool Works Inc. Shakeproof Div. Elgin, Illinois	81073 Grayhill Inc. La Grange, Illinois
72794 Dzus Fastner Co. Inc. West Islip, New York	74306 Piezo Crystal Co. Div. of PPA Industries Inc. Carlisle, Pennsylvania	78277 Sigma Instruments Inc. South Braintree, Mass.	81312 Litton Systems Inc. Winchester Electronics Div. Watertown, Connecticut
72928 Gulton Industries Inc. Gudeman Div. Chicago, Illinois	74542 Hoyt Elect.Instr. Works Inc. Penacook, New Hampshire	78290 Struthers Dunn Inc. Pitman, New Jersey	81439 Therm-O-Disc Inc. Mansfield, Ohio
72982 Murata Erie N. America Inc. Erie, Pennsylvania	74840 Illinois Capacitor Inc. Lincolnwood, Illinois	78553 Eaton Corp. Engineered Fastener Div. Cleveland, Ohio	81483 International Rectifier Corp. Los Angeles, California
73138 Beckman Industrial corp. Helipot Div. Fullerton, California	74970 Johnson EF Co. Waseca, Minnesota	78592 Stoeger Industries South Hackensack, New Jersey	81590 Korry Electronics Inc. Seattle, Washington
73168 Fenwal Inc. Ashland, Massachusetts	75042 TRW Inc. IRC Fixed Resistors Philadelphia, Pennsylvania	79136 Waldes Kohinoor Inc. Long Island City, New York	81741 Chicago Lock Co. Chicago, Illinois
73293 Hughes Aircraft Co. Electron Dynamics Div. Torrance, California	75297 Litton Systems Kester Solder Div. Chicago, Illinois	79497 Western Rubber Co. Goshen, Indiana	82227 Airpax Corp. Cheshire Div. Cheshire, Connecticut
	75376 Kurz-Kaseh Inc. Dayton, Ohio		82240 Simmons Fastner Corp. Albany, New York

Federal Supply Codes for Manufacturers (cont)

82305 Palmer Electronics Corp. South Gate, California	84171 Arco Electronics Commack, New York	89536 John Fluke Mfg. Co., Inc. Evrcrt, Washington	91802 Industrial Devices Inc. Edgewater, New Jersey
82389 Switchcraft Inc. Sub of Raytheon Co. Chicago, Illinois	84411 American Shizuki TRW Capacitors Div. Ogallala, Nebraska	89597 Fredericks Co. Huntingdon Valley, Penna.	91833 Keystone Electronics Corp. New York, New York
82415 Airpax Corp Frederick Div. Frederick, Maryland	84613 FIC Corp. Rockville, Maryland	89709 Bunker Ramo-Eltra Corp. Amphenol Div. Broadview, Illinois	91836 King's Electronics Co. Inc. Tuckahoe, New York
82872 Roanwell Corp. New York, New York	84682 Essex Group Inc. Peabody, Massachusetts	89730 General Electric Lamp Div. Newark, New Jersey	91929 Honeywell Inc. Micro Switch Div. Freeport, Illinois
82877 Rotron Inc. Custom Div. Woodstock, New York	85367 Bearing Distributing Co. San Francisco, California	90201 Mallory Capacitor Co. Sub of Emhart Industries Inc. Indianapolis, Indiana	91934 Miller Electric Co. Woonsocket, Rhode Island
82879 ITT Royal Electric Div. Pawtucket, Rhode Island	85372 Bearing Sales Co. Los Angeles, California	90215 Best Stamp & Mfg. Co. Kansas City, Missouri	91984 Maida Development Co. Hampton, Virginia
83003 Varo Inc. Garland, Texas	85480 W. H. Brady Co. Industrial Product Milwaukee, Wisconsin	90303 Duracell Inc. Technical Sales & Marketing Bethel, Connecticut	91985 Norwalk Valve Co. S. Norwalk, Connecticut
83014 Hartwell Corp. Placentia, California	85932 Electro Film Inc. Valencia, California	91094 Essex Group Inc. Suflex/IWP Div. Newmarket, New Hampshire	92914 Alpha Wire Corp. Elizabeth, New Jersey
83055 Signalite Fuse Co. (Now 71744)	86577 Precision Metal Products Co. Peabody, Massachusetts	91247 Illinois Transformer Co. Chicago, Illinois	93332 Sylvania Electric Products Semiconductor Products Div. Woburn, Massachusetts
83058 TRW Assemblies & Fasteners Group Fasteners Div. Cambridge, Massachusetts	86684 Radio Corp. of America (Now 54590)	91293 Johanson Mfg. Co. Boonton, New Jersey	94144 Raytheon Co. Microwave & Power Tube Div. Quincy, Massachusetts
83259 Parker-Hannifin Corp. O-Seal Div. Culver City, California	86928 Seastrom Mfg. Co. Inc. Glendale, California	91462 Alpha Industries Inc. Logansport, Indiana	94222 Southco Inc. Concordville, Pennsylvania
83298 Bendix Corp. Electric & Fluid Power Div. Eatonville, New Jersey	87034 Illuminated Products Inc. (Now 76854)	91502 Associated Machine Santa Clara, California	94988 Wagner Electric Corp. Sub of McGraw-Edison Co. Whippany, New Jersey
83315 Hubbell Corp. Mundelein, Illinois	88219 GNB Inc. Industrial Battery Div. Langhorne, Pennsylvania	91506 Augat Inc. Attleboro, Massachusetts	95146 Alco Electronic Products Inc. Switch Div. North Andover, Massachusetts
83330 Kulka Smith Inc. A North American Philips Co. Manasquan, New Jersey	88245 Winchester Electronics Liton Systems-Useco Div. Van Nuys, California	91507 Froeliger Machine Tool Co. Stockton, California	95263 Leecraft Mfg. Co. Long Island City, New York
83478 Rubbercraft Corp. of America West Haven, Connecticut	88486 Triangle PWC Inc. Jewitt City, Connecticut	91637 Dale Electronics Inc. Columbus, Nebraska	95275 Vitramon Inc. Bridgeport, Connecticut
83553 Associated Spring Bames Group Gardena, California	88690 Essex Group Inc. Wire Assembly Div. Dearborn, Michigan	91662 Eleo Corp. A Gulf Western Mfg. Co. Connector Div. Huntingdon, Pennsylvania	95303 RCA Corp. Receiving Tube Div. Cincinnati, Ohio
83740 Union Carbide Corp. BATTERY PRODUCTS Div. Danbury, Connecticut	89020 Amerace Corp. Buchanan Crimptool Products Div. Union, New Jersey	91737 ITT Cannon/Gretnar (Now 08718)	95348 Gordo's Corp. Bloomfield, New Jersey
	89265 Power-Brumfield (See 77342)		95354 Methode Mfg. Corp. Rolling Meadows, Illinois

Federal Supply Codes for Manufacturers (cont)

95573
Campion Laboratories Inc.
Detroit, Michigan

95712
Bendix Corp.
Electrical Comp. Div.
Franklin, Indiana

95987
Weckesser Co. Inc.
(Now 85480)

96733
SFE Technologies
San Fernando, California

96853
Gulton Industries Inc.
Measurement & Controls Div.
Manchester, New Hampshire

96881
Thomson Industries Inc.
Port Washington, New York

97525
EBCO Inc.
Santa Ana, California

97540
Whitehall Electronics Corp.
Master Mobile Mounts Div.
Fort Meyers, Florida

97913
Industrial Electronic
Hardware Corp.
New York, New York

97945
Pennwalt Corp.
SS White Industrial Products
Fiscataway, New Jersey

97966
CBS
Electronic Div.
Danvers, Massachusetts

98094
Machlett Laboratories Inc.
Santa Barbara, California

98159
Rubber-Teck Inc.
Gardena, California

98278
Malco A Microdot Co.
South Pasadena, California

98291
Sealectro Corp.
BICC Electronics
Trumbull, Connecticut

98372
Royal Industries Inc.(Now 62793)

98388
Lear Siegler Inc.
Accurate Products Div.
San Deigo, California

99120
Plastic Capacitors Inc.
Chicago, Illinois

99217
Bell Industries Inc.
Elect. Distributor Div.
Sunnyvale, California

99378
ATLEE of Delaware Inc.
N. Andover, Massachusetts

99392
Mepco/Electra Inc.
Roxboro Div.
Roxboro, North Carolina

99515
Electron Products Inc.
Div. of American Capacitors
Duarte, California

99779
Bunker Ramo- Eltra Corp.
Barnes Div.
Lansdown, Pennsylvania

99800
American Precision Industries
Delevan Div.
East Aurora, New York

99942
Mepco/Centralab
A North American Philips Co.
Milwaukee, Wisconsin

U.S. SALES OFFICE AREAS

AL, Huntsville

John Fluke Mfg. Co., Inc.
4920J Corporate Drive
Huntsville, AL 35805-6202
(205) 837-0581

AZ, Phoenix

John Fluke Mfg. Co., Inc.
2211 S. 48th Street
Suite B
Tempe, AZ 85282
(602) 438-8314

Tucson

(602) 790-9881

CA, Southern

John Fluke Mfg. Co., Inc.
P.O. Box 19676
Irvine, CA 92713-9676
16969 Von Karman
Suite 100
Irvine, CA 92714
(714) 863-9031

Burbank

John Fluke Mfg. Co., Inc.
2020 N. Lincoln Street
Burbank, CA 91504
(213) 849-7181

Northern

John Fluke Mfg. Co., Inc.
2300 Walsh Ave., Bldg. K
Santa Clara, CA 95051
(408) 727-0513

San Diego

(619) 292-7656

CO, Denver

John Fluke Mfg. Co., Inc.
14180 East Evans Ave.
Aurora, CO 80014
(303) 695-1000

CT, Hartford

John Fluke Mfg. Co., Inc.
Glen Locken East
41-C New London Turnpike
Glastonbury, CT 06033
(203) 659-3541

DC, Washington

(301) 770-1570

FL, Clearwater

(813) 799-0087

Miami

(305) 462-1380

Orlando

John Fluke Mfg. Co., Inc.
940 N. Fern Creek Ave.
Orlando, FL 32803
(305) 896-4881

Tampa

(813) 251-9211

GA, Atlanta

John Fluke Mfg. Co., Inc.
2700 Delk Rd., Suite 150
Marietta, GA 30067
(404) 953-4747

IL, Chicago

John Fluke Mfg. Co., Inc.
1150 West Euclid Ave.
Palatine, IL 60067
(312) 705-0500

IN, Indianapolis

John Fluke Mfg. Co., Inc.
8777 Purdue Rd.
Suite 101
Indianapolis, IN 46268
(317) 875-7870

LA, Baton Rouge

(504) 924-1203

New Orleans

(504) 455-0814

MA, Boston

John Fluke Mfg. Co., Inc.
Middlesex Tech Center
900 Middlesex Turnpike
Building 8
Billerica, MA 01821
(617) 863-2400

MD, Baltimore

(301) 792-7060

Rockville

John Fluke Mfg. Co., Inc.
5640 Fishers Lane
Rockville, MD 20852
(301) 770-1570

MI, Detroit

John Fluke Mfg. Co., Inc.
33031 Schoolcraft
Livonia, MI 48150
(313) 522-9140

MN, Bloomington

John Fluke Mfg. Co., Inc.
1801 E. 79th St., Suite 9
Bloomington, MN 55420
(612) 854-5526

MO, St. Louis

John Fluke Mfg. Co., Inc.
11756 Borman Drive
Suite 160
St. Louis, MO 63146
(314) 993-3805

NC, Greensboro

John Fluke Mfg. Co., Inc.
1310 Beaman Place
Greensboro, NC 27408
(919) 273-1918

NJ, Paramus

John Fluke Mfg. Co., Inc.
P.O. Box 930
Paramus, NJ 07653-0930
West 75 Century Road
Paramus, NJ 07652
(201) 262-9550

NM, Albuquerque

(505) 881-3550

NY, Rochester

John Fluke Mfg. Co., Inc.
4515 Culver Road
Rochester, NY 14622
(716) 323-1400

OH, Cleveland

John Fluke Mfg. Co., Inc.
7830 Freeway Circle
Middleburg Heights, OH 44130
(216) 234-4540

Dayton

John Fluke Mfg. Co., Inc.
5450 Far Hills Avenue
Suite 111
Kettering, OH 45429
(513) 436-2224

OK, Northeast

(918) 749-0190

OR, Portland

(503) 227-2042

PA, Philadelphia

John Fluke Mfg. Co., Inc.
200 Lindenwood Drive
Malvern, PA 19355
(215) 647-9550

Pittsburgh

(412) 261-5171

TX, Austin

(512) 459-3344

Dallas

John Fluke Mfg. Co., Inc.
1801 Royal Lane
Suite 307
Dallas, TX 75229
(214) 869-0311

El Paso

(915) 533-3508

Houston

(713) 240-5995

San Antonio

John Fluke Mfg. Co., Inc.
10417 Gulfdale
San Antonio, TX 78216
(512) 340-0498

UT, Salt Lake City

(801) 268-9331

WA, Seattle

John Fluke Mfg. Co., Inc.
5020 148th Ave. N.E.
Suite #110
Redmond, WA 98052
(206) 881-6966

U.S. Government Specialists

Army (301) 770-1570, MD
(203) 837-0581, AL

Navy (714) 863-9031, CA
(301) 770-1570, MD

USAF (513) 436-2224, OH
(512) 340-2621, TX

Security (301) 770-1570, MD

Service Center Areas

CA, Burbank (213) 849-4641

CA, Santa Clara (408) 727-0513

CO, Denver (303) 695-1000

FL, Orlando (305) 896-4881

IL, Chicago (312) 705-0500

MD, Rockville (301) 770-1576

NJ, Paramus (201) 262-9550

TX, Dallas (214) 869-2848

WA, Everett (206) 356-5560

For more information on Fluke products or Sales Offices you may dial (800) 426-0361 toll-free in most of the U.S.A.
From Alaska, Hawaii, or Washington phone (206) 356-5400. From Canada and other countries phone (206) 356-5500.



John Fluke Mfg. Co., Inc., P.O. Box C9090, Everett, WA 98206
Fluke (Holland) B.V., P.O. Box 2269, 5600 CG, Eindhoven, The Netherlands. Phone (040) 458045
Litho in U.S.A. 11/87

INTERNATIONAL SALES OFFICES

Argentina •

Crasin S.A.
Virrey del Pino 4071 DPTO E-66
1430 CAP FED
Buenos Aires, Argentina
Tel: 54-1-532-5248
TLX: (390) 22284 COASN AR

Asia

Fluke Asia Ltd.
Shun Tak Centre, Room 1501
200 Cunnaught Road
Central, Hong Kong
Tel: 852-5-482116
TLX: (780) 87058 FLUKE
FAX: (852) 5-479863

Australia •

Elmeasco Instruments Pty, Ltd.
P.O. Box 30
Concord, N.S.W. 2137
Australia
Tel: 61-2-736-2888
TLX: (790) 25887 A/B: ELSCOAA 25887
FAX: 61-2-733663

Elmeasco Instruments Pty, Ltd.
P.O. Box 623
12 Maroondah Highway
Ringwood, Victoria 3134
Australia
Tel: 61-3-879-2322
TLX: (790) 30418 A/B: ELTENTAA 30418
FAX: (61) (3) 879-4310

Elmeasco Instruments Pty, Ltd.
P.O. Box 274
Salisbury, Queensland 4107
Australia
Tel: 61-7-875-1444
TLX: (790) 44062 A/B: ELMQLDAA44062

Elmeasco Instruments Pty, Ltd.
P.O. Box 154
Prospect, South Australia 5082
Tel: 61-8-344-9000 TLX: (790) A/B:87519

Elmeasco Instruments Pty, Ltd.
P.O. Box 413
Scott House
46-48 Kings Park Road
West Perth, Western Australia 6005
Australia
Tel: 61-9-481-1500
TLX: (790) 94765 A/B: ASECSAA 94765
FAX: (61) (9) 322-2075

Austria ■

Walter Rekirsch Elektronische
Geräte GmbH & Co.
Vertrieb KG
Obachgasse 28
1220 Vienna, Austria
Tel: 43-222-25-36-26
TLX: (847) 134759
FAX: 43-222-25-72-75

Bahrain •

Basma W.L.L.
P.O. Box 5701
Manama, Bahrain
Tel: 973-251364; TLX: (955) 9003
FAX: (965) 245218

Bangladesh •

Motherland Corporation
24 Hatkhola Road, Tikatuli
Dacca-3, Bangladesh
Tel: 257249
TLX: (950) 642022 PCO BJ
Cable: "MOTHERLAND" DACCA

Belgium ■

N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

Bolivia •

Coasin Bolivia S.R.L.
Casilla 7295
La Paz, Bolivia
Tel: 591-2-40962
TLX: (336) 3233 COALAP BV
Cable: COALAP

Brazil •

ATP/Hi-Tek Eletronica Ltda
Al. Amazonas 422,
Alphaville, 06400, Barueri
Sao Paulo, Brazil
Tel: 55-11-421-5477;
TLX: (391) 1171413 HITK BR

Brunei •

Rank O'Connor's SDN BHD
No. 8 Blok D
Sulfi Shophouse Complex
Mile 1 Jalan Tutong
Bandar Seri Begawan
Negara Brunei Darussalam
Tel: 673-2-23109 or 23557
TLX: (799) BU 2265 RANKOC

Canada •

Fluke Electronics Canada Inc.
101, 1144 - 29th Avenue N.E.
Calgary, Alberta T2E 7P1
Canada
Tel: (403) 291-5215
Fax: (403) 291-5219

Fluke Electronics Canada Inc.
400 Britannia Road East
Unit #1
Mississauga, Ontario L4Z 1X9
Tel: (416) 890-7600
Fax: (416) 890-6866

Fluke Electronics Canada Inc.
1255 Trans Canada Hwy.
Suite 130
Dorval, Quebec
H9P 2V4 Canada
Tel: (514) 685-0022
TLX: (514) 685-0039

Fluke Electronics Canada Inc.
1690 Woodward Drive
Suite 216
Ottawa, Ontario
K2C 3R8 Canada
Tel: (613) 723-9453
Fax: (613) 723-9458

Chile •

Intronica Chile, Ltda.
Casilla 16228 (Mail)
Santiago 9, Chile
Tel: 56-2-2321886
TLX: (332) 346351 INTRON CK

China, Peoples Republic of •

Fluke International Corporation
P.O. Box 9065
Beijing
People's Republic of China
Tel: 86-01-65-7281
TLX: (716) 222237 FBSC CN

Instrimpex - Fluke Service Center
57, Xisi Dong Da Jie
Xicheng-qu
Beijing
Peoples Republic of China
Tel: 86-01-65-7281

Colombia ■

Sistemas E Instrumentacion, Ltda.
Carrera 13, No. 37-43, Of. 401
Ap. Aereo 29583
Bogota DE, Colombia
Tel: 57-232-4532
TLX: (396) 45787 COASN CO

Cyprus ■

Chris Radiovision, Ltd.
P.O. Box 1989
Nicosia, Cyprus
Tel: 357-21-66121; TLX: (826) 2395

Cyprus, Northern ■

Ucok Buroteknik
2C & 2D Muftu Ziyai Street
Lefkosa, Northern Cyprus
Mersin 10, Turkey
Tel: 90-741-357-20-71777
TLX: (821) 57267

Czechoslovakia ■

Amtest Associates, Ltd.
Amtest House
75-79 Guildford Street
Chertsey, Surrey KT16 9AS
England
Tel: 44-9328-68355
TLX: (851) 928855 AMTEST G
Fax: 44-9328-61919

Denmark ■

Tage Olsen A/S
Ballerup Byvej 222
2750 Ballerup
Denmark
Tel: 45-2-656111
TLX: (855) 35293 TOAS DK

Eastern European Countries ■

Amtest Associates, Ltd.
Amtest House
75-79 Guildford Street
Chertsey, Surrey KT16 9AS
England
Tel: 44-9328-68355
TLX: (851) 928855 AM*EST G
Fax: 44-9328-61919

German Branch Office
Amtest Associates, Ltd.
Zugspitzstrasse 2A
P.O. Box 1107
8011 Vaterstetten
West Germany
Tel: 49-81-067117; TLX: (841) 528332

Ecuador •

Proteco Coasin Cia., Ltda.
P.O. Box 228-A
Ave. 12 de Octubre 2285
y Orellana
Quito, Ecuador
Tel: 593-2-529684
TLX: (393) 22085 ESIND

Egypt

Electronic Engineering Liaison Office
P.O. Box 2891 Horreya
11361 Heliopolis, Cairo
Egypt
Tel: 20-2-695705; TLX: (927) 22782

England

N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

Fiji •

Awa New Zealand Limited
37 Freeston Road
Walu Bay, P.O. Box 858
Suva, Fiji
Tel: 679-312079; TLX: (792) FJ2347
FAX: 679-314379

Finland ■

Instrumentarium Elektronika
P.O. Box 64
02631 Espoo 63
Finland
Tel: 356-0-5281
TLX: (857) 124426 HAVUL SF
FAX: (358) 0-5021073
Teletex: (857) 8-100155 INSTRUE

France ■

M.B. Electronique S.A.
606, Rue Fournay
P.O. Box 31
78530 BUC, France
Tel: 33-1-39-56-81-31
TLX: (842) 695414
Fax: (33) (1) 3956-53-44

German Branch Office

Amtest Associates, Ltd.
Zugspitzstrasse 2A
P.O. Box 1107
8011 Vaterstetten
West Germany
Tel: 49-81-067117
TLX: (841) 528332

Germany, West

N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

Greece

Heflonio Scientific Representations Ltd.
11, Vrassida Street
Athens 612, Greece
Tel: 30-1-7211140; TLX: (863) 219330

Hong Kong •

Schmidt & Co (H.K.), Ltd
18th Floor, Great Eagle Centre
23 Harbour Road
Wanchai, Hong Kong
Tel: 852-5-8330-222
TLX: (780) 74766 SCHMC HK
FAX: 852-5-8918754

India •

Hinditron Services Pvt., Ltd.
69/A.L. Jagmohandas Marg
Bombay 400 006, India
Tel: 91-22-8121316, 8125344
TLX: (953) 1175326 HSPL IN

Bangalore Office
Hinditron Services Pvt., Ltd.
8th Main Road
33/44A Raj Mahal Vias Extension
Bangalore 560 080, India
Tel: 91-812-363139
TLX: (953) 08452741
Cable: TEKHIND BANGALORE

Calcutta Office
Hinditron Services Pvt. Ltd.
5th Floor, Castle House
5/1A, Hungerford St.
Calcutta 700 017, India
Tel: 91-33-432628
TLX: (953) 214153

New Delhi Sales
Hinditron Services Pvt. Ltd.
204-206 Hemkunt Tower
98 Nehru Place
New Delhi, 110019, India
Tel: 91 (11) 6410380 or 6414493
TLX: (953) 3161458 HSPL IN

New Delhi Service
Hinditron Services Pvt., Ltd.
Field Service Center
15, Community Centre
Panchshila Park
New Delhi 110 017, India
Tel: 91 011 6433675

Cable: Tekcentre Delhi
Hyderabad Office
Hinditron Services Pvt. Ltd.
Field Service Center
Emerald Complex, 5th Floor
1-7-264
114 Sarojini Devi Road
Secunderabad 500 003, India
Tel: 91 842 821117
TLX: (953) 04256973 HSPL IN

Indonesia •

P.T. Lamda Triguna
P.O. Box 6/JATJG
Jakarta, 13001
Indonesia
Tel: 62 21 8195365
TLX: (796) 46171 LAYARIA

Iran ■

Arma Baynalmellefi Co., Ltd.
P.O. Box 951570
Pasdaran-Tehran
Tel: 98-21-248717; TLX: (951) 213648

Ireland

N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

Israel ■

R.D.T. Electronics Engineering Ltd.
P.O. Box 43137
Tel Aviv 61430
Israel
Tel: (3) 972-3-483211
TLX: (922) 371452 RDT IL
Fax: 972-3-492190

Italy ■

Sistrel S.p.A.
Via le Erminio Spalla 41
00142 Rome, Italy
Tel: 39-6-504-1367; TLX: (843) 625857
Fax: (39) 6-504137



John Fluke Mfg. Co., Inc. / PO Box C9090 / Everett, WA 98206 / (206) 356 5400

Litho in U.S.A. 11/87

Milan Office
Sistrel S.p.A.
Via Pelizza da Volpedo 59
20092 Cinisello Balsamo
Milan, Italy
Tel: 39-2-6181893, TLX: (843) 334643
Fax: (39) 2-6182440

Naples Office
Sistrel S.p.A.
Via Cintia
Parco S. Paolo 35
80126 Naples, Italy
Tel: 39-81-7678700
Fax: (39) 81-7661361

Japan ●
John Fluke Mfg. Co., Inc.
Japan Branch
Sumitomo Higashi Shinbashi Bldg.
1-1-11 Hamamatsucho
Minato-ku, Tokyo 105, Japan
Tel: 81-3-434-0181
TLX: (781) 2424331 FLUK JPY
FAX: 81-3-434-0170

Osaka Sales Office
John Fluke Mfg. Co., Inc.
Japan Branch
Katsushige Building
2-45 Kohraibashi
Higashi-ku, Osaka 541
Japan
Tel: 81-6-229-0871 FAX: 81-6-229-1098

Korea, Republic of ●
Myoung Corporation
Yeo Eui Do P.O. Box 14
Seoul, Korea
Tel: 82-2-784-9942 MYOUNG
TLX: (787) K24283
FAX: (82) 2-784-2387

Kuwait ■
Al Bahar International Group
P.O. Box 26672 Safat
13127 Safat, Kuwait
Kuwait, Arabian Gulf
Tel: 965-848601, 847598
TLX: (959) 44822

Lebanon and Jordan ■
Mabek (Electronics Division)
P.O. Box 13-5657
Beirut, Lebanon
Tel: 812523
TLX: (923) 22889 LIBANK LE

Malaysia ●
Mecomb Malaysia Sdn Bhd
P.O. Box 24
46700 Petaling Jaya, Selangor, Malaysia
Tel: 60-3-774-3422
TLX: (784) MA37764 MECOMB
Fax: (6) 03-774-3414

Malta ■
Fabian Enterprises
20, Msida Road
Gzira, Malta
Tel: 513283/40216, TLX: (838) 1837

Mexico ●
Mexicana de Electronica
Industrial, S.A. (Mixel)
Diagonal No. 27
Entre Calle de Eugenia Y Ave
Colonia del Valle
C.P. 03100, Mexico
Tel: (905) 660-4323
TLX: (383) 1771038 FAIRME

Executone De Monterrey, S.A.
Ave. Gonzalitos NTE 545
Monterrey N.L., Mexico
Tel: 90-5-480400, 472625
TLX: (383) 382659

Morocco ■
Oussama S.A.
Angle Boulevard Emile Zola et
Rue Rethel
P.O. Box 2007 Casa
Casablanca
Morocco
Tel: 212-24-13-38, TLX: (933) 28879 M

Nepal ●
Associated Enterprises
GPO Box 790, Pyaphal Tole
Kathmandu, Nepal
Tel: 13868, TLX: (947) 2568 (ASOENT NP)

Netherlands
N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

New Zealand ●
Auckland Office
Northrop Instruments & Systems, Ltd.
459 Khyber Pass Road
Private Bag, Newmarket
Auckland 1, New Zealand
Tel: 64-9-501-801, 501-219
TLX: (791) 21570
FAX: 64-9-543430

Wellington Office
Northrop Instruments & Systems Ltd.
Information Technology Group
First Floor, Northrop Bldg.
189-191 Willis Street
P.O. Box 2406
Wellington, New Zealand
Tel: 64-4-856-658
TLX: (791) 3380

Christchurch Office
Northrop Instruments & Systems Ltd.
Information Technology Group
110 Mandeville Street
P.O. Box 8388
Christchurch, New Zealand
Tel: 64-3-488-874
TLX: (791) 4801

Norway ■
Morgensterne & Co A/S
Konghellegate 3
P.O. Box 6688, Rodelokka
Oslo 5, Norway
Tel: (2) 356110, TLX: (856) 71719

Oman ■
OHI Telecommunications LLC
P.O. Box 889
Muscat, Oman
Tel: 968-603606
TLX: (926) 5052 TELECOM ON

Pakistan ●
International Operations (PAK), Ltd.
505 Muhammadi House
I. Chundrigar Road
P.O. Box 5323, Karachi, Pakistan
Tel: 92-21-221127, TLX: (952) 24494 PIO PK

Peru ●
Importaciones y Representaciones
Electronicas S.A.
Avda. Franklin D. Roosevelt 105
Lima 1, Peru
Tel: 51-14-28-8650
TLX: (394) 25663 PE IREING

Philippines, Republic of ●
Spark Radio & Electronics, Inc.
Greenhills P.O. Box 610
San Juan, Metro Manila, Zip 3113
Philippines
Tel: 63-2-775192, 704096
TLX: (722 or 732) 27901 RLA PH

Portugal ■
Decada Espectral
Equipamentos de Electronica
Av. Bombeiros Voluntarios
Lote 102B, Miraflores/Alges
1495 Lisbon, Portugal
Tel: 351-1-4103420, TLX: (832) 15515

Romania ■
Amtest Associates Ltd.
Amtest House
75-79 Guildford Street
Chertsey, Surrey KT16 9AS
England
Tel: 44-9328-68355, TLX: (851) 928855
FAX: 44-9328-61919

Saudi Arabia ■
Electronic Equipment Marketing Co.
P.O. Box 3750
Riyadh, Saudi Arabia
Tel: 966-1-477-1650
TLX: (928) 401120 ZUHAIJR SJ

Scotland
N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

Singapore, Republic of ●
Frank O'Connor's (PTE) Ltd.
O'Connor House
98 Pasir Panjang Road
Singapore 0511
Republic of Singapore
Tel: 65-4737944, TLX: (786) RS21023 OCONSN
FAX: 4724508 CABLE CINECOM SINGAPORE

South Africa ●
Prolea PNI
P.O. Box 39127
Bramley 2018
Republic of South Africa
Tel: 27-11-786-3647
TLX: (960) 4-24409 SA
FAX 27-11-786-1807

Spain ■
ESSA
Equipos y Sistemas S.A.
C/Polonico Morales, 13-B
Madrid 16, Spain
Tel: 34-1-458-0150
TLX: (831) 42856 EYS E

Sri Lanka ●
Computerlink Data Systems, Ltd.
331 Union Place
Colombo 2, Sri Lanka
Tel: 94-1-28641
TLX: (954) 22455 COLINK CE

Sweden
N.V. Philips
Industrial & Electro-Acoustic Systems
5600 MD Eindhoven
The Netherlands
Tel: 040-785539
Tlx: 35000 phtc nl
Fax: 040-785651

Switzerland ■
Traco Electronic AG
Jenatschstrasse 1
8002 Zurich
Switzerland
Tel: 41-1-201-0711
TLX: (845) 815570 TRCOCH

Syria ■
Mabek (Electronics Division)
P.O. Box 4238
Damascus, Syria

Taiwan ●
Schmidt Electronics Corp.
5th Fl, Cathay Min Sheng
Commercial Building,
344 Min Sheng East Road
Taipei Taiwan R.O.C.
Tel: 886-2-501-3468
TLX: (785) 11111 SCHMDT
Fax: (886) 2-502-9692

Thailand ●
Measuretronix Ltd.
2102/63 Ramkhamhaeng Rd.
Huamark Bangkok 10240
Thailand
Tel: 66 (2) 3742516, 3741632
TLX: (788) 82796 HUAMARK TH

Tunisia ■
Selep S.A.R.L.
6, Rue de Sparte
Tunis - 1000 RP
Tunisia
Tel: 216-1-248093, TLX: (934) 13030

Turkey ■
Erkman Elektronik Aletler
Ticaret Anonim Sirketi
Necatibey Cad 92/3
Karakoy, Istanbul, Turkey
Tel: 90 (11) 441546, TLX: (821) 24399

United Arab Emirates ■
Harris Al-Afaq, Ltd.
P.O. Box 8141
Kamal Hamza Bldg.
Traffic Police Road
Dubai, U.A.E.
Tel: 971-4-283625
TLX: (958) 48168 AFAQEM

Al-Sanani Con. Trad. Est.
P.O. Box 7187
Abu-Dhabi, U.A.E.
Tel: 971-2-821370, TLX: (858) 23966

Uruguay ●
Coasin Uruguay S.A.
Casilla de Correo 1400
Montevideo 2529
Montevideo, Uruguay
Tel: 598-2-789204, 789015
TLX: (398) UY23010 COAUR

USSR
Amtest Associates Ltd.
Amtest House
75-79 Guildford Street
Chertsey, Surrey KT16 9AS
England
Tel: 44-9328-68355, TLX: (851) 928855
FAX: 44-9328-61919

Venezuela ●
Coasin C.A.
Calle 9 Con Calle 4, Edif Edinurbi
Apartado de Correos NR-70.136
Los Ruices
Caracas 1070-A, Venezuela
Tel: 58 (2) 241-03-09
TLX: (395) 21027 EMVEN VC

Yugoslavia ■
Amtest Associates Ltd.
Amtest House
75-79 Guildford Street
Chertsey, Surrey KT16 9AS
England
Tel: 44-9328-68355, TLX: (851) 928855
FAX: 44-9328-61919

■ Supplied and Supported by —
Fluke (Holland) B.V.
P.O. Box 2269
5600 CG Eindhoven
The Netherlands
Tel: (040) 45805, TLX: (844) 51846
FAX: 31-40-457515

● Supplied and Supported by —
Fluke International Corporation
P.O. Box C9090
Everett, WA 98206 U.S.A.
Tel: (206) 356-5500
Tel: 185103 FLUKE UT
FAX: 206-356-5116

The following countries are represented by:
Fluke (Holland) B.V.
P.O. Box 2269
5600 CG Eindhoven
The Netherlands
Tel: (040) 45805, TLX: 51846
FAX: 31-40-457515

Abu Dhabi	Mauritania
Afghanistan	Menorca
Albania	Mongolia
Algeria	Niger
Angola	Nigeria
Benin	Qatar
Bornholm	Rodhos
Botswana	Russia
Chad	Sardinia
Corsica	Saudia Arabia
Czechoslovakia	Scotland
Dubai	Senegal
Ethiopia	Sierra Leone
Guinea	Somalia
Ibiza	Togo
Iceland	Upper Volta
Iraq	Wales
Kuwait	Watin Sahara
Liberia	Yemen
Libya	Zaire
Luxembourg	Zambia
Maderia	Zimbabwe
Mallorca	

TECHNICAL SERVICE CENTERS

U.S.A.

CA, Burbank

John Fluke Mfg. Co., Inc.
Tel: (213) 849-4841

CA, Santa Clara

John Fluke Mfg. Co., Inc.
Tel: (408) 727-8121

CO, Denver

John Fluke Mfg. Co., Inc.
Tel: (303) 695-1000

FL, Orlando

John Fluke Mfg. Co., Inc.
Tel: (305) 896-4881

IL, Palatine

John Fluke Mfg. Co., Inc.
Tel: (312) 705-0500

MD, Rockville

John Fluke Mfg. Co., Inc.
Tel: (301) 770-1576

NJ, Paramus

John Fluke Mfg. Co., Inc.
Tel: (201) 262-9550

TX, Dallas

John Fluke Mfg. Co., Inc.
Tel: (214) 869-0311

WA, Seattle

John Fluke Mfg. Co., Inc.
Tel: (206) 356-5560

Other Countries

Argentina, Buenos Aires

Coasin S.A.
Tel: 552-5248
TLX: (390) 22284

Australia, Brisbane

Elmeasco Instruments Pty. Ltd.
Tel: 552-5248

Australia, Concord

Elmeasco Instruments Pty. Ltd.
Tel: 736-2888
TLX: (790) AA25887

Australia, Ringwood

Elmeasco Instruments Pty. Ltd.
Tel: 879-2322
TLX: (790) AA36206

Austria, Vienna

Walter Rekersch
Elektronische Gerate GmbH & Co.
Tel: 253626
TLX: (847) 134759

Belgium, Brussels

Fluke (Belgium) NV/SA
Tel: 2164090
TLX: (846) 26312

Brazil, Sao Paulo

Hi-Tek Electronica Ltda.
Tel: 421-5477
TLX: (1391) 1171413

Canada, Calgary

Fluke Electronics Canada Inc.
Tel: (403) 291-5215
Fax: (403) 291-5219

Canada, Mississauga

Fluke Electronics Canada Inc.
Tel: (416) 890-7600
Fax: (416) 890-6866

Canada, Montreal

Fluke Electronics Canada Inc.
Tel: (514) 685-0022
Fax: (514) 685-0039

Canada, Ottawa

Fluke Electronics Canada Inc.
Tel: (613) 723-9453
Fax: (613) 723-9458

Chile, Santiago

Intronica Chile Ltda.
Tel: 232 1886
TLX: (332) 346351

China, Peoples Republic of

Instrimpex - Fluke Service Center
Tel: 65-7281

Colombia, Bogota

Sistemas E Instrumentacion, Ltda.
Tel: 232-4532
TLX: (396) 45787

Denmark, Ballerup

Tage Olson A/S
Tel: 658111
TLX: (855) 35293

Ecuador, Quito

Proteco Coasin Cia., Ltda.
Tel: 526759
TLX: (393) 2865

Egypt and Sudan

Electronic Engineering Liaison Ofc
Tel: 2455705
TLX: (927) 22782

England, Chertsey, Surrey

Amtest Associates, Ltd
Tel: 68355

England, Watford, Herts

Fluke (Great Britain) Ltd
Tel: 40511
TLX: (851) 934583

Finland

Instrumentarium Elektronikka
Tel: 358-0-5281
TLX: (857) 124426

France

M.B. Electronique S.A.
Tel: 1-39-56-81-31
TLX: (842) 695414

Greece, Athens

Hellenic Scientific Representations
Tel: 7211140
TLX: (863) 219330

Hong Kong, Wanchai

Schmidt & Co (H.K.) Ltd.
Tel: 8330-222
TLX: (780) 74766

India, Bangalore

Hinditron Services Pvt. Ltd.
Tel: 363139
TLX: (953) 845741

India, Bombay

Hinditron Services Pvt. Ltd.
Tel: 6300043
TLX: (953) 11-72247

India, New Delhi

Hinditron Services Pvt. Ltd.
Tel: 6433675
TLX: (953) 316458

India, Secunderbad

Hinditron Services Pvt. Ltd.
Tel: 821117
TLX: (953) 1556973

Indonesia, Jakarta Pusat

P.T. Lamda Triguna
Tel: 8195365
TLX: 46171 LAYARIA

Israel, Tel Aviv

R.D.T. Electronics Engineering Ltd.
Tel: 483211
TLX: (922) 371452

Italy, Milan

Sistrel S.p.A.
Tel: 6181893
TLX: (843) 334643

Japan, Tokyo

John Fluke Mfg. Co., Inc.
Japan Branch
Tel: 434-0181
TLX: (781) 242-4331

Korea, Republic of

Myoung Corp.
Tel: 784-9942
TLX: MYOUNG K24283

Malaysia, Selangor

Mecomb Malaysia Sdn. Bhd.
Tel: 3-743422
TLX: (784) MA37764

Mexico

Mexicana de Electronica
Industrial, S.A. (Mexel)
Tel: 5-660-4323
TLX: (383) 1771038

Netherlands, Tilburg

Fluke (Nederland) B.V.
Tel: 352455
TLX: (844) 52683

New Zealand, Auckland

Northrop Instruments
& Systems Ltd.
Tel: 501-801
TLX: (791) NZ21570

New Zealand, Wellington

Northrop Instruments
& Systems Ltd.
Tel: 856-658
TLX: (791) NZ3380

Norway, Oslo

Morgenstjerne & Co. A/S
Tel: 356110
TLX: (856) 71719

Pakistan, Karachi

International Operations (PAK), Ltd.
Tel: 221127, 239052
TLX: (952) 24494

Peru, Lima

Importaciones Y Representaciones
Electronicas S.A.
Tel: 288650
TLX: (394) 25663

Philippines

Spark Radio & Electronics Corp.
Tel: 2-775192
TLX: (722) 27901

Portugal, Lisboa

Equipamentos de Electronica e
Cientificos, S.A.R.L.
Tel: 2103420
TLX: (832) 15515

Republic of Singapore

Rank O'Connor's (PTE) Limited
Tel: 4737944
TLX: (786) RS21023

Republic of South Africa, Bramley

Protea PNI
Tel: 11-786-3647
TLX: (960) 4-24409

Spain, Madrid

Equipos y Sistemas S.A.
Tel: 458-0150
TLX: (831) 42856

Sweden, Vallingby

Kaliber AB
Tel: 380350
TLX: (854) 14077

Switzerland, Zurich

Traco Electronic AG
Tel: 2010711
TLX: (845) 815570

Taiwan, Taipei

Schmidt Electronics Corp.
Tel: 5013468
TLX: (785) 11111

Thailand, Bangkok

Measuretronix Ltd.
Tel: 374-2516
TLX: (788) 82796

Turkey, Istanbul

Erkman Elektronik Aletler
Tel: 441546
TLX: (821) 24399

Uruguay, Montevideo

Coasin Uruguay S.A.
Tel: 789015
TLX: (398) UY23010

Venezuela, Caracas

Coasin C.A.
Tel: 241-0309
TLX: (395) 21027

West Germany, Ismaning/Munich

Fluke (Deutschland) GmbH
Tel: 9605240
TLX: (841) 0522472



John Fluke Mfg. Co., Inc., P.O. Box C9090, Everett, WA 98206


Fluke (Holland) B.V., P.O. Box 2269, 5600 CG, Eindhoven, The Netherlands. Phone (040) 458045

Litho in U.S.A. 11/87

INDEX

The number following each index entry indicates the page number.

- Access Procedure, 4-1
- Acceptance Test Procedure, 4-8
- AC Line Voltage Selection, 4-6
- Accessories, 6-1
- Backup Operating Power, 2-5
- Battery,
 - Charge Duration, 1-1
 - Charging, 2-6
 - External, 1-1
 - Internal, 3-1
 - Charger Adjustments, 4-11
 - Damage, 2-6
 - Life, 2-6
 - Pack Option, 6-1
- Calibration, 4-8, 4-9
- Cleaning, 4-6
- Error Sources, 2-6
 - Lead Wire, 2-7
 - Output Impedance, 2-7
 - Minimizing, 2-6
 - Thermal, 2-6
- External Symptom Troubleshooting, 4-14
- Fluke Address, 2-1
- Front and Rear Panel Features, 2-2, 2-3
- Fuse Replacement, 2-5
- Guarded Operation, 2-6
- Input Power Requirements, 2-1
- Installation, 2-1
- Internal Voltage Measurements, 4-16
- Long Term Stability, 2-10
- Output Short Recovery, 1-1
- Oven Temperature Thermistor Nominal Value, 2-7
- Rack-Mounting (see installation)
- Service, 4-1
- Service Centers, 7-1
- Shipping Information, 2-1
- Specifications, 1-2, 1-3, 1-4
- Stability (see long term stability)
- Supply Power, 1-1
- Temperature Range, 1-1
- Transit Case Option, 6-1
- Transportation, 2-1
- Troubleshooting, 4-14



Section 7A

Manual Change Information

7A-1. INTRODUCTION

7A-2. Table 7A-1 defines the assembly revision levels documented in this manual. To identify the configuration of the pca's used in your instrument, refer to the revision letter (marked in ink) on the component side of each pca.

7A-3. NEWER INSTRUMENTS

7A-4. As changes and improvements are made to the instrument, they are identified by incrementing the revision letter marked on the affected pca assembly. These changes are documented on a supplemental change/errata sheet which, when applicable, is inserted at the front of the manual.

Section 8

Schematic Diagrams

FIGURE	TITLE	PAGE
8-1.	Interconnect Diagram	8-3
8-2.	A1 LED, A2 Motherboard and A6 Battery Module PCAs	8-4
8-3.	A3 Pre-Regulator PCA	8-6
8-4.	A4 Regulator PCA	8-8
8-5.	A5 Reference and A5A8 Piggyback PCA's	8-10
8-6.	A8 Preheater PCA	8-12

NOTES






Figure 8-1. Interconnect Diagram




A1 LED

732A-1606



A2 Motherboard

732A-1605



A6 Battery Module

732A-1654

Figure 8-2. A1 LED, A2 Motherboard and A6 Battery Module PCAs




Figure 8-2. A1 LED, A2 Motherboard and A6 Battery Module PCBs (cont)





Figure 8-3. A3 Pre-Regulator



NOTES: UNLESS OTHERWISE SPECIFIED
 1. ALL RESISTANCES ARE IN OHMS AND ALL CAPACITANCE ARE IN MICROFARADS
 2. ALL RESISTORS ARE 1/4 W 5% CC.

REF. DESIG.	
LAST USED	NOT USED
R20	
C10	C7, 8
CR28	CR27, 17
Q6	
S2	
DS1	

Figure 8-3. A3 Pre-Regulator (cont)






Figure 8-4. A4 Regulator PCA




NOTES: UNLESS OTHERWISE SPECIFIED
 1. ALL RESISTANCES ARE IN OHMS AND ALL CAPACITANCE ARE IN MICROFARADS.
 2. ALL RESISTORS ARE 1/4 W 5% CC.

REF	DES
LAST USED	NOT USED
R27	R24
C10	
CR-15	CR1,2,4,5,9,11
Q13	Q9-11,14
DS1	
VR9	VR3-B

Figure 8-4. A4 Regulator PCA (cont)




732A-1608




732A-1601

Figure 8-5. A5 Reference and A5A8 Piggyback PCAs




732A-1001
-1007

Figure 8-5. A5 Reference and A5A8 Piggyback PCAs (cont)



732A-1653



732A-1053

Figure 8-6. A8 Preheater PCA

CHANGE/ERRATA INFORMATION

ISSUE NO: 1 6/86

This change/errata contains information necessary to ensure the accuracy of the following manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCB is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change/errata.

MANUAL

Title: 732A & 732A/AN
Print Date: May 1986
Rev.- Date: ---

C/E PAGE EFFECTIVITY

Page No.	Print Date
1	6/86

ERRATA #1

On page 4-9, paragraph 4-38, following step 6, add the following NOTE:

NOTE

The 10V calibration potentiometer (accessible through a hole in the front panel) has a range of approximately ± 5 ppm. If this range is insufficient to calibrate the 732A 10V output to your standards, the appropriate jumpers on the A7 Calibration PCA must be made or opened. The A7 Calibration PCA is shown in Figure 5-1. Access the A7 Calibration PCA by first removing the top dust cover, then the top guard cover and finally the top foam block. With the 10V calibration potentiometer centered, note the deviation of the 732A 10V output from your standards in ppm. Cutting a jumper increases the 732A 10V output the amount labeled above the jumper. Likewise, completing a jumper decreases the output the amount of the jumper label. After 10V output has been calibrated to your standards, the 1V and 1.018V outputs must also be calibrated.