



# Process Eye Professional Recall (V4)

SP104006 Rev 1.0  
January 2005

As part of our continuous product improvement policy, we are always pleased to receive your comments and suggestions about how we should develop our product range. We believe that the manual is an important part of the product and would welcome your feedback, particularly relating to any omissions or inaccuracies you may discover.

You can send your comments to:-

MKS Instruments, Spectra Products  
Cowley Way  
Crewe  
Cheshire  
CW1 6AG  
U.K.

+44 (0) 1270 250150 Tel.  
+44 (0) 1270 251939 Fax.

email – [manual\\_info@mksinst.com](mailto:manual_info@mksinst.com)

Windows is a trademark of the Microsoft Corporation.

All other brand or product names are trademarks or registered trademarks of their respective companies.

## ***I. Getting Help***

We are always pleased to provide assistance where we can. If you are experiencing any difficulties, or need help, please feel free to call your local MKS Spectra facility and ask for the Customer Support Group. Please have the following information ready so that our technical staff may help you quickly and effectively:

The Process Eye software version number which is printed on the CD ROM, or can also be found on screen by selecting Help | About from the menu bar in Process Eye.


Support Contact Numbers:

United Kingdom  
+44 (0) 1270 250150 Tel.

<b>I. Getting Help .....</b>	<b>3</b>
<b>Section 1 - Starting Recall .....</b>	<b>5</b>
<b>Section 2 - The Main Recall Screen .....</b>	<b>6</b>
1 Title Bar .....	6
2 Menu Bar .....	6
3 File List .....	6
4 Data View Window.....	7
5 Toolbars.....	7
5.1 Default Toolbar.....	7
5.2 Trend Display Toolbar .....	8
5.3 Chart Display Toolbar .....	9
5.4 Scan Navigation Toolbar .....	10
6 Status Bar .....	10
<b>Section 3 - Working with Data Files .....</b>	<b>11</b>
3.1 Opening a file .....	11
3.2 Opening a file in the File List.....	12
3.3 Scrolling through the data file list .....	12
3.4 Closing a file.....	13
3.5 Exporting a file in text format.....	14
3.6 Creating a Baseline File.....	18
<b>Section 4 - Recall Toolbars and Utilities.....</b>	<b>25</b>
4.1 Standard Views .....	25
4.2 Preferences .....	26
4.3 Printing .....	27
4.4 Exiting Recall .....	28
4.5 The Edit Menu .....	29
4.6 The Scan Pull Down Menu .....	30
4.7 View Pull Down Menu .....	31
4.8 The Window Pull Down Menu.....	41


## Section 1 - Starting Recall

Recall can be started in a number of ways:



From the <Start> menu select **All Programs\MKS RGA Applications\Recall**


From the Process Eye or Recipe Wizard program by clicking on the Recall button.



Or from the Process Eye, or Recipe Wizard program, by right-clicking on a recipe and selecting **Recall Data** from the list.

# Section 2 - The Main Recall Screen

The main page of the Recall program consists of six parts:



## 1 Title Bar

The title bar shows the name of the file currently in view. The close, minimize and maximize buttons work in the normal manner.

## 2 Menu Bar

Recall is no different from other Windows based programs, all the available functions are available via the menus.

## 3 File List

The File List contains a list of all the Recall data files in the current directory.





**Cut** cuts the highlighted section of text.



**Copy** copies the highlighted section of text to the clip board.



**Paste** pastes the contents of the clip board into the recipe.



**Cascade windows** cascades all open windows.



**Tile horizontally** tiles all open windows horizontally.



**Tile vertically** tiles all open windows vertically.



**Bring active file to the front** brings all open views for the active file to the front.



**Print** prints the contents of the currently open recipe file.



**Help** opens the online help file for recipe wizard.

## 5.2 Trend Display Toolbar

Three buttons constitute the Trend Display Toolbar:



**View Properties** is used to add a new trend or to edit an existing trend.

Clicking on the View Properties button will cause the Trend Display Pane and Channel Properties windows to be displayed. See Adding Trends for more details of this window.

Note the button only becomes available when there is data to trend i.e. after the second scan is complete.



**Key Display** controls the display of the key table for the active trend views. The button will not be available if no trend view is available.



**Key Mode** changes the key table view between vertical and horizontal views.

### 5.3 Chart Display Toolbar



The nine buttons that make up the Chart Display Toolbar control certain aspects of the active data windows:



**Properties** displays the settings for the currently active recipe view.



**Grid** displays a Y-axis grid in the currently active view window. Default is disabled.



**Log/Lin** switches between a linear and logarithmic Y-axis. Default is linear.



**Y-axis Minor Ticks** enables the display of minor ticks on the Y-axis. Default is major ticks.



**Time** switches between real and elapsed time on the X-axis from the start of the recipe. Default is elapsed time.



**Days** displays the date and time on the X-axis. Default is time only.



**X-axis Minor Ticks** enables the display of minor ticks on the X-axis. Default is major ticks.



**Add Scroll Bar** adds a scroll bar to the X-axis. Default is no scroll bar.



**Add Cursor** adds a measurement cursor to the active Bar chart display. The cursor can be moved to the required mass and will display the mass number and partial pressure. As many cursors as are required can be added.

## 5.4 Scan Navigation Toolbar



The last six buttons control the viewing of stored data:

-  **First Scan** views the oldest scan stored in the file.
-  **Jump Back** views the scan 10 scans previous to the current one being viewed.
-  **Previous Scan** allows you to view the previous scan.
-  **Next Scan** allows you to view the next scan.
-  **Jump Forward** views the scan 10 scans subsequent to the current one being viewed.
-  **Last Scan** views the newest scan acquired in the current recipe run.


## 6 Status Bar

The status bar is along the bottom of the screen and displays “mouse over” help information. It can be switched off by selecting **Status Bar** from the **View** pull down menu.


## Section 3 - Working with Data Files

### 3.1 Opening a file

To open a file using the file open dialog:




Select **Open (Dialog)** from the **File** pull down menu, or right-click on a file in the file list and select **Open New Folder**.



In the "Open" dialog, select the folder and file to open and then click **<Open>**.

### 3.2 Opening a file in the File List



A file from the file list can be opened by selecting **Quick Open** from the **File** pull down menu, or by right-clicking on the file and selecting **Open This**.

### 3.3 Scrolling through the data file list

Once a file is open, other files in the same folder can be opened sequentially using the **Open previous** and **Open next** icons.



The same features are also available from the **File** pull down menu.


These buttons cause the current file to close and the next or previous file in the list one to open.

Opening files in the way allows the next file to be opened with any changes to the current displays being applied to the new file.

**Note:** Changes will not be applied to any views unless the view lock is switched off.



### 3.4 Closing a file



To close an active file, select **Close** from the **File** pull down menu, or click on the close button on the text view.



**Note:** Clicking on the close button on any trend or chart view simply closes that view.



To close all open views, use the Close All icon.




### 3.5 Exporting a file in text format


Data in Process Eye is stored in a digital format that is not easily read by any other program.

Recall therefore has the ability to export data to a tab delimited text in quotes ASCII file.


To export a file in the file list, right-click on the **File** and select **Export as Text** from the list.




If no files are open, selecting **File / Export as Text** will display a file open dialog where single or multiple files can be selected for export.



If the file selected is already open then the following message will be displayed.



This will cause the "Export Data to File" dialog to appear. Here we can specify which data is exported, the options are explained below.



## General Information

Selecting this option includes the header and instrument information in the text file.

```
"[Header Information]"
"Recipe Name"      "C:\Process Eye\Recipes\Barchart Interactive.SI-r"
"Recipe Date"     "15/09/2003 17:16:01"
"Software version" "Process Eye V3"
"File version"    7

"[Instrument_1]"
"Instrument Name" "MiniLab"
"Serial number"  "LM76-00199002"
"Application"    "Atmospheric Inlet Mini-Lab"
"Sensor"         "Standard Enclosed Source"
"Inlet system"   "Atmospheric"
"Inlet 1 Factor" 1.00
"Inlet 2 Factor" 1.00
"Full scale amps" 2.00e-006
"Electronic Gain 1" 1.0
"Electronic Gain 2" 100.0
"Electronic Gain 3" 20000.0
"Faraday gains" 3
"Multiplier gains" 2
"Detector type" 1
"Aux I/O connection" 0
"Maximum mass" 100
"SEM identification" 1500
"Filament type" 0
"Mains frequency" 0
"Peak resolution" 32
"Peaks per amu" 1
```

## Data Acquisition Settings

Selecting this option includes the acquisition setting used on the measurement.

```
"[Acquisition Settings_1]"
"Data Set" "DataSet1"
"Data Stream" "Data"
"Measurement" "BarChart1"
"Acquisition mode" 0
"Channel count" 100
"Instrument" "LM76-00199002"
"First mass" 1.00
"Last mass" 100.00
"Accuracy" 5
"Detector" "Faraday"
"Sensitivity (A/mbar)" 1.50e-006
"Full scale reading (mbar)" 1.33e-002
"Electronic gain ID" 2
"Detector gain" 1.00
"Ion Source" "Standard Electron Energy"
"Emission current" 1.00
"Electron energy" 40.0
"Ion energy" 7.00
"Extractor voltage" -20.0
"Detector voltage" -600.0
"Filament" 1
```

**Note:** These are the settings for each measurement, if there is more than one measurement there will be a separate section for each.

## Alarms and Events

The alarms and events are events that are written to the event log while the recipe is running. To include this information check the box.

```
"[Events]"
"Time" "Duration" "Severity" "Event" "Text1" "Text2"
"28/02/2003 14:01:34" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:01:54" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:02:13" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:02:34" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:02:54" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:03:14" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:03:34" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:03:54" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:04:14" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:04:35" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:04:55" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:05:14" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
"28/02/2003 14:05:34" "n/a" "Yellow" "Calibration" "Sensitivity Calibration Failed"
```

## Data

The pressure of each mass in the measurement. It is exported in the pressure units selected in Recall preferences.

```
"[Scan Data (Pressures in mbar)]"  
"Time"      "Scan"      "Mass 1"    "Mass 2"    "Mass 3"  
"25/09/2003 22:00:09"  14    6.7234e-007 -6.5168e-008  
"25/09/2003 22:00:15"  15    7.2161e-007 5.2452e-008 -4  
"25/09/2003 22:00:21"  16    6.4055e-007 -3.1789e-009
```

## Column Headings

Column headings are identifiers such as the channel names, in the above example column headings are "Time" "Scan" "Mass 1" etc.

## Scan Number

Adds the scan number to the file.

## Millisecond Data

The time will include milliseconds, for example "25/09/2003 22:00:09" will become "25/09/2003 22:00:09.415"

### 3.6 Creating a Baseline File

A baseline file is required for statistical process control, this file contains an average and standard deviation value for each value in a scan.


The file is then used to compare live against stored data and allows the live data to be displayed as a standard deviation from the baseline file instead of as a pressure.

A baseline file would normally include data from one recipe run during a repetitive process. The recipe is created and run normally to generate a number of standard data files. Files relating to a good process run can be then incorporated into the baseline file and the original recipe modified to include this file.

Select **Baseline Wizard** from the **File** pull down menu.



This starts the wizard, the first page is for information only, click **<Next>**.




To create a new baseline file, select the **Create a new baseline file** option, an incrementing file name will be generated automatically and the file stored in the currently active folder.

To change the filename or directory for the baseline file, click the **<Browse>** button.

**Note:** The only files that can be added to a baseline file are those in the currently open folder and listed in the file list. You cannot add files from more than one directory, or from another directory other than the current one.

To add files to an existing baseline file select the **Add files to the existing Baseline file highlighted below** option.




Highlight the file to modify and click **<Next>**.

Click **<Properties>** to view the properties of the selected baseline file.


The files in the current folders are checked to ensure that they were created using the same version of the recipe.


If not, a list of revisions will be displayed and files created by each revision can be included by checking the box next to that revision.

If revisions of recipes have different measurement properties, the data files can not be included in the same data file




If data from recipes with different measurement properties has been selected, the following warning dialog is displayed and one of the recipe revisions would have to be removed before proceeding.






Select one of the revisions and click <Next>




A list of measurements in the selected data file is displayed, you can choose to include all, or any number of the measurements in the baseline file. Simply check the box next to the measurement to include it.

Click <Next>.



The wizard is now ready to create an empty baseline file with the correct structure, click <Next> to continue.


The baseline file is now ready for data to be added.




Because the latest revisions of recipe data files were chosen, the only valid files to add are **Mon000016** and **Mon000017**.

If the earlier revision was chosen, **Mon000001** through to **Mon000015** could be included.

Select the files relating to good process runs by putting a check in the box next to that file.




To continue, click <Next>.



If the data files have differing numbers of scans, the files can be aligned from the beginning so the data from the first scan of each file is compared, then the next and so on.

Alternatively, the files can be aligned to the last scan, so all the data from the last scan of each file is compared and then the next to last scan and so on.


The baseline file can now be included in the actual recipe.  
Edit the recipe using **Recipe Wizard** and add the file on the “Statistical Process Monitoring” dialog.




## Section 4 - Recall Toolbars and Utilities

### 4.1 Standard Views

To modify and create views, use the **View Manager**




Select **Standard Views** from the **File** pull down menu.




The "Manage Views" dialog appears; see the Process Eye Professional manual for more information on View Manager.

## 4.2 Preferences

Preferences are only available if there are no open files.



Select **Preferences** from the **File** pull down menu, then the option required.



Selecting **Defaults** opens the "Preferences" dialog where changes to the way a file opens can be made.


Also available from the **Preferences** menu is the unit of pressure Recall will use in its displays.

When data is exported as text, all pressures are converted to the units selected here.

## 4.3 Printing

Only trend and graph displays can be printed in Recall, if a text window is the active window, or if there are no open files, the print option is disabled.


The print function prints the data currently displayed, if a view has been edited, it will be the edited version that is printed and not the original.



Select **Print** from the **File** pull down menu, or click on the print icon.



## 4.4 Exiting Recall



To exit Recall, select **Exit** from the **File** pull down menu, or click on the close button.



## 4.5 The Edit Menu



The edit pull down menu is only visible when a file is open, the menu is only enabled when modifying text in the comment view (see Page 35)



To undo a change made in the comment view select **Undo** from the **Edit** pull down menu or use the keyboard shortcut **Ctrl-Z**.

 To cut a section of text from the comment view, highlight a section of text then select **Cut** from the **Edit** pull down menu, click on the cut icon. or use the keyboard shortcut **Ctrl-X**

 To copy a section of text from the comment view, highlight a section of text then select **Copy** from the **Edit** pull down menu, click on the copy icon. or use the keyboard shortcut **Ctrl-C**

 To paste a section of text into the comment view, select <Paste> from the Edit pull down menu, click on the paste icon, or use the keyboard shortcut **Ctrl-V**.

## 4.6 The Scan Pull Down Menu

The scan pull down menu is only visible when a file is open.



"First Scan" button displays the oldest scan stored in the file.



"Jump Backwards" button displays the scan 10 scans previous to the scan in view.



"Go to Previous" button displays the scan previous to the current one in view.



"Go to Next" button displays the scan after the current one in view.




"Jump Forwards" displays the scan 10 scans after the current one in view.



"Last Scan" button displays the oldest scan stored in the file.

## 4.7 View Pull Down Menu

### Toolbars



Selecting **Toolbars** from the **View** pull down menu opens the “Toolbars” dialog.




Here you can select if a toolbar is displayed or not.


The **Show Tool tips** option displays “mouse over” help.

The **Cool Look** option changes the menu buttons to 3D style.

## Status Bar




The Status bar can be disabled by removing the check-mark.




The File List can be disabled in the same way.

## File List Properties

There are three additional pieces of information that can be displayed in the file list window, these are Comments, Date/Time and File Size



Examples of each option are shown below:




## Main Window Style

This is the text window which appears with each view, it can display one of three pieces of information:


- The contents of the alarm log,
- The recipe properties,
- The contents of the "comment" field.

By default, the alarm log is normally displayed. Closing this view will also close the file.



**Note:** This option is only visible in the **View** pull down menu when at least one data file is open.

## The Alarm Log view.



Date/Time	Severity	Event	Message	Duration	Data Set
28/02/2003 14:01:20	Blue	Data Files	Storing data to D:\Process Eye 2000\...	n/a	
28/02/2003 14:01:34	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:01:54	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:02:13	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:02:34	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:02:54	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:03:14	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:03:34	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:03:54	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:04:14	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:04:35	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:04:55	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:05:14	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1
28/02/2003 14:05:34	Yellow	Calibration	Sensitivity Calibration Failed: Factor = ...	n/a	DataSet1


The alarm log view displays the contents of the alarm log.

You can use the **Show** pull-down to sort the alarms by alarm severity



## Recipe Properties View


The Recipe Properties view displays information on the instrument and recipe settings.



The information displayed can be filtered using the **Show Properties** pull-down.

## Comment View


The Comment View displays any comments already stored with the file. You can add new comments simply by clicking on the page and typing the new comment.



Once a comment has been added it can be saved by selecting <Save> from the pull down menu or clicking on the Save icon.


## Managing Views

To manage the views in the currently active file:



Select <View Manager> from the View pull down menu.

The "Data File View Manager" dialog will appear.




Any current swept measurement views are displayed in the top list. In the above example there is a view for the measurement Barchart1, to remove it uncheck the box.

New trends can be added by clicking on <Add>.

## The Add Trend Dialog.


The left-hand window displays a list of all the currently defined trend views. The right-hand window displays a list of measurements.




To add a trend, select one of the view definitions from the left-hand list and a measurement to apply it to in the right-hand list.

For example; Selecting Default Trend and Barchart1 adds a default trend using data from Barchart1

The new view will now appear in the "Data File View Manager" dialog.




Click <OK>.



The new view is then displayed.

## View Properties



To open the view properties of an active view you may either, select **Properties** from the **View** pull down menu, or use the Properties icon.



**Note:** If the view lock is enabled then properties will not be available.



## 4.8 The Window Pull Down Menu

The window pull down menu is only visible when a data file is open.




### Add a New Trend


To add a new trend view to the current file, select **New Trend View** from the **Window** pull down menu, or click on the add trend icon



Use the "Trend Display Pane & Channel Properties" dialog to configure the new trend.



Clicking **Edit Channel** allows the user to change the properties of the selected channel, such as the mass it is measuring or if it relates to a user channel.



The list of known gases relating to mass numbers is contained in the text file - **C:\Program Files\Spectra\RGA Applications\Workstation\Common Gases.txt**.

This can be edited or added to, in order to give a more accurate description of the users system. In the above example "Mass 2 – Hydrogen" has been edited to read "Test Gas".

## Close All Windows




To close all open windows select **Close All Windows** from the **Window** pull down, or use the icon.




## Arrange All Windows

To arrange all open windows from all the open data files:



Select **Arrange All Windows** from the **Window** pull down and then choose one of the standard windows arrangement methods.

## To arrange windows on an active File



To arrange the windows on the currently active file, choose from the four options shown.

## Selecting an active view

To select a view and make it the active view, choose it from the list of currently open views.

