

# Analysis of Different Measurement Setups for a Programmable Josephson Voltage Standard

Ralf Behr, Johannes Kohlmann, Jan-Theodor B. M. Janssen, Peter Kleinschmidt, Jonathan M. Williams, Sophie Djordjevic, Jean-Pierre Lo-Hive, François Piquemal, Per-Otto Hetland, Dominique Reymann, Gunnar Eklund, Christian Hof, Blaise Jeanneret, Oleg Chevtchenko, Ernest Houtzager, Helko E. van den Brom, Andrea Soso, Domenico Andreone, Jaani Nissilä, and Panu Helistö

**Abstract**—The electrical characteristics of two different 1-V binary programmable Josephson arrays, an superconductor/insulator/normal conductor/insulator/superconductor-type Josephson array, and an externally shunted superconductor/insulator/superconductor-type Josephson array, were investigated at ten metrology institutes. Various operational parameters were evaluated and compared using different Josephson array voltage standard setups at microwave frequencies around 70 GHz. The results of the measurements show that both arrays have been working very well and the main differences were not imposed by the arrays themselves, but by the different measurement setups of the laboratories.

**Index Terms**—Comparison, Josephson arrays, programmable voltage standard, superconductor/insulator/normal conductor/insulator/superconductor (SINIS), superconductor/insulator/superconductor (SIS) Josephson junctions.

## I. INTRODUCTION

**J**OSEPHSON Array Voltage Standards (JAVS) based on programmable Josephson arrays have increasingly gained in importance in recent years. These programmable arrays are superior to SIS arrays used for conventional JAVS, because the steps are inherently stable and any voltage can be set at a high speed limited only by the bias electronics if the array is binary-divided [1], [2]. This ability to set voltages rapidly leads to reduced measurement uncertainties for certain dc applications, such as FRDC measurements on thermal converters [3], [4], resistance calibrations with a double Josephson potentiometer [5], and probably the Watt experiment [6] and the metrological

triangle [7]. Furthermore, these programmable arrays have the potential for ac metrology [8], [9]. It is, therefore, anticipated that programmable Josephson arrays will replace JAVS based on conventional SIS arrays over the next few years. In particular, the commercial fabrication of these programmable arrays is under way [10] and they will soon be available on the market.

Nevertheless, there is a concern that errors may exist during practical measurement because the arrays are not operated with zero current. Any current in a series resistance would generate an additional voltage that would hinder useful calibrations. When the programmable Josephson array is binary-divided, an easy self-check can be performed to verify the flatness of the steps of constant voltage and to determine the operating point [11], [12].

The aim of the co-operation presented here was to compare JAVS setups for programmable Josephson arrays. The behavior of the two arrays have been investigated for that purpose within the framework of the ongoing EUROMET project no. 626, coordinated by PTB. In this project, the investigations were focussed on dc measurements. The results obtained with a certain JAVS provide information about special features of the setup used. This information has been made available to every participating laboratory. Therefore, the individual setups can be improved and proper operation of the JAVS based on a programmable array will be ensured.

## II. ARRAYS AND MEASUREMENT SETUPS

The two Josephson arrays used in this study have been realized using the reliable superconductor/insulator/superconductor (SIS) array was fabricated at VTT. It consists of 3488 Josephson junctions divided into 5 bits. The junctions are arranged in 16 parallel microwave stripline paths. The 1-V level is achieved at 70 GHz, when the array is operated on the second step. The intrinsically shunted superconductor/insulator/normal conductor/insulator/superconductor (SINIS) array was made at PTB. It has a nominal number of 8192 junctions divided into 6 bits, but the largest bit has a superconductive shorted junction and therefore just 4095 junctions. The shorted junction has no effect on the behavior of the array. The microwave stripline is distributed into 64 parallel paths. A more detailed description of the layouts and the fabrication processes is given elsewhere [13]–[16].

The Josephson arrays were sent from laboratory to laboratory. The time for measuring both arrays was limited to three weeks in

Manuscript received June 17, 2002; revised November 4, 2002.

R. Behr and J. Kohlmann are with the Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany.

J.-T. B. M. Janssen, P. Kleinschmidt, and J. M. Williams are with the National Physical Laboratory (NPL), Middlesex, U.K.

S. Djordjevic, J.-P. Lo-Hive, and F. Piquemal are with the BNM-LNE, Fontenay-aux-Roses, France.

P. O. Hetland is with the Justervesenet (JV), Kjeller, Norway.

D. Reymann is with the Bureau International des Poids et Mesures (BIPM), Sèvres Cedex, France.

G. Eklund is with the SP Swedish National Testing and Research Institute, Borås, Sweden.

C. Hof and B. Jeanneret are with the METAS, Swiss Federal Office of Metrology and Accreditation, Bern-Wabern, Switzerland.

O. Chevtchenko, E. Houtzager, and H. van den Brom are with the NMi Van Swinden Laboratorium, Delft, The Netherlands.

A. Soso and D. Andreone are with the Istituto Elettrotecnico Nazionale Galileo Ferraris (IEN), Torino, Italy.

J. Nissilä is with the MIKES, Electricity and Time, Espoo, Finland.

P. Helistö is with the VTT Microsensing, Espoo, Finland.

Digital Object Identifier 10.1109/TIM.2003.811570

most cases. The parameters investigated are those related to critical currents, sensitivity to flux trapping, and functionality of all bits, as well as step width as a function of the microwave power and microwave frequency. To perform the measurements, every institute had to mount the arrays to a cryoprobe and to use the equipment of the laboratory's Josephson voltage standard. All participants use an EIP counter to phase lock the frequency, with the exception of IEN. The Gunn oscillators vary in power and frequency, i.e., from 60 to 85 GHz. As various kinds of waveguides were used, with attenuation ranging from 1 dB for oversized circular waveguides to 6 dB for rectangular ones, a wide range of maximum microwave power from about 3 to 50 mW was applied at the finline antenna of the Josephson arrays. In addition, several institutes carried out direct or indirect comparisons with SIS-type, SNS-type, and SINIS-type Josephson arrays to evaluate the calibration capability of the arrays and to verify the quantization of steps.

### III. RESULTS

The results obtained for the critical currents and step width are summarized in Fig. 1. Without microwave power applied, the parameters do not differ significantly. For the SINIS array, all measurements of the critical current are within 900 and 1050  $\mu\text{A}$ . For the externally shunted SIS array, all measurements of the critical current are within 180 to 220  $\mu\text{A}$ , with the exception of two unexplained measurements at about 140  $\mu\text{A}$ . All participants used the bias source they would normally use with SIS Josephson arrays, so noise levels are expected to be sufficiently low for the evaluation of the arrays (references or the link to the different articles describing the Josephson measurement setups are found in [17] under *Metrology Area "Electricity and Magnetism," Comparison identifier BIPM.EM-K10.a&b*).

Under microwave irradiation, the parameters, such as the maximum step width, vary from one laboratory to the next. This is mainly due to different operating frequencies and/or microwave power available in conjunction with the array's step width dependencies upon frequency and power [17]. Due to these variations, it is necessary to individually determine the optimum operating point for each setup. A Josephson array with a well-known step width in dependence of power and frequency might be useful to establish these conditions. Fig. 2 shows the step width for the SINIS-type array as a function of the applied microwave power at the finline antenna. A large first step has been observed over a wide range of microwave power. As NPL, SP, and NMi used cryoprobes with high attenuation, the resulting maximum step width for the SINIS-type array was found to be smaller than at other laboratories (see Fig. 1). The step width of the shunted array as a function of frequency is shown in Fig. 3. The step width slightly varies over the large microwave frequency range from about 60 to 78 GHz. At high frequencies, the step width decreases.

Special incidents or measurements are very important in this kind of co-operation as they give further information about problems that could appear with the arrays. In contrast to SIS-type Josephson arrays, operated at zero current, the current-voltage characteristic changes with temperature. The bias




Fig. 1. (Top): Measured critical current of the arrays without microwave irradiation. Arrows to the left indicate the SINIS-type array; arrows to the right the shunted SIS-type array. (Bottom): Measured maximum step width of the first step of the SINIS-type array and of the first and second step of the shunted SIS-type array under microwave irradiation. The step width is measured by detecting a perceptible change of the step voltage to define the step limits (see Fig. 4).




Fig. 2. Step width in dependence upon the microwave power at a frequency of 69.5 GHz for six segments of the SINIS-type array. For clarity, the curves corresponding to the positive steps (step p) have been shifted by 0.5 mA. The measurements were made by METAS.

currents must therefore be adjusted if necessary. For instance, a continuous drift of the step position of an SINIS-type array to higher bias currents is an indication of a falling helium level.

A few institutes observed that the arrays suddenly trapped flux, sometimes without identifiable reason. In one case, it was due to cooling the array too fast. Overall, both arrays showed very small rate of trapped flux.

For measurements with rapidly changing voltages, the choice of filters is a crucial problem. On one hand, large steps are expected to have high immunity to noise that would allow small




Fig. 3. Step width as a function of frequency for the shunted SIS-type array. The microwave power is estimated at the finline antenna of the array and unknown for the BIPM measurements. The measurements were made by BNM/LNE using an oversized circular waveguide, BIPM using a dielectric waveguide, and IEN using a stainless steel waveguide.




Fig. 4. Flatness of a step of the SINIS-type array measured at the 1-V level using a second JAVS. Removing additional filters between the EM N1a null-detector and the SINIS-type array increases the cut-off frequency of the filter and results in a sloped step. The measurements were made by PTB.

filters to be used. On the other hand, narrow bandwidth filters might support noise-induced phase slippage that might result in sloped steps and very large measurement uncertainties. Fig. 4 shows the flatness of a step of the SINIS array measured by PTB at the 1-V level using a second Josephson voltage standard. Removing additional filters between the EM N1a null-detector and the cryoprobe, a sloped step appeared probably due to phase slippage in the array induced by the noise of the null-detector. The same behavior was found for the externally shunted array. While these measurements indicate that noise immunity and inherent stability of the steps is not the same, i.e., even large steps do not have *per se* a high noise immunity, this problem is easy to overcome by appropriate filtering.

In contrast to conventional JAVS, the steps are inherently stable, but occur at nonzero current. As any small series resis-




Fig. 5. Comparisons made with both arrays. PTB: direct comparison with another SINIS-type array. BNM/LNE, BIPM, and IEN: direct comparison with SIS-type arrays. SP: indirect comparison with an SIS-type array via Weston cells. METAS: direct comparison with an SNS-type array.

tance within the series arrays could lead to errors, the flatness of the steps has to be verified carefully. This was investigated with self-checks and direct comparisons. Self-checks were carried out at NPL and METAS. Two parts of the array with the same number of Josephson junctions were biased in series opposition. By varying the current on the steps, a slope can easily be found. The voltage difference was measured with a null-detector with sub-nanovolt resolution. To obtain an exact null voltage, the numbers of Josephson junctions need to be equal. Unfortunately, the shorted junction prevented to meet this condition. Thus, the uncertainty was increased to about 1 nV due to the small unbalanced offset voltage.

Fig. 5 shows the indirect and direct comparisons of both arrays with other arrays. All measurements agree very well, e.g., at BIPM, the direct comparisons agree to  $-0.05 \pm 0.12$  nV for the SINIS array and  $-0.09 \pm 0.09$  nV for the shunted SIS array. The measurement uncertainty corresponds to a one standard deviation of the mean. More details of the comparison procedure at BNM-LNE are given in [18].

#### IV. CONCLUSION

Both programmable Josephson arrays were found to be working very well in all laboratories. The main differences in electrical parameters were caused by the different measurement setups of the laboratories. The agreement of the output voltages at the 1-V level between the SINIS-type array or the externally shunted SIS-type array and other Josephson arrays was found to be better than 0.5 nV for all direct comparisons.

#### REFERENCES

- [1] C. A. Hamilton, C. J. Burroughs, and R. L. Kautz, "Josephson D/A converter with fundamental accuracy," *IEEE Trans. Instrum. Meas.*, vol. 44, pp. 223–225, Apr. 1995.
- [2] S. P. Benz, C. A. Hamilton, C. J. Burroughs, T. E. Harvey, and L. A. Christian, "Stable 1 volt programmable voltage standard," *Appl. Phys. Lett.*, vol. 71, pp. 1866–1868, Sept. 1997.
- [3] C. J. Burroughs, S. P. Benz, C. A. Hamilton, and T. E. Harvey, "Thermoelectric transfer differences of thermal converters measured with a Josephson source," *IEEE Trans. Instrum. Meas.*, vol. 48, pp. 282–284, Apr. 1999.
- [4] T. Funck, R. Behr, and M. Klonz, "Fast reverse DC measurements on thermal converters using a SINIS Josephson junction array," *IEEE Trans. Instrum. Meas.*, vol. 50, pp. 322–325, Apr. 2001.

- [5] R. Behr, L. Grimm, T. Funck, J. Kohlmann, H. Schulze, F. Müller, B. Schumacher, P. Warnecke, and J. Niemeyer, "Application of Josephson series arrays to a DC quantum voltmeter," *IEEE Trans. Instrum. Meas.*, vol. 50, pp. 185–187, Apr. 2001.
- [6] W. Beer, A. L. Eichenberger, B. Jeanneret, B. Jeckelmann, P. Richard, H. Schneider, A. R. Pourzand, A. Courteville, and R. Dändliker, "The OFMET watt balance: Progress report," *IEEE Trans. Instrum. Meas.*, vol. 50, pp. 583–587, Apr. 2001.
- [7] F. Piquemal and G. Genevès, "Argument for a direct realization of the quantum metrological triangle," *Metrologia*, vol. 37, pp. 207–211, 2000.
- [8] S. P. Benz, C. J. Burroughs Jr, P. D. Dresselhaus, and L. A. Christian, "AC and DC voltages from a Josephson arbitrary waveform synthesizer," *IEEE Trans. Instrum. Meas.*, vol. 50, pp. 181–184, Apr. 2001.
- [9] J. M. Williams, T. J. B. M. Janssen, R. Behr, J. Kohlmann, and F. Müller, "Synthesis of precision AC waveforms using a SINIS Josephson junction array," in *CPEM Conf. Dig.*, 2002, pp. 434–435.
- [10] M. Schubert, T. May, G. Wende, L. Fritzsche, and H.-G. Meyer, "Coplanar strips for Josephson voltage standard circuits," *Appl. Phys. Lett.*, vol. 79, pp. 1009–1011, Aug. 2001.
- [11] C. J. Burroughs, S. P. Benz, T. E. Harvey, and C. A. Hamilton, "1 volt DC programmable Josephson voltage standard," *IEEE Trans. Appl. Superconduct.*, vol. 9, pp. 4145–4149, June 1999.
- [12] B. Jeanneret, A. Rüfenacht, and C. Burroughs, "High precision comparison between SNS and SIS Josephson voltage standard," *IEEE Trans. Instrum. Meas.*, vol. 50, pp. 188–191, Apr. 2001.
- [13] J. Hassel, H. Seppä, L. Grönberg, and I. Suni, "SIS junctions with frequency dependent damping for a programmable Josephson voltage standards," *IEEE Trans. Instrum. Meas.*, vol. 50, pp. 195–198, Apr. 2001.
- [14] —, "One volt Josephson voltage standard based on frequency dependently damped SIS junctions," *Rev. Sci. Instrum.*, to be published.
- [15] R. Behr, H. Schulze, F. Müller, J. Kohlmann, and J. Niemeyer, "Josephson arrays at 70 GHz for conventional and programmable voltage standards," *IEEE Trans. Instrum. Meas.*, vol. 48, pp. 270–273, Apr. 1999.
- [16] O. Kieler, R. Behr, F. Müller, H. Schulze, J. Kohlmann, and J. Niemeyer, "Improved 1 V programmable Josephson voltage standard using SINIS junctions," *Physica C*, vol. 372–376, pp. 309–311, 2002.
- [17] BIPM.EM-K10.a&b. [Online]. Available: <http://kcdb.bipm.fr/BIPM-KCDB>
- [18] J. P. Lo-hive, S. Djordjevic, P. Cancela, F. Piquemal, R. Behr, C. J. Burroughs, and H. Seppä, "Characterization of Binary Josephson Series Array of Different Types at BNM-LNE and Comparisons With Conventional SIS Arrays," *IEEE Trans. Instrum. Meas.*, vol. 52, Apr. 2003.

**Jan-Theodor B.M. Janssen**, photograph and biography not available at the time of publication.

**Peter Kleinschmidt** was born in Lüneburg, Germany, on February 5, 1969. He studied physics at the University of Heidelberg, Heidelberg, Germany, and at the University of Massachusetts, Amherst.

In 1997, he joined the Centre for Electromagnetic Metrology, National Physical Laboratory, Middlesex, U.K., where he has been working on the development of cryogenic current comparator ratio bridges, programmable Josephson voltage standards, and the European COUNT project.

**Jonathan M. Williams**, photograph and biography not available at the time of publication.



**Sophie Djordjevic** was born in France in 1972. She received the Ph.D. degree in solid state physics from the University Pierre et Marie Curie, Paris, France, in 1998, where she was working on high-temperature superconductors.

After a postdoctoral training at CEA, Saclay, France, where she worked on the Wigner transition in semiconductor heterostructures, she joined the Bureau National de Métrologie, Laboratoire National d'Essais (BNM-LNE), Fontenay aux Roses, France, in 2002. She is now involved in the work on

ac and dc Josephson voltage standards.

**Jean-Pierre Lo-Hive**, photograph and biography not available at the time of publication.



**Ralf Behr** was born in Stade, Germany, on January 25, 1963. He received the M.S. and Ph.D. degrees in physics in 1990 and 1996, respectively, from the University of Hamburg, Hamburg, Germany, where he was working in the field of vortex dynamics in high-temperature superconductors.

In 1996, he joined the Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany, where he has been engaged in the development of the Josephson voltage standard.

Dr. Behr is a member of the Deutsche Physikalische Gesellschaft (DPG).



**Johannes Kohlmann** was born in Dortmund, Germany, on November 23, 1960. He received the M.S. and Ph.D. degrees in physics from the Georg-August-University, Göttingen, Germany, in 1986 and 1990, respectively, where he was working in the field of low temperature solid state physics.

In 1990, he joined the Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany, where he has been engaged in the research and development of the Josephson potentiometer and Josephson junction series arrays for the PTB's

voltage standard. Since 1997, he has been Head of a PTB project on a programmable Josephson voltage standard.

Dr. Kohlmann is a member of the Deutsche Physikalische Gesellschaft (DPG).



**François P. M. Piquemal** was born in Bois d'Arcy, France, in October 1960. He received the Doctorate degree in solid state physics from the Université de Jussieu, Paris, France, in 1988, where he was involved in the study of magnetic properties of heavy-fermion compounds in normal and superconducting states.

Since 1988, he has been with the National Metrology Laboratory, BNM-LNE (formerly BNM-LCIE), Fontenay-aux-roses, France, implied in the domain of electricity and magnetism. At

the beginning, his activity was focused on the quantum Hall effect and the resistance metrology, and then it extended onto single electron tunneling experiments. He is now in charge of the electrical metrology department of the BNM-LNE. His main research interests include the quantum electrical metrology, the instrumentation based on cryogenic current comparators, and the determination of fundamental constants.


**Per-Otto Hetland**, photograph and biography not available at the time of publication.

**Dominique Reymann**, photograph and biography not available at the time of publication.



**Gunnar Eklund** received the M.Sc. degree in physics from the University of Gothenburg, Sweden, in 1985.

He joined the Laboratory for Electricity and Time at SP, Borås, Sweden, in 1985 and is now working in the fields of low frequency impedance measurements, Josephson array measurements, and quantized Hall effect measurements.



**Christian Hof** was born in Solothurn, Switzerland, in 1970. He received the Ph.D. degree in solar cell physics from the University of Neuchâtel, Neuchâtel, Switzerland, in 1999.

After spending a brief time working in the industry he joined the Swiss Federal Office of Metrology and Accreditation (METAS), Bern-Wabern. He is mainly involved in research on a capacitance standard based on counting electrons and on Josephson junction arrays.



**Blaise Jeanneret** was born in Fleurier, Switzerland, in 1959. He received the Ph.D. degree in experimental condensed matter physics from the University of Neuchâtel, Neuchâtel, Switzerland, in 1989.

He then spent two years as a Guest Scientist at the National Institute of Standards and Technology, Boulder, CO, working on high-temperature superconductors. In 1993 he joined the Institute for Micro- and Optoelectronics of the Swiss Federal Institute of Technology, Lausanne, where he was

involved in research on the quantum Hall effect. In 1996, he moved to the Swiss Federal Office of Metrology and Accreditation (METAS), Bern-Wabern, where he continued his work on the quantum Hall effect. He became head of the quantum metrology laboratory at METAS in 1997. His present work is focused on quantum electrical metrology.

**Oleg Chevtchenko**, photograph and biography not available at the time of publication.

**Ernest Houtzager**, photograph and biography not available at the time of publication.

**Helko E. van den Brom** was born in Utrecht, The Netherlands, in 1971. He received the M.Sc. degree in theoretical solid state physics from Utrecht University, Utrecht, in 1995, and the Ph.D. degree in experimental solid state physics from Leiden University, The Netherlands in 2000.

In 2000, he joined the Nmi Van Swinden Laboratorium, Delft, The Netherlands, where his focus is on the development of Josephson and SET-based electrical quantum standards. Currently, he is coordinator of the European COUNT project toward a SET-based standard for electrical current.



**Andrea Sosso** received the degree in electronic engineering in 1990 and the Ph.D. degree in metrology, with a thesis on this subject, in 1994.

After graduation, he had been involved in the study of a device for the comparison of electrical and mechanical power. He became a Researcher in the Electrical Metrology Department, Istituto Elettrotecnico Galileo Ferraris (IEN), Torino, Italy, in 1996. He participated in the development and analysis of the ac quantum Hall effect measurement system set up at the IEN, and spent one year at the National Institute of Standards and Technology working in JAVS dc voltage calibrations and on the maintenance of a voltage standard. He is currently responsible of the dc voltage standard at IEN. His main interests concern dc voltage standards, precision measurements, and design of electronic circuits

**Domenico Andreone**, photograph and biography not available at the time of publication.



**Jaani Nissilä** was born in 1969. He received the D.Sc. (Tech.) degree in applied physics from the Department of Engineering Physics and Mathematics, Helsinki University of Technology, Helsinki, Finland, in 2001. His thesis focused on defects in semiconductors.

He joined the Centre for Metrology and Accreditation (MIKES), Espoo, Finland, at the beginning of 2002. Currently, he is participating in the development of ac voltage standards based on Josephson junctions. He is also involved in the development of

RF metrology and has been involved in the development of instrumentation for positron annihilation spectroscopies.



**Panu Helistö** was born in Finland in 1956. He received the M.Sc. and Dr.Tech. degrees in technical physics from the Helsinki University of Technology, Helsinki, Finland, in 1981 and 1987, respectively.

He has studied time coherence of resonant gamma radiation, physics of steel machinability, and biomedical engineering and signal processing at Helsinki University of Technology. Since 1998, he has worked on electrical metrology at VTT Information Technology, Espoo, Finland, and the Centre for Metrology and Accreditation (MIKES), Espoo.

Presently, he leads the quantronics group at VTT Information Technology, developing applications of Josephson junction devices.